

Monte Carlo applied to efficiency computation

K.M. Waino and G.F. Knoll, Nucl. Instr. And Meth. A 44 (1966) 213.

De Castro Faria and Levesque, Nucl. Instr. And Meth. A 46 (1967) 325.

K. Debertin, B. Grosswendt

Efficiency calibration of semiconductor detectors by primary standard sources and Monte Carlo calculations

Nucl. Instr. and Meth. 203 (1982) 343-352

A.I. Hawari and R.F. Flemming

High accuracy determination of the shape of the efficiency curve of the HPGe detector in the energy range 900 to 1300 keV

Nucl. Instrum. and Methods A 353 (1994) 106-108

G. Haase, D.Tait, A. Wiechen

Determination of full energy peak efficiency for cylindrical volume sources by the use of a point source standard in gamma spectrometry

Nucl. Instrum. and Methods in Phys. Res. A361 (1995) 240-244

A.I. Hawari, R.F. Flemming and M.A. Ludington

High-accuracy determination of the relative full energy peak efficiency curve of a coaxial HPGe detector in the energy range 700-1300 keV

Nucl. Instr. And Meth. A398 (1997) 276.

M.A. Ludington, R.G. Helmer

High accuracy measurement and Monte Carlo calculations of the relative efficiency curve of an HPGe detector from 433 to 2754 keV

Nucl. Instrum. and Methods in Phys. Res. A446 (2000) 506-521

J. Ródenas, A. Martinavarro, V. Rius

Validation of the MCNP code for simulation of Ge-detector calibration

Nucl. Instrum. and Methods in Phys. Res. A450 (2000) 88-97

M. Garcia-Talavera, H. Neder, M.J. Daza and B. Quintana

Towards a proper modeling of detector and source characteristics in Monte Carlo simulations

Appl. Radiat. Isot. 52 (2000) 777-783

I.O.B. Ewa, D. Bodizs, Sz. Czifrus, Zs. Molnar

Monte Carlo determination of full energy peak efficiency for a HPGe detector

Appl. Radiat. Isot. 55 (2001) 103-108

D. Karamanis, V. Lacoste, S. Andriamonje, GF. Barreau, M. Petit

Experimental and simulated efficiency of a HPGe detector with point-like and extended sources

Nucl. Instrum. and Methods in Phys. Res. A487 (2002) 477-487

J.C. Hardy, V.E. Jacob, M. Sanchez-Vega, R.T. Effinger, P. Lipnik, V.E. Mayes, D.K. Willis and R.G. Helmer

Precise efficiency calibration of an HPGe detector : source measurement and Monte Carlo calculations with sub-percent precision

Appl. Radiat. Isot. 56 (2002) 65-69

Z. Wang, B. Kahn, J.D. Valentine

Efficiency calculation and coincidence summing correction for germanium detectors by Monte Carlo simulation

IEEE Trans. Nucl. Sci. 49 N°4 (2002), 1925-1931

J. Ródenas, A. Pascual, I. Zarza, V. Serradell, J. ortiz, L. Ballesteros

Analysis of the influence of germanium dead layer on detector calibration simulation for environmental radioactive samples using the Monte Carlo method

Nucl. Instrum. and Methods in Phys. Res. A496 (2003) 390-399

D. Karamanis

Efficiency simulation of HPGe and Si(LI) detectors in gamma-and X-ray spectroscopy

Nucl. Instrum. and Methods in Phys. Res. A505 (2003) 282-285

R.G. Helmer, J.C. Hardy, V.E. Jacob, M. Sanchez-Vega, R.G. Neilson and J. Nelson

The use of Monte Carlo calculations in the determination of a Ge detector efficiency curve

Nucl. Instrum. and Methods in Phys. Res. A511 (2003) 360-381

S. Hurtado, M. García-Léon, R. García-Tenorio

GEANT4 code for simulation of a germanium gamma-ray detector and its application to efficiency calibration

Nucl. Instrum. and Methods in Phys. Res. A518 (2004) 764-774

R.G. Helmer, N. Nica, J.C. Hardy, V.E. Jacob

Precise efficiency calibration of an HPGe detector up to 3.5 MeV, with measurements and Monte Carlo calculations

Appl. Radiat. Isot. 60 (2004) 173-177

S. Hurtado, M. García-Léon, R. García-Tenorio

Monte Carlo simulation of the response of a germanium detector for low-level spectrometry measurements using GEANT4

Appl. Radiat. Isot. 61 (2004) 139-143

J. Saegusa, K. Kawasaki, A. Mihara, M. Ito, M. Yoshida

Determination of detection efficiency curves of HPGe detectors on radioactivity measurement of volume samples

Appl. Radiat. Isot. 61 (2004) 1383-1390

C.M. Salgado, C. Conti, P. Becker

Determination of HPGe detector response using MCNP5 for 20-150 keV X-rays

Appl. Radiat. Isot. 64 (2006) 700-705

V. Peyres E. García-Toraño

Efficiency calibration of an extended-range Ge detector by a detailed Monte Carlo simulation

Well-type detector

J.-M. Laborie, G. Le Petit, D. Abt, M. Girard

Monte Carlo calculation of the efficiency response of a low-background well-type HPGe detector

Nucl. Instrum. and Methods in Phys. Res. A 479 (2002) 618-630

Efficiency transfer

M. Jurado Vargas, N. Cornejo Díaz, D. Pérez Sánchez

Efficiency transfer in the calibration of a coaxial p-type HPGe detector using the Monte Carlo method

Appl. Radiat. Isot. 58 (2003) 707-712