

¹²⁴ Sb - ACTIVITY MEASUREMENT AND DETERMINATION OF PHOTON EMISSION INTENSITIES

par

Marie-Martine BÉ, B. CHAUVENET, AND ALL THE PARTICIPANTS IN THE EUROMET ACTION 907 : M.-N. AMIOT, C. BOBIN, M.-C. LÉPY, T. BRANGER, I. LANIÈCE, A. LUCA, M. SAHAGIA, A.-M. WÄTJEN, K. KOSSERT, O. OTT, O. NÄHLE, P. DRYAK, J. SOCHOROVÀ, P. KOVAR, P. AUERBACH, T. ALTZITZOGLOU, S. POMMÉ, G. SIBBENS, R. VAN AMMEL, J. PAEPEN, A. IWAHARA, J.U. DELGADO, R. POLEDNA, L. JOHANSSON, A. STROAK, C. BAILAT, Y. NEDJADI, P. STRING

CEA SACLAY

DIRECTION DE LA RECHERCHE TECHNOLOGIQUE LABORATOIRE D'INTÉGRATION DES SYSTÈMES ET DES TECHNOLOGIES DÉPARTEMENT DES TECHNOLOGIES DU CAPTEUR ET DU SIGNAL LABORATOIRE NATIONAL HENRI BECQUEREL



DIRECTION DES SYSTÈMES D'INFORMATION

RAPPORT CEA-R-6222

2009

CEA / SACLAY 91191 GIF-SUR-YVETTE CEDEX FRANCE

RAPPORT CEA-R-6222 – Marie-Martine Bé, B. Chauvenet, M.-N. Amiot, C. Bobin, M.-C. Lépy, T. Branger, I. Lanièce, A. Luca, M. Sahagia, A.-M. Wätjen, K. Kossert, O. Ott, O. Nähle, P. Dryak, J. Sochorovà, P. Kovar, P. Auerbach, T. Altzitzoglou, S. Pommé, G. Sibbens, R. Van Ammel, J. Paepen, A. Iwahara, J.U. Delgado, R.Poledna, L. Johansson, A. Stroak, C. Bailat, Y. Nedjadi, P. String

«124Sb - Mesure de l'activité et détermination des intensités photoniques»

Résumé - La traçabilité de l'antimoine 124, en terme de mesure d'activité, est très limitée. Les valeurs, venant de laboratoires internationaux de métrologie, contribuant au Système International de Référence sont au nombre de trois. Deux laboratoires ont utilisé la méthode $4\pi\beta$ - γ de mesure en coïncidence, le troisième un détecteur NaI cristal puit. Les deux premiers résultats sont en accord, mais le troisième diffère de 2 %. Diverses hypothèses ont été évoquées pour expliquer cette divergence parmi lesquelles la cohérence du schéma de désintégration. Par ailleurs, ce radionucléide émet des rayonnements gamma de haute énergie et donc pourrait être utilisé pour l'étalonnage en rendement des détecteurs gamma, si les intensités d'émission sont bien connues. Ces remarques ont conduit à proposer un exercice international de mesure de l'activité et des intensités photoniques. Cet exercice, dirigé par le CEA/List-LNE/LNHB, a été enregistré comme projet Euromet 907. Il a été demandé aux participants de mesurer l'activité d'une solution d'antimoine 124 fournie par le LNHB à l'aide de toutes les méthodes disponibles dans leur laboratoire, dans le but de confirmer ou non l'existence d'un problème lié à une, ou plusieurs, techniques de mesure. De plus, les participants ont été sollicités pour mesurer les intensités des émissions photoniques émises lors de la désintégration. Ces résultats ont été comparés avec les valeurs disponibles dans la littérature et un nouveau schéma de décroissance est proposé. Huit laboratoires nationaux de métrologie ont participé à cet exercice

2009 - Commissariat à l'Énergie Atomique - France

RAPPORT CEA-R-6222 – Marie-Martine Bé, B. Chauvenet, M.-N. Amiot, C. Bobin, M.-C. Lépy, T. Branger, I. Lanièce, A. Luca, M. Sahagia, A.-M. Wätjen, K. Kossert, O. Ott, O. Nähle, P. Dryak, J. Sochorovà, P. Kovar, P. Auerbach, T. Altzitzoglou, S. Pommé, G. Sibbens, R. Van Ammel, J. Paepen, A. Iwahara, J.U. Delgado, R.Poledna, L. Johansson, A. Stroak, C. Bailat, Y. Nedjadi, P. String

«¹²⁴Sb – Activity measurement and determination of photon emission intensities »

Abstract - The international traceability of antimony 124, in term of activity, is very limited. The results of ¹²⁴Sb activity measurements sent to the SIR (BIPM - International System of Reference, BIPM.RI(II)-K1.Sb-124.) are scarce. Up to now, only three laboratories have contributed. Two of them carried out measurements using the $4\pi\beta$ - γ coincidence counting technique and the third one using the $4\pi\gamma$ method with a well-type crystal detector. The first two results are in agreement but the last one differs significantly from them, by 2 %. The decay scheme consistency cannot be excluded when trying to explain those discrepancies. In other respects, this nuclide emits high-energy gamma rays, and then could be selected as a valuable standard radionuclide for the calibration of gamma-ray detectors in that energy range, given well known photon intensities. Those considerations led to the proposal of an international exercise and to the realisation of this Euromet project, registered as project nº 907, coordinated by CEA/List-LNE/LNHB. The first part of this exercise was dedicated to activity measurements and to their comparison. For this purpose, participants were asked to make use of all the direct measurement techniques available in their laboratory in order to confirm or not the existence of possible biases specific to some measuring methods. In addition, this exercise offered the opportunity of improving the uncertainties of the gamma-ray intensities. Then, participants were asked, in the second part of the exercise, to carry out X-ray and gamma-ray intensity measurements. These results have been compared to previous published values and new decay scheme data are proposed. Eight international laboratories participated in this exercise.

2009 - Commissariat à l'Énergie Atomique - France

- Rapport CEA-R-6222 -

CEA Saclay

Direction de la Recherche Technologique Laboratoire d'Intégration des Systèmes et des Technologies Département des Technologies du Capteur et du Signal Laboratoire National Henri Becquerel

¹²⁴ Sb – ACTIVITY MEASUREMENT AND DETERMINATION OF PHOTON EMISSION INTENSITIES

par

Marie-Martine Bé, B. Chauvenet, and all the participants in the Euromet action 907 :

M.-N. Amiot, C. Bobin, M.-C. Lépy, T. Branger, I. Lanièce, A. Luca, M. Sahagia, A.-M. Wätjen,
K. Kossert, O. Ott, O. Nähle, P. Dryak, J. Sochorovà, P. Kovar, P. Auerbach, T. Altzitzoglou,
S. Pommé, G. Sibbens, R. Van Ammel, J. Paepen, A. Iwahara, J.U. Delgado, R.Poledna,
L. Johansson, A. Stroak, C. Bailat, Y. Nedjadi, P. String.

¹²⁴Sb – Activity measurement and determination of photon emission intensities

Part A – Euromet 907 Results

1. Introduction

2. Decay scheme of antimony 124

3. Programme and procedure

- 3.1. General description
- 3.2. Properties of the distributed solution
- 3.3. Impurity checks
- 3.4. List of participating laboratories

4. Activity measurements

- 4.1. Methods
- 4.2. Source preparation
- 4.3. Results
- 4.4 Findings for the activity measurements

5. Half-life measurement

6. Photon emission intensities determination

- 6.1. Experimental arrangements
- 6.2. Efficiency calibration and analysis procedure
- 6.3. Uncertainties
- 6.4. Statistical treatment of data
- 6.5. Results of **g** ray intensity measurements
- 6.6. Comparison with previous published data
- 6.7. Euromet 907, x-ray intensity results
- 6.8. Findings for the determination of photon emission intensities

7. Conclusions on the overall exercise

Appendix A1: "Characteristic of the radioactive solution"

Appendix A2: Questionnaire "Description of the measurement technique"

Appendix A3: Description of gamma spectrometry arrangements and techniques

Appendix A4: Erratum Participant 6

Part B – ¹²⁴Sb Decay data evaluation

- I- Comments on evaluation of decay data
- II- Tables and decay scheme data

Euromet 907

¹²⁴Sb – Activity measurement and determination of photon emission intensities

Part A – Euromet 907 Results

1. Introduction

¹²⁴Sb is a fission product. As such, this radionuclide has to be taken into account in the management of short-lived wastes of nuclear industry. This radionuclide can be used also as a tracer in industry because of its high-energy high-intensity photon emissions (603, 723, 1691 keV) and its relatively short half-life (60,2 d).

The international traceability of this radionuclide is very limited. The results of ¹²⁴Sb activity measurements sent to the SIR (BIPM – International System of Reference, BIPM.RI(II)-K1.Sb-124.) are scarce. Up to now, only three laboratories have contributed. Two of them carried out measurements using the $4\pi\beta$ - γ coincidence counting technique and the third one using the $4\pi\gamma$ method with a well-type crystal detector. The first two results are in agreement but the last one differs significantly from them, by 2 %, [C. Michotte *et al.* 2006 *Metrologia* **43**, Tech. Suppl. 06007]. The decay scheme consistency cannot be excluded when trying to explain those discrepancies.

On the other hand, this nuclide emits high-energy gamma rays, and then could be selected as a valuable standard radionuclide for the calibration of gamma-ray detectors in that energy range, given well known photon intensities.

Those considerations led to the proposal of an international exercise and to the realisation of this Euromet project, registered as project n° 907, coordinated by LNE-LNHB. The first part of this exercise was dedicated to activity measurements and to their comparison. For this purpose, participants were asked to make use of all the direct measurement techniques available in their laboratory in order to confirm or not the existence of possible biases specific to some measuring methods. In addition, this exercise offered the opportunity of improving the uncertainties of the gamma-ray intensities. Then, participants were asked, in the second part of the exercise, to carry out x-ray and gamma-ray intensity measurements.

In the following sections, the participants in the Euromet project 907 will be referred to as E907-n, where n is a serial number.

2. Decay scheme of antimony 124

¹²⁴Sb disintegrates through β^{-} transitions to excited levels in ¹²⁴Te. The total energy (*Q*) released to reach the ¹²⁴Te ground state level amounts to 2904 keV. The decay scheme is rather complex, as it includes more than 25 β^{-} transitions and about 70 γ transitions with energies spread between 148 keV and 2807 keV. However, only about a tenth of them have probabilities greater than 1 %. A simplified decay scheme is shown below. For a comprehensive decay scheme, see the "Decay Data Evaluation" part of this report.

In order to get comparable activity measurement results, participants were recommended to use the same nuclear decay data, *i.e.* the evaluated data published in Nuclear Data Sheets (NDS) 80,4 in 1997. This evaluation was based on γ -ray intensity values measured from 1984 to 1990. All these measurements were carried out relatively to the most intense γ -ray line of energy 602 keV, taken as the reference line.

In this exercise, the participants were asked to give their results in terms of both absolute values and relative values referred to the 602 keV line.

The 124 Sb half-life value adopted for the exercise was 60,20 (3) d.



Figure 2.1. Simplified decay scheme of ¹²⁴Sb

3. Programme and procedure

3.1. General description

Ampoules of ¹²⁴Sb solution were prepared and sent to the participants by the end of February 2007. They were accompanied with a description of the presently adopted decay scheme data, a short note describing the solution (see Appendix A1) and a questionnaire (see Appendix A2).

The results had to be sent to LNE-LNHB, with the completed questionnaire. The participants were requested to determine their uncertainties according to the rules expressed in the "Guide to the expression of uncertainty in measurement" (GUM, 1993). The results included the following items: - standardization of the ¹²⁴Sb solution, for which participating laboratories had to describe the

- standardization of the ¹²⁴Sb solution, for which participating laboratories had to describe the techniques using the questionnaire;

- half-life measurements, enabling, using various methods, to detect possible impurities in the solution; when available, these results could be used to determine the half-life value with a better accuracy;
- measurements of photon emission intensities for gamma- and K x-rays, for which participants were invited to give their results in terms of both absolute and relative values.

In addition, participants were requested to search for impurities in the solution at the receipt of the ampoule and two months later.

The last participant results were received at LNE-LNHB in March 2008.

Activity measurements and possible SIR submission

The participants were asked to report in the questionnaire the main characteristics of the equipment(s) used, the measurement method(s), the procedure(s) and main correction factors, the results and associated uncertainties, etc.

LNE-LNHB intended to send its own activity measurement result for submission to the SIR for establishing the traceability of the comparison exercise to the KCDB. Two ampoules of the solution were then delivered to the BIPM.

An ampoule of the same solution was sent to each participant (see next paragraph).

Then, each laboratory had the opportunity:

- to send an ampoule and the associated activity value to the BIPM for SIR submission and direct contribution to the KCRV (Key Comparison Reference value) of ¹²⁴Sb;
- to participate in the KCDB (Key Comparison Data Base) via the LNE-LNHB SIR result.

Measurements of x- and gamma photon emission intensities

The participants had to return, on the same questionnaire, a description of the main experimental details, of the peak analysis method and the results for the measured x- and γ -ray emissions. They were asked to give their results in terms of both absolute values and relative values referred to the 602 keV line.

It is worth noting that the absolute values measured by the participants depended on their activity measurement result, contrary to the relative ones only depending on their spectrometric measurements.

All the results were requested to be quoted at the reference time, March 1, 2007, 0 h UTC.

3.2. Properties of the distributed solution

The radioactive solution of ¹²⁴Sb was prepared by LNE-LNHB. Its activity concentration was of about 1,56 MBq.g⁻¹ (March 1, 2007, 12 h UTC). The solution was hydrochloric (2 mol.L⁻¹) with a SbCl₃ carrier concentration of 23μ g.g⁻¹. Each ampoule was filled with 5 ml of radioactive solution.

The participants were informed about a problem of antimony volatility observed by LNE-LNHB during the process of preparation of solid sources. Then the following advice was written on the form "Characteristics of the radioactive solution" sent to the participants.

"During the preparation of solid sources, a problem of volatility with ¹²⁴Sb was encountered. To try to solve it, we followed the procedure:

1) a weighed deposit of radioactive solution of ¹²⁴Sb was made for each solid source,

2) a drop with an equivalent mass of $2,5 \text{ mol.L}^{-1}$ NaOH was added on each deposit.

The purpose was to precipitate the element antimony as insoluble oxychlorides (SbOCl, ...)."

The homogeneity of the batch of ampoules was checked using an ionisation chamber. Ten ampoules (Volume 5mL) were prepared, each of them was measured 10 times, the experimental standard deviation of the measurement results was 0,02 %, which confirmed the excellent homogeneity of the batch.

3.3. Impurity checks

Purity tests of the initial solution were achieved using x- and gamma-ray spectrometry (coaxial HPGe ~ 100 cm³ and planar HPGe). ¹²⁵Sb, ¹²⁵Te^m and ¹²²Sb were more specifically scrutinized.

A first series of measurements was carried out using both detectors at the date of February 9, 2007. The acquisition time was 250 000 seconds. No impurities were found, and detection limits were determined:

- _
- ¹²⁵Sb: $A(^{125}Sb) / A(^{124}Sb) = 4 \ 10^{-4} \ Bq/Bq$, ¹²²Sb: $A(^{122}Sb) / A(^{124}Sb) = 2 \ 10^{-4} \ Bq/Bq$, ¹²⁵Te^m: $A(^{125}Te^m) / A(^{124}Sb) = 10^{-3} \ Bq/Bq$. -

A second series of measurements was undertaken on July 6, 2007, under the same conditions, which confirmed the previous measurements and the absence of gamma-emitting impurities.

When reported, these results were confirmed by the participants. Nevertheless, participant 5 identified ¹²⁵Sb as an impurity, whose quantity was found between the decision and detection limits, *i.e.* 0,03 kBq/g ($2 \ 10^{-5} \ Bq/Bq$) and 0,05 kBq/g ($3,5 \ 10^{-5} \ Bq/Bq$) respectively.

3.4. List of participating laboratories

EUROMET 9	007 : Sb		
Laboratory	Address	Persons involved in the	
		measurements	Contact email address
	Czech Metrological Institute	P. Dryak, J. Sochorovà, P.	pdryak@cmi.cz
CMI	Radiova 1,	Kovar, P. Auerbach	
	CZ 10200 Praha 10, Czech Republic		
	"Horia Hulubei" National Institute	A. Luca (γ-ray spectrometry),	aluca@ifin.nipne.ro
IEIN UU	of Physics and Nuclear Engineering	M. Sahagia and A.C. Wätjen,	msahagia@ifin.nipne.ro
11,114-1111	P.O. Box MG-6	$(4\pi\beta-\gamma CC \text{ and sources})$	cristina.watjen@nipne.ro
	Bucharest, Romania	preparation)	
IRA-	Rue du Grand-Pré 1 CH - 1007	C. Bailat, Y. Nedjadi, P.	Claude.bailat@chuv.ch
METAS	Lausanne	String	
	Laboratorio Nacional de Metrologia	A. Iwahara, M.C.M de	delgado@ird.gov.br
IRD-	das Radiaçoes Ionizantes	Oliveira, J.U. Delgado, C.J.	
LNMRI	Av. Salvador Allende s/n° Recreio	da Silva	
	Rio de Janeiro, Brazil		
	Institute for Reference Material and	T. Altzitzoglou, S. Pommé,	Timotheos.altzitzoglou@cec.e
	Measurements	R. Van Ammel, G. Sibbens,	uropa.eu,
IRMM	Retieseweg 111, 2440 Geel,	J. Paepen, J. Camps	stefaan.pomme@ec.europa.eu
	Belgium		
	Laboratoire National Hanri	C. Bobin $(4\pi\beta - \gamma CC)$	
		M.C. Lépy (γ- spectrometry)	mm.be@cea.f
LNE-LNHB	CEA	M.N. Amiot (IC)	christophe.bobin@cea.fr
	01101 Gif-sur-Vyetta Caday France	T. Branger (sources	marie-christine.lepy@cea.fr
	51151 Oll-Sul- I velie Ceuex, Flance	preparation)	
	National Physical Laboratory	L. Johansson, L. Keightley,	lena.johansson@npl.co.uk;
NDI	Queens Road	A. Stroak	lynsey.keightley@npl.co.uk
INTL	Teddington, Middlesex, TW11		
	OLW, U.K.		
	Dhysikalisch Tachnische	K. Kossert (LSC)	Karsten.Kossert@ptb.de
	Pundosonstalt	O. Nähle $(4\pi\beta - \gamma CC)$	
PTB	Bundesallae 100	O. Ott, R. Dersch (γ-	
	28 116 Braunschweig Germany	spectrometry)	
	56 110 Draunsenweig, Germany		

\mathbf{L} able \mathbf{S}_{11} , \mathbf{L}_{151} of participating

4. Activity measurements

4.1. Methods

Following the recommendations, most participants used several methods to measure the activity of ¹²⁴Sb. The eight participants thus sent a total of 18 individual measurement results.

They were invited to give their individual results for each method they used and their adopted final value. When a participant used several methods, his final value was either the average of all the different results obtained or of a selection of some of them, or the result of the method in which he was the most confident. Some results were thus eliminated by some participants, because of an unexplained discrepancy when compared to others.

The details are given in Table 4.1. Each line lists a measurement method, the number of participants who used it, the identification numbers of those participants, and the identification numbers of the participants who included that method in their final value.

It is worth noting that all participants carried out measurements with the $4\pi\beta(PC,PPC)-\gamma$ coincidence or anti-coincidence method and included it in their final value.

Method	Number of results	Participants using the method (<i>serial number</i>)	Participants using the method in their final result (<i>serial number</i>)
$4\pi \beta$ (PC,PPC)- γ coincidences or anti-	8	1-2-3-4-5-6-7-8	1-2-3-4-5-6-7-8
coincidences			
$4\pi \beta(LS)-\gamma$ coincidences	2	2-3	2
CIEMAT/NIST liquid scintillation	2	5-8	5
counting			
$4\pi\gamma$ NaI(Tl) counting	3	1-3-8	8
$4\pi\gamma$ CsI(Tl) counting	1	8	8
Well-type ionization chamber	2	1-3	-

Table 4.1. Activity measurement methods used by participants

4.2. Source preparation

In addition to the pilot laboratory, three participants gave some information about antimony volatility related to source preparation.

Participant E907-1 reported experimental evidence that the recommended procedure was not sufficient to completely avoid Sb emanation, possibly because of the difficulty of fully homogenizing $SbCl_3$ and NaOH by drop addition onto the source support. Instead, they diluted the original solution with an aqueous solution containing NaOH and chelating agent EDTA, after which no emanation could be detected. Such a problem was not encountered by the pilot laboratory when testing the procedure.

Participant E907-5 observed a bad quality of samples (low efficiency and changes of the appearance of samples) when applying the recommended procedure. Thus, many attempts were made to improve the sample preparation. Some techniques showed evidence of antimony evaporation during source preparation. Finally a technique was applied in which the original solution and tartaric acid were used in a light-protected metal-free environment. No antimony evaporation was observed applying that technique.

Participant E907-6 followed the recommended procedure: after preparing the sources by drop deposition of an aliquot of the original solution onto VYNS films, a drop of NaOH in concentration 2,5 mol/L of approximate equal mass was added. The effect was the reduction of the beta detection efficiency down to 0,70 - 0,80. These sources were used in reporting the result. Sources with lower NaOH concentrations *i.e.* 0,25 mol/L and 0,025 mol/L were also prepared. The effect was to obtain better efficiencies, 0,85 and 0,92 respectively, but the loss of mass was more than 0,5 %; consequently, these results were not used.

4.3. Results

The results obtained for all methods are reported in Table 4.2 and Figure 4.1. This leads to an unweighted mean value of 1451 (12) kBq/g.

The following sections focus on the results obtained for each method in more details.



Figure 4.1. Measurement results for all methods

Participant	Methods	Activity	Standard
n°		concentration	uncertainty
		(kBq/g)	(kBq/g)
1	$4\pi(PC)\beta-\gamma$	1443,6	4,7
	4πγ(NaI)	1440,2	4,8
	Ionization chamber	1429,6	9,2
2	$4\pi(PC)\beta-\gamma$	1457,6	5,5
	$4\pi(LS)\beta-\gamma$	1455,6	3,3
3	$4\pi(PC)\beta-\gamma$	1452,5	4,1
	$4\pi(LS)\beta-\gamma$	1455,5	3,7
	$4\pi\gamma$ (NaI)	1438,4	2,8
	Ionization chamber	1442	10
4	$4\pi(PC)\beta-\gamma$	1451,1	5,4
5	$4\pi(PC)\beta-\gamma$	1461,9	10.1
	Ciemat/Nist(LS)	1444,6	5,4
6	$4\pi(PC)\beta-\gamma$	1466,4	8,1
7	$4\pi(PC)\beta-\gamma$	1437	4,2
8	$4\pi(PC)\beta-\gamma$	1465	10
	$4\pi\gamma$ (NaI)	1470	22
	4π (CsI)	1465	10
	Ciemat/Nist(LS)	1438	7

Table 4.2.Results	for all t	the methods	used by	participants
	,		· · ·	1 1

4π **b**-gcoincidence or anti-coincidence measurements

 $4\pi \beta$ - γ coincidence or anti-coincidence measurements using a proportional counter in the beta channel were carried out by all participants. In addition, two participants (2 and 3) made measurements using a liquid scintillation counter in that channel. The results are presented in Figure 4.2. In that figure, the mean value (1454,4 kBq/g) is the unweighted mean value of the 4π (PC) β - γ results and the standard deviation (10,4 kBq/g or 0,72 %) corresponds to the distribution of those results. Only one participant in eight is an outlier within the one standard uncertainty limits. The two 4π (LS) β - γ results are shown for comparison purpose; they look quite consistent with other results.



Figure 4.2. **4p b-g** coincidence or anti-coincidence measurement results

The participant laboratories which carried out both coincidence and anticoincidence measurements with a proportional counter and a liquid scintillation detector in the beta channel got very close results. This observation is of particular interest since there was an identified problem of antimony volatility which should have a different impact on solid point sources and liquid scintillation sources. This effect could be a potential source of dispersion of the results, independently from any other origin, and thus cause confusion in comparing the results of the different measurement methods.

The maximum beta detection efficiency achieved as reported by participants spread from 70 % to 98 % for solid point sources, and from 90 % to 98 % for liquid scintillation sources. Participant 3 reported that the adding of NaOH lowered by about 5 to 10 % the beta detection efficiency of solid sources due to additional self absorption.

In order to control the antimony evaporation observed during the production process of his solid sources, participant 5 measured repeatedly some sources over a period of 8 months. An increase of the activity result was observed which could not be explained by the detected impurity (125 Sb) alone, which was measured at the ampoule reception with a detection limit of 0,05 kBq/g (see section 3.3). So, that participant introduced two uncertainty components, one of 0,20 % for impurities, and one of 0,58 % for source stability, due to possible changes of source properties due to environmental influences, which is predominant in his uncertainty budget. To our knowledge, no other participant made that control over a similar period of time.

From the $4\pi\beta\gamma$ coincidence or anti-coincidence measurements, it can be concluded that this set of results is consistent.

4π**-g**NaI and **4p** CsI measurements

Three participants (1, 3 and 8) carried out 4π - γ measurements using a well-type NaI(Tl) crystal detector and one (8) using also a 2×2 π CsI crystal detector. The results are presented in Figure 4.3. In that figure, the mean value (1450 kBq/g) is the unweighted mean of the 4π - γ NaI results and the standard deviation (18 kBq/g, or 1,22 %) corresponds to the distribution of those results. The 4π CsI result is added for comparison purpose.



Figure 4.3. 4p-g NaI and 4p CsI measurement results

The reported maximum ¹²⁴Sb detection efficiencies were respectively 0,895, 0,956 and 0,91 for participants 1, 3 and 8. For participant 8, the predominant uncertainty component came from efficiency calculation (1,5%); a substantial component of this uncertainty concerned the contribution of high-energy beta radiation which is particularly abundant in the decay of ¹²⁴Sb, when working out the efficiency curve by using a Monte Carlo code. The reported detection efficiency of the 4π CsI detector was 1, with an uncertainty of efficiency calculation of 0,6%. It can be noted that since that efficiency value is maximal, any change of that quantity can only be a reduction and then leads to an increase of the activity concentration.

The results obtained surround those obtained with coincidence counting. In spite of their high value, the results of participant 8 are consistent with coincidence results, considering their large uncertainties. The results of participants 1 and 3 are in agreement but are low compared to coincidence measurement results. Participant 1 carried out measurements both with solid point sources and liquid sources in sealed glass scintillation vials filled with 14 mL Ultima Gold. The results reported for the two types of sources were in good agreement, demonstrating no loss of activity. Considering its uncertainty, the value of participant 3 is inconsistent with his coincidence result and the mean value of all coincidence results.

Liquid scintillation (Ciemat-Nist) measurements

Two participants (5 and 8) carried out measurements using the Ciemat-Nist method. Their results are respectively 1444,6 (54) kBq/g and 1438 (7) kBq/g. These results are consistent with each other, but are low compared to those obtained by coincidence counting (see Fig. 4.1). However, the number of results is too small to get a significant conclusion about a possible trend resulting from the method itself. Participant 5 gave some details on the efficiency calculation, taking into account the seven most dominant beta branches and twenty seven gamma transitions, and cascades with one beta and up to four gammas in coincidence. Decay data were successively taken from "Nucléide-CD" (2005) and from Nuclear Data Sheets (NDS) 80,4 (1997), leading to an activity larger by about 0,055 % with "Nucléide". Participant 8

included the 13 most dominant beta decay branches and twenty five gamma-ray transitions and cascades with one beta and up to four gamma-rays in coincidence, with decay data taken from NDS (1997).

As for coincidence measurements, participant 5 noticed evidence of a significant increase of activity over a period of ten months of measurements, then he added two uncertainty components, one of 0,20 % for impurities, and one of 0,05 % for sample stability. Participant 8 measured his sources 17 times over a period of seven months. Some of those sources needed 5-6 days after preparation to become stable. Then they were mostly stable (to better than 0,1 - 0,2 %) for a period of 75 days, after which the count rate started to increase. An uncertainty component of 0,2 % for sample stability was added.

Well-type ionization chamber measurements

Two participants (1 and 3) carried out measurements of the ¹²⁴Sb solution in glass ampoules using a pressurized well-type ionization chamber. The chamber of participant 1 (Centronics IG11/A20) was filled with argon at 2 MPa, and the chamber of participant 3 (Vinten 671) with nitrogen at 1 MPa. Both participants determined the response of their instrument by computation from semi-empirical efficiency curves, obtained from Monte Carlo calculation and primary standard solutions. Participant 3 pointed out that the response to beta particles was obtained from an experimental curve obtained with pure beta-emitters; the contribution of beta particles to the response of the chamber was less than 1 % for ¹²⁴Sb.

The results are 1429,6 (92) kBq/g and 1442 (10) kBq/g for participants 1 and 3, respectively. Participant 1 gave in addition a result obtained using the data of "NDS" instead of "Nucléide" (1422,8 (62) kBq/g), the difference is of about 0,5 %.

These two results are consistent within the uncertainty limits and, for both participants, they are consistent with their respective coincidence measurement values (see Table 4.2 and Fig. 4.1).

Final results

Each participant was asked to select his final activity concentration value from his individual results. This value was considered as his reference value for the activity comparison, and the one to be used by him to derive absolute photon intensities. Those reference values are presented in Table 4.3 and Figure 4.4.

Participant n°	Procedure / choice of Activity participant concentrati		Relative standard	Comments
		(kBq/g)	uncertainty (%)	
1	$4\pi(PC)\beta-\gamma$	1443,6	0,32	$4\pi\gamma$ and ionization
				chamber not selected
2	Weighted mean of	1456,1	0,38	Adoption of uncertainty
	$4\pi(\text{PC})\beta-\gamma + 4\pi(\text{LS})\beta-\gamma$			budget of $4\pi(PC)\beta-\gamma$
3	$4\pi(PC)\beta-\gamma$	1452,5	0,28	$4\pi(LS)\beta-\gamma$ and
				ionization chamber not
				selected;
				$4\pi\gamma$ considered as an
				outlier
4	$4\pi(PC)\beta-\gamma$	1451,1	0,37	Single method used
5	Unweighted mean of	1453,3	0,60	Uncertainty enlarged to
	$4\pi(PC)\beta-\gamma$ + Ciemat/Nist(LS)			take account of the
				discrepancy between
				the two methods
6	$4\pi(PC)\beta-\gamma$	1466,4	0,55	Single method used
7	$4\pi(PC)\beta-\gamma$	1437	0,29	Single method used
8	Weighted mean	1465	0,48	Ciemat/Nist LS not
	(internal variance) of			selected, interpreted as
	$4\pi(\text{PC})\beta-\gamma + 4\pi\gamma(\text{NaI}) + 4\pi(\text{CsI})$			an outlier

Table 4.3. Final results of participants

The unweighted mean value obtained with those results is 1453 (10) kBq/g, the associated relative standard deviation of the distribution is 0,68 %. This value remains close to the general mean value (1451 (12) kBq/g) and to the mean value of coincidence counting results (1454,4 (10,4) kBq/g), this latter method being the predominant one for that exercise.

Participant 1 selected only his $4\pi(PC)\beta-\gamma$ coincidence result.

Participant 3 did the same choice. For supporting that choice, he considered that the very good agreement obtained using the coincidence method with two different types of sources, solid and liquid, should remove doubts about source stability. The $4\pi\gamma$ result was discarded because of unexplained discrepancy at this stage. A possible source of error could be the underestimation of the contribution of high-energy beta particles which is particularly abundant in the decay of ¹²⁴Sb, when working out the efficiency curve by using a Monte Carlo code.

Participant 5 observed some inconsistency between his $4\pi(PC)\beta-\gamma$ coincidence and Ciemat/Nist LS results. Suspecting unknown systematic effects, he adopted the unweighted mean of the two values, with an enlarged uncertainty.

Participant 8 adopted the weighted mean of three methods, excluding the discrepant Ciemat/Nist LS result. No explanation for this discrepancy was found at the deadline of the exercise.



Figure 4.4. Final results given by participants

4.4 Findings for the activity measurements

In spite of some lack of detailed information given in some questionnaires, several points come to light:

- When preparing solid sources, several laboratories noticed antimony volatility and tried to deal with it.

- The results of the $4\pi\beta$ - γ coincidence measurements are consistent (except one outlier). There was a very good agreement between measurements made with solid point sources and liquid scintillation sources.

- The results obtained by means of other methods $(4\pi-\gamma, \text{ ionisation chambers and LSC})$ are in general consistent with each other but lower than the coincidence results, except for participant 8 with larger uncertainties (Fig. 4.3). Participant 3 mentioned for his $4\pi-\gamma$ result a possible underestimation of high-energy beta emissions when calculating the detector response.

- When simulations by means of Monte Carlo calculations as well as LSC efficiency calculations were done using various sets of decay data, no significant differences in the computed results were observed.

5. Half-life measurement

Participant 8 carried out the half-life measurement of 124 Sb by means of a well-type ionization chamber. The decay of the source was followed during 240 days, *i.e.* about four half-lives. The chamber main characteristics are described in Table 5.1.

IONIZATION CHAMBER DESCRIPTION					
Producer and type	Centronic IG12/A20				
Shape (U-shape, cylindrical)	cylindrical, re-entrant tube				
Gas nature	Argon				
Gas pressure	2 MPa				
Well diameter	50,8 mm				
Thickness of the wall of the well	0,8 mm				
IC external diameter	184 mm				
IC height	427 mm				
Material of the wall of the well	Steel				
Material of the electrodes	Aluminium				
Thickness of lead shielding	50 mm				
CURRENT	MEASUREMENT				
Reference of the electrometer	Keithley 6517A in voltage mode, using				
	NPL current integrator and IRMM stable				
	air capacitor (shielded in 50 mm lead)				
Current measurements relative to a	No				
long-lived reference source (Y/N)					
MEASUI	REMENT DATA				
Typical current pA	222 pA at reference date				
Background current pA	0,05 pA				
IC response pA/MBq	N.A.				
Typical integration time	28 s at the beginning				
	334 s at the end of measurement campaign				
Number of measurements	Source: 10138				
	Background: 5811				

Table 5.1. Details on ionization chamber and measurement conditions

Data treatment

One raw data point corresponds to one charge cycle of the feedback capacitor from 0 to 9 V. This raw data is corrected for background, impurities and decay during the integration time, to obtain the activity of the source at the beginning of the charge cycle. The background is considered constant over the measurement campaign, and the decay correction is applied using a preliminary half-life of 60,20 (3) days. The data are limited to 240 days, after which the uncertainty of the background becomes too important. Raw data points due to instrument malfunctioning are also rejected, leaving 10138 points. These are averaged, day-by-day, yielding 46 data points for fitting. A least-squares fit is applied to determine the half-life, weighting the data points with the inverse of the variances (?²=0,955). Residuals after the fit are shown in Figure 5.1.



Figure 5.1. Residuals from the least-square fitting of experimental points

Uncertainty budget

Table 5.2 lists all the uncertainty components which were taken into account to calculate the combined standard uncertainty of the half-life. The calculation of the individual uncertainty contributions is done using the variance propagation formula:

$$\frac{\mathbf{s}_{T_{1/2}}}{T_{1/2}} \approx \frac{2}{\mathbf{l}T} \sqrt{\frac{2}{n+1}} \frac{\mathbf{s}_A}{A}$$

in which *T* is the duration of the measurement campaign (240 days), and *n* corresponds with the frequency of the uncertainty component (n = 1 for low-frequency components).

Process	Uncertainty on A	п	Propagation factor	Uncertainty of $T_{1/2}$ (relative 10^{-4})
	(relative ~ 10 ⁻⁴)			``````````````````````````````````````
Background, systematic	1,9	1	0,78	1,5
component				
Medium term instability	1,0	1	0,78	0,78
Ampoule repositioning	2,1	56	0,15	0,32
Long term instability	0,40	1	0,78	0,31
Background, random component	9,6	5811	0,015	0,14
Impurities	0,18	1	0,78	0,14
High frequency instability	2,4	10138	0,011	0,026
Statistical nature of decay	0,94	10138	0,011	0,010
Time measurement	0,36	10138	0,011	0,004
Preliminary decay correction	$4,5 \times 10^{-6}$	10138	0,011	5×10^{-8}
	Standard comb	oined unce	rtainty ($\times 10^{-4}$)	1,8

Table 5.2. Uncertainty budget of the ¹²⁴Sb half-life measurement

The final result is **60,212** (**11**) **d**. This value is in very good agreement with the one recommended for that exercise.

6. Photon emission intensities determination

Six participants carried out measurements of the γ -ray emission intensities. As recommended, they sent their results both in absolute and relative values, the most intense line (602 keV) being the reference. Only three of those participants measured x-ray emission intensities.

6.1. Experimental arrangements

The γ -ray emission intensities were measured by the six participants with a coaxial-type germanium detector (HPGe). The x-ray emission intensities were measured by the three participants with a planar HPGe. The main features of the experimental arrangements are listed in Table A.1 "Instrumentation - general characteristics, geometry arrangements" of Appendix A3.

6.2. Efficiency calibration and analysis procedure

The list of radionuclides and their x- and γ - rays which were used as standards by participants to determine the full-peak efficiency calibration curve of their detectors are detailed in Table A.2 (Appendix A3). Since ¹²⁴Sb emits gamma rays in a large energy range from 148 keV to 2800 keV, the "traditional" efficiency curve of the HPGe gamma detectors had to be extended.

Participants 2 and 6 used point sources emitting gamma photons up to 1408 keV. Participant 6 thus sent results for gamma lines of energies up to 1918 keV only.

Participants 3 and 5 used point sources emitting up to 3 MeV (⁵⁶Co) and 2,7 MeV (⁸⁸Y) respectively.

Participants 7 and 8 used a Monte Carlo calculation method combined with experimental results from point sources, up to 2615 keV (²³²Th (²⁰⁸Tl)) and 1332 keV (⁶⁰Co) respectively. It is noteworthy that participant 7 used a Monte Carlo simulation method to work out the

It is noteworthy that participant 7 used a Monte Carlo simulation method to work out the efficiency curve of his detector and claimed an efficiency uncertainty in the range 0,15 - 0,25 %. This is, at least, two times lower than those of the other participants.

A short description of the efficiency interpolation peak analysis method is given in Table A.3 (Appendix A3). Most participants carried out peak analyses using a software provided by a manufacturer and, often, they finalised or made an analysis more specific with their own calculations. Participants 2 and 7 did not give clear explanations of the peak analysis method they used.

The main applied corrections were due to the coincidence summing effect, pile-up and dead-time losses. Some numerical examples are given in Tables 6.1 and additional details in Table A.4 (Appendix A3). In most cases, the calibration of the spectrometer and the measurement of Sb-124 sources were carried out with the same geometry arrangement, so that the absorption and self-absorption effects were included in the efficiency calibration.

Code number	E907-3		E907-5		E907-7		E907-8	
Detector	HP-Ge (10	0 cm^3)	HP-Ge (127 cm^3)		HP-Ge (180 cm^3)		HP-Ge (360 cm^3)	
Source-to-detector	12,5 cm		18,5 cm		25 cm		12 cm	
distance								
Gamma ray	Summing	Pile-up	Summing	Pile-up	Summing	Pile-up	Summing	Pile-up
	effect		effect	& dead	effect		effect	& dead
				time				time
602,7 keV	1,02	-	1,003	1,01	1,003	1	1,0014	1,022
1690,9 keV	-	-	1,004	1,01	1,0035	1	1,0066	1,022
2293,7 kev	0,18	-	0,333	1,01	0,35	1	1	1,022

Table 6.1. Main corrections applied to photon intensity measurements

6.3. Uncertainties

The participants were asked to give detailed uncertainty budgets. The main components are summarized in tables below for the two most intense emissions taken as examples.

a) 602-keV photon emission

The main components of the uncertainty come from the peak detection efficiency and from the activity value.

The combined standard uncertainties determined by the participants are in the range 0,3 % - 1,5 %.

Code Number	Relative standard uncertainties (%)								
	Counting statistics	Peak area	Efficiency	Correction factor	Others	Source activity	Combined standard uncertainty		
2	0,3	0,1	0,6	< 0,2	~ 0,2	0,4	0,72		
3	0,6		0,5	0,1	0,3	0,28	0,87		
5	0,034	0,061	0,20	0,038		0,65	0,68		
6	0,22		0,98			0,55	1,14		
7	0,1	0,2	0,07	0,05	0,01	0,25	0,3		
8	0,05	0,1	1,2	0,7	0,3	0,5	1,5		

Table 6.2. Uncertainty budget of the photon emission intensity results for the 602-keV line, in %.

b) 1690-keV photon emission

The following table gives an example for a γ - ray emission intensity given both in absolute and relative values with respective uncertainties.

The standard combined uncertainties determined by the participants are in the range 0,4 % - 2,4 %.

Code Number	Relative standard uncertainties (%)									
	Counting statistics	Peak area	Efficiency	Correction factor	Others	Combined standard uncertainty on relative value	Source activity	Combined standard uncertainty		
2	0,3	0,1	2	< 0,2	~ 0,2	2,4	0,4	2,4		
3	0,6		0,5	0,1	0,3	1,2	0,28	0,87		
5	0,075	0,061	0,36	0,038		0,43	0,65	0,75		
6	0,5		1,6			1,9	0,55	1,9		
7	0,1	0,2	0,17	0,05	0,01	0,37	0,25	0,4		
8	0,01		1,6	0,8	0,3	1,13	0,5	1,9		

Table 6.3. Uncertainty budget of the photon emission intensity results for the 1690-keV line, in %.

c) X-ray emissions

The combined standard uncertainties reported by the participants are in the range 5 - 20 %, the main component coming from the detector efficiency.

Code Number	Relative standard uncertainties (%)								
	Counting statistics	Peak area	Efficiency	Correction factor	Others + weighting	Source activity	Combined standard uncertainty		
2							21		
3	0,56		0,8			0,28	1,3		
8	0,13	0,28	2,5	0,6	0,3	0,5	4,3		

Table 6.4. Uncertainty budget of the photon emission intensity results for the Ka x-ray emissions, in %.

6.4. Statistical treatment of data

When the same measurement technique was applied by several participants, a weighted average, a_w , was calculated using the combined uncertainties of the individual values as weights.

For *n* independent values a_i , each with a combined standard uncertainty u_{ci} , a weight p_i proportional to the inverse of the square of the individual standard uncertainty u_{ci} can be assigned to each value.

$$a_w = \frac{\sum_{i=1}^n p_i a_i}{\sum_{i=1}^n p_i} \quad ,$$

where the weights are $p_i = 1/u_{c_i}^2$.

Internal and external uncertainties can be assigned to the weighted mean value. The square root of the internal variance $\mathbf{s}^{2}_{int}(a_{w})$ is the expected uncertainty of the mean, based on the individual *a priori* variances $u_{c_{i}}^{2}$ and the application of the propagation law of uncertainty.

$$\mathbf{s}_{int}(a_w) = \left[\sum_{i} (1/u_{c_i}^2)\right]^{-1/2}$$

The external uncertainty is given by the formula,

$$\boldsymbol{s}_{ext}(a_{w}) = \left[\frac{\sum_{i} (a_{i} - a_{w})^{2} / u_{c_{i}}^{2}}{(n-1)\sum_{i} 1 / u_{c_{i}}^{2}}\right]^{\frac{1}{2}}$$

The external variance $\mathbf{s}_{ext}^2(a_w)$ takes the scatter of the data into account, and is based on the amount by which each result a_i deviates from the mean, measured as a fraction of each given uncertainty u_{c_i} .

A measure of the consistency of the data is given by the ratio,

$$\boldsymbol{s}_{ext} / \boldsymbol{s}_{int} = \sqrt{\boldsymbol{c}^2 / (n-1)}$$
.

If this ratio is significantly greater than unity, at least one of the input data surely has an underestimated standard uncertainty value u_{c_i} which should be increased.

6.5. Results of g ray intensity measurements

The participants were asked to measure all the photon emission intensities, the weak as well as the intense lines, in order to improve the decay scheme data, some of the weakest lines having been observed in the past without any confirmation since.

All the measured values of photon emission intensities published in the literature are given relative to the 602-keV line; this is why the participants were requested to send their results both in absolute values, *i.e.* per 100 % Sb-124 disintegrations and in relative values, the 602-keV γ -ray line being taken as the reference line. This was thus intended to make possible the comparison of the Euromet 907 results with the other available results and to give a better statistical treatment of the data when working out the decay scheme evaluation.

In general, for the most intense lines, the results sent by the participants were in good agreement. The set of data, for the two most intense lines are detailed below. Figures 6.1 and 6.2 give the diagram for the 602- and 1690-keV lines. For a comprehensive statistical study of all sets of data, see Part B – ¹²⁴Sb Decay Data Evaluation.

a) 602-keV photon-emission absolute intensity

Despite the relatively large dispersion of activity measurement results, the set of photon emission absolute intensity data is consistent, with a reduced c^2 of 0,1.

The value of participant 6 was found to be an outlier, based on Chauvenet's criterion.

The value given by participant 7 amounts to 58 % of the total statistical weight, because the associated uncertainty is significantly lower than the uncertainties given by the other participants, due to an uncertainty of the calibration efficiency of 0,07 % and an uncertainty on the activity value of 0,25 %; both seem optimistic.

Nevertheless, since the single (UWM) and weighted means (WM) are close to each other and the set of data is consistent, the weighted mean is adopted as reference.

			_
Participant	I_{abs602} (%)	<i>u</i> _c	
E907-2	97,5	0,7	
E907-3	97,8	0,9	
E907- 5	97,6	0,7	
E907- 6	91	1	
E907-7	97,84	0,34	
E907-8	98,1	1,5	
Chi2	0,1		$c^{2}/(n-1)$
Chi2 criterion	3,3		
UWM	97,787		Unweighted mean
WM	97,769		Weighted mean
$u_{\rm c}$ (int)	0,26		Internal uncertainty
$u_{\rm c}$ (ext)	0,07		External uncertainty
Adopted	97.77	0.26	

Table 6.5. Results for the 602-keV photon-emission absolute intensity

b) 1690-keV photon-emission absolute intensity

The second line in terms of emission intensity is the 1690 keV γ -ray. The spread of the values is larger than for the 602 keV line. However, ignoring the value of participant 6, the set of the five remaining data is consistent with a reduced c^2 of 2,2. The value of participant 7 amounts to 65 % of the total statistical weight.

Participant	$I_{\rm abs1690}(\%)$	<i>u</i> _c
E907- 2	45,56	1,09
E907- 3	47,04	0,40
E907- 5	47,10	0,35
E907- 6	44,70	0,77
E907- 7	47,65	0,18
E907- 8	46,03	0,87
Chi2	2,2	
Chi2 criterion	3,3	
UWM	46,676	
WM	47,392	
$u_{\rm c}$ (int)	0,15	
$u_{\rm c}$ (ext)	0,22	
Adopted	47,39	0,22

Table 6.6. Results for the 1690-keV photon-emission absolute intensity

c) Diagrams for the 602-, 722-, 1690-, and 2090-keV lines

From Figures 6.1, 6.2, 6.3 and 6.4, it appears that the spread of the results is greater for the gamma lines of high energy, namely 1690 and 2090 keV, than for those of lower energy. This could be due to difficulties in the detector efficiency calibration at high energy; as already mentioned, only three participants (3, 5, 7) used standard point sources with energies greater than 1400 keV.



Figure 6.1. Results for the 602-keV photon-emission absolute intensity



Figure 6.2. Results for the 722-keV photon-emission absolute intensity



Figure 6.3. Results for the 1690-keV photon-emission absolute intensity



Figure 6.4. Results for the 2090-keV photon-emission absolute intensity

6.6. Comparison with previous published data

As presented in the introduction, a possible explanation for the discrepancy in the activity results obtained by means of various methods could be due to problems in the decay scheme data, this is why a part of this exercise was dedicated to photon emission intensity measurements. There are two ways for working out the Sb-124 decay scheme:

- the first one is to use absolute γ -emission intensity values; this requires the knowledge of the nuclide activity concentration of the solution;
- the second one results from the fact that the sum of all the transitions (including the 602-keV transition) arriving to the ground state level of the daughter nuclide must be equal to 100 % of disintegrations of the parent nuclide.

In the case of Sb-124, the situation is simplified since no β transition populating the ground state level of tellurium 124 is expected; then the sum of the gamma transition probabilities with energies 2807, 2693, 2681, 2455, 2323, 2294, 2182, 2039, 1657, 1325, 602 keV which populate it must be equal to 1. This can be expressed by:

$$\sum_{i} I_{g_{i}} \times I_{abs602} (1 + a_{T_{i}}) = 1$$

where I_{g} is the relative intensity of a gamma emission, α_{Ti} is the total conversion coefficient of the corresponding gamma transition, I_{abs602} the absolute value of the 602-keV gamma line.

 I_{abs602} , the absolute value of the 602-keV gamma line, is then deduced from the measured relative I_{gi} values and the total conversion coefficients α_{Ti} ,

$$I_{abs602} = \frac{1}{\sum_{i} I_{g_{i}} [1 + a_{T_{i}}]}$$

The same statistical analysis, as previously, was done with larger sets of data, with the relative photon intensity results sent by the participants and including other results published in available journals. The reference line is $I\gamma_{602} = 100$.

To use the formula above, the total conversion coefficient α_{Ti} of each gamma transition is required. The α_{Ti} coefficients were interpolated from the Band's tables (2002Ba85) using the program BrIcc with the "frozen orbital approximation" (Kibédi *et al.* 2008Ki07). All transitions with a measured polarity are of E2 type and the energies of the involved transitions are relatively high, then the internal conversion coefficient values derived from theoretical calculations can be considered as very reliable.

As an example, details are given in table 6.7 hereafter for the 1325 keV gamma transition which reaches the ground state level. A relative γ -emission intensity of 1,623 (7) is adopted (see Part B – ¹²⁴Sb Decay Data Evaluation, for details). All the relative I_{γ} values were calculated in that way.

From the α_{Ti} and the measured relative $I_{\gamma i}$ values, the absolute emission intensity of the 602-keV gamma-ray is then derived to be 97,775 (20) %. This value is in full agreement with the value of 97,777 (26) % obtained from absolute measurements.

	1325 keV		
	$I_g(\%)$	<i>u</i> _c	
E907- 2	1,637	0,033	
E907- 3	1,599	0,027	
<i>E907- 5</i>	1,603	0,016	
E907- 6	1,582	0,137	
E907- 7	1,621	0,007	
E907- 8	^(o) 1,768	0,040	
Patil (2006Pa16)	1,707	0,026	
Goswamy (1993Go10)	1,61	0,03	
Jianming (1988Yo05)	1,645	0,028	
Mardirosian (1984Ma13)	1,69	0,29	
Iwata (1984Iw03)	1,584	0,023	
Johnson (1974Jo03)	1,67	0,04	
Meyer (1990Me15)	1,66	0,04	
Sharma (1979Sh08)	1,71	0,04	
Chi2	2,0		
Chi2 criterion	2,2		
UWM	1,6399		
WM	1,6233		
$u_{\rm c}$ (int)	0,0051		
$u_{\rm c}$ (ext)	0,0073		
Adopted	1,623	0,007	

Table 6.7. Results for the 1325-keV photon-emission relative intensity

The most intense absolute γ -ray emission intensities obtained from the absolute measurements made by the six participants in Euromet 907 are compared, in the following table, with those obtained from the sets of 14 relative values (same participants + 8 published data) and conversion coefficients (ICC). The agreement is quite good.

gray energy (keV)	Absolute intensity from absolute measurements	Absolute intensity from relative measurements and ICC
602	97,77 (26)	97,775 (20)
645	7,414 (21)	7,422 (15)
709	1,3635 (43)	1,363 (5)
713	2,269 (11)	2,273 (7)
722	10,712 (31)	10,708 (22)
968	1,880 (6)	1,887 (10)
1045	1,835 (6)	1,852 (14)
1325	1,583 (6)	1,587 (7)
1368	2,615 (9)	2,620 (8)
1690	47,39 (22)	47,46 (19)
2090	5,491 (26)	5,493 (24)

 Table 6.8. Values of the absolute intensities obtained from absolute measurements and from relative measurements and ICC values

From the comparison above, it is concluded that the absolute γ -ray emission intensities can be considered as reliable and the decay scheme consistent. (see also further comments in Part B - ¹²⁴Sb Decay Data Evaluation). Then the problems in activity measurements should not be due to the decay scheme data.

6.7. Euromet 907, x-ray intensity results

Three laboratories (2, 3, 8) carried out specific measurements of the x-ray emission intensities, using a HPGe planar detector (Table A.1, Appendix A3). Participants 3 and 8 used radioactive sources with low-energy photon radiation emissions, *e.g.* ²⁴¹Am (Table A.2, Appendix A3), to calibrate their spectrometer in the x-ray energy range.

	E90	7-2	E907	7-3	E90	7-8	Calcu	lated
Energy (keV)	I %	<i>u</i> _c	I %	<i>u</i> _c	I %	<i>u</i> _c	I %	<i>u</i> _c
27,2 (Kα ₂)			0,128	0,002	0,130	0,003	0,1252	0,0018
27,5 (Kα ₁)			0,264	0,004	0,230	0,006	0,233	0,003
30,9 (Kβ' ₂)			0,068	0001	0,063	0,002	0,0667	0,0012
31,7 (Kβ' ₁)			0,017	0,0005	0,0136	0,0006	0,0145	0,0005
Κα	0,35	0,07	0,392	0,0045	0,359	0,007	0,358	0,0035
Κβ	0,087	0,018	0,085	0,0011	0,076	0,0018	0,081	0,0013
K x Total	0,437	0,072	0,476	0,005	0,436	0,007	0,439	0,004

Table 6.9. X-ray emission intensity results, in % of disintegrations

The total K x-ray values obtained by participants 2 and 8 are in agreement, the value given by participant 3 deviates from the others, mainly due to a $K\alpha_1$ higher intensity value.

The calculated x-ray emission intensities were derived from the decay scheme data, *i.e.* from the evaluated γ -ray absolute emission intensities and the internal conversion coefficients, as shown in Part B – ¹²⁴Sb Decay Data Evaluation, by using the EMISSION program. They are in good agreement with the values of participants 2 and 8.

The agreement found between the measured values and the calculated values derived from all the available measured relative $I_{\gamma i}$ values and the internal conversion coefficients confirm the consistency of that decay scheme.

6.8. Findings for the determination of photon emission intensities

Six participants in eight carried out measurements of the γ - ray emission intensities. All of them used the same kind of apparatus, however significant differences in their reported uncertainty budget were observed, not only due to the activity uncertainty.

The sets of absolute values measured by the participants are generally consistent. Similarly, the set of relative values are consistent and moreover they are in agreement with the other values published elsewhere. So, no major problem was observed.

The dispersion of participants' results obtained for the absolute intensity of the most intense gamma-ray line (602 keV) is significantly smaller (almost three times less) than the one found for activity concentrations. Since those absolute intensities are inversely proportional to activity concentration values, this means that the activity and spectrometry measurement results of a participant are correlated, *i.e.* exposed to a same cause of variation. The volatility of antimony could explain that observation to some extent, even if some participants gave experimental evidence of the resolution of that problem.

Most of the participants made an effort to determine the weak gamma rays in order to improve the overall decay scheme (see Part B - ¹²⁴Sb Decay Data Evaluation, for details).

The decay scheme evaluation was established on the basis of all the available relative gamma emission intensities, then the activity measurement results had no influence and the weight of the strong 602 keV line reduced to that of its internal conversion coefficient. The resulting gamma intensities were fully consistent with those obtained from absolute gamma emission measurements. From this comparison, it was concluded that the absolute γ -ray emission intensities can be considered as reliable and the decay scheme consistent.

7. Conclusions on the overall exercise

All along this exercise, various problems or questions appeared.

a) Volatility of the antimony solution

When preparing its own solid sources, LNE-LNHB noticed a volatility of antimony. This difficulty was pointed out and a remedy proposed to other participants for their sources. However, only two participants reported the same problem and dealt with. A possible evaporation of the initial solution could explain some low activity results. This assumption is supported by the smaller dispersion of absolute intensity values found by participants for the 602 keV gamma ray compared to their activity concentration values.

b) Activity measurements

Two participants carried out coincidence or anticoincidence measurements with both a proportional counter and a liquid scintillation detector in the beta channel, for which they got very close results. Moreover, the set of results from the coincidence and anticoincidence measurements is consistent. This can confirm the robustness of these results.

The other results $(4\pi - \gamma, \text{ ionisation chambers and LSC})$ are generally lower but are consistent all together. By now, there is no identified and validated explanation to those lower results.

c) Efficiency curve of HPGe detectors

In this exercise, a participant reported results obtained with an HPGe detector for which the efficiency curve was determined using a Monte Carlo calculation method combined with experimental results. The related uncertainty was significantly lower than for the other participants. The uncertainties claimed by that participant seem optimistic but were accepted for determining the relative photon emission intensities. The assessment of uncertainties resulting from such calculations, associated with measured data, is a question to be addressed since this kind of procedure should become widespread in the future.

In spite of all those questions, the comparison of the absolute photon emission intensities deduced from the measurements carried out in that exercise with the values deduced from the relative photon emission intensities resulting from this exercise and other available measurements, and internal conversion coefficients, demonstrates that inconsistencies in activity measurement results are not due to decay scheme data and then, other causes must be sought. ¹²⁴Sb – Activity measurement and Determination of photon emission intensities

Characteristics of the radioactive solution

LNE-LNHB

Radionuclide : Sb-124

Ampoule number :

Mass of the solution : - g

Chemical composition of the standard solution : SbCl₃ in HCl 2M

Advice: During the preparation of solid sources, a problem of volatility with Sb-124 was encountered. To try to solve it we followed the procedure : 1) a weighed deposit of radioactive solution of Sb-124 was made for each solid source, 2) a drop with an equivalent mass of 2,5 mol/L NaOH was added on each deposit. The purpose was to precipitate the element antimony as insoluble oxychlorides (SbOCl, ...).

Activity per gram of solution : $\sim 1,5$ MBq/g (reference date)

Impurity : Gamma and X-ray impurities are under checking and will be checked again in three months. The results of the LNHB tests will be sent to the participants. All the participants are requested to check for impurities at the receipt of the ampoule and

All the participants are requested to check for impurities at the receipt of the ampoule and three months later.

Homogeneity of all the ampoules less than 0,1 %

Reference Date : March 1, 2007 – 0H UTC

Saclay, February 2007

APPENDIX A2

¹²⁴Sb - Activity measurement

EUROMET PROJECT 907

I. <u>GENERALITIES</u>

Participating laboratory :

Name of the contact person :

e-mail address :

Date of receipt of ampoule (DD-MM-YYYY) :

Date of report to LNHB (DD-MM-YYYY) :

Reference date 1 March 2007, 0h UTC **

Half-life : :(60.20 ± 0.03) d **

Ampoule number :

List of all the methods used and the name of the corresponding reporting files

method (acronym)	filename	Name(s) of the person(s) who carried out the measurements

II. PRELIMINARY MEASUREMENTS

II. 1. <u>Method used for preliminary measurements:</u>

Please indicate the method used: calibrated ionization chamber; well crystal, other...

Posults obtained	Radioactivity concentration, in kBq.q ⁻¹	Date of this measurement	
Results obtained	(at reference date)		
Before opening the original ampoule	(
After transfer to another ampoule (if			
relevant)			

Total mass of solution found in the ampoule (g) :

^{*} UTC coordinated universal time

^{**}Nuclear Data Sheet, vol 80, 4(1997), H.Limura et al.

II. 2. Adsorption tests

Please take into account the adsorption tests in the evaluation of the final results.

Please, first rinse the original "empty" ampoule with water or with a solution chemically similar to the radioactive solution. Then chose one of the two methods given below:

- Adsorption tests carried out with liquid scintillation counting by using water-immiscible cocktail to measure the residual activity
- Adsorption tests carried out with proportional counting or g spectrometry by using an aggressive solution to remove most of the activity and prepare solid source(s) to measure the residual activity

Please explain the measuring procedure used:

Activity remaining in the "empty" original ampoule (Bq) .:

Date of this test

Correction factor deduced from these adsorption tests and its standard incertainty:

II. 3. Impurity checks :

Method of measurement :

Nuclide	Radioactivity concentration, in kBq.g ⁻¹ (at reference date)	Impurity to ¹²⁴ Sb ratio and its uncertainty
III. PRIMARY MEASUREMENT METHODS

III. 1. Coincidences measurement

- a Preparation sources
- Dilution

Dilution carried out? (Y/N):

If yes please indicate the diluent use (chemical, concentration) :

Dilution factor:

Sources

Solid sources	
Type of film (VYNS,)	
and typical thickness	
Metallic coating and	
typical thickness	
Is the film coated on	
both sides? (Y/N)	
Metallic ring inner	
diameter	
Metallic ring outer	
diameter	
Metallic ring thickness	
Number of sources	
prepared	
Typical mass	

Liquid sources		
Scintillator		
Chemical used to		
stabilize the cocktail		
quenching agent		
Type of vials		
Volume of cocktail		
Number of sources prepared		
Typical mass		

• Weighing

Balance used:

Date of last calibration:

Traceability to SI:

Weighing conditions: (temperature humidity)

Typical value of buoyancy correction:

b Method and instrumentation

DESCRIPTION OF THE PHOTON COUNTING DETECTOR				
(channel 1)				
Scintillato	or detector		Semiconductor	r detector
Crystal material			Nature	
Number of crystals			Number of detectors	
Crystal diameter			Туре	
Well diameter			Diameter	

Window material	
Distance between	
photon counter and	
source	
Resolution at	
FWHM	
Solid angle	
Well type(Y/N)	
Crystal height	
Well depth	
Thickness	

Window material	
Distance between	
photon counter and	
source	
Resolution at	
FWHM	
Solid angle	
Well type(Y/N)	
Coaxial(Y/N)	
Planar(Y/N)	
Volume	

DE	SCRIPTION OF THE EI	LECTRON hannel 2)	I COUNTING COUNTER
Proportion	al counter		LS Coun
Type (pill box, gas flow,)			Type of counter
Pressurized ? (Y/N)			Type of phototubes
Solid angle (2 or 4 π)			Number of phototubes
Height of each half			Operating temperature
Anode material			Coincidence resolving time between the phototubes
Anode diameter			Efficiency variation method
Anode length			
Anode distance from source			
Gas nature			
Gas pressure			
voltage applied			
Wall material			
Wire length			

nel 2)		
	LS Coun	ter
	Type of counter	
	Type of phototubes	
	Number of phototubes	
	Operating temperature	
	Coincidence resolving	
	time between the	
	phototubes	
	Efficiency variation	
	method	

ASSOCIATED ELECTRONICS				
Chanı	nel 1		Channel	2
Discrimination level			Discrimination level or	
or window			window	
type of dead time			type of dead time	
Minimum dead time			Minimum dead time	
Method used for			Method used for	
measurement			measurement	
Live time clock Y/N				
Pulser technique Y/N				
loss free counting				
Y/N				
Pile-up rejector Y/N				
	COUI	NTING DA	ATA	
Typical count rate			Typical count rate	
Background rate			Background rate	
Typical time for one			Typical time for one	
measurement			measurement	
Number of sources			Number of sources	
measured			measured	
Maximum ¹²⁴ Sb			Maximum ¹²⁴ Sb	
efficiency achieved			efficiency achieved	

Distance between	
detector and source	

Distance between detector and source

c Detailed Uncertainty Budget

Uncertainty components^{*}, in % of the activity concentration, due to

Relative standard uncertainties	$u_{\rm i} \ge 10^4$	
	evaluated method	
Contributions due to	А	В
counting statistics weighing background dead time resolving time Gandy effect pile-up decay data quenching extrapolation of efficiency curve Accidental coincidences other		
Relative combined standard uncertainty, u_{c}		

d Final result for a given method

radioactivity concentration of the ¹²⁴ Sb solution on the reference date kBq g ⁻¹ ,	combined uncertainty kBq g ⁻¹	combined uncertainty %

Remarks

III. 2. <u>4πγ measurement</u>

- a Preparation sources
- Dilution

Dilution carried out? (Y/N):

If yes please indicate the diluent use (chemical, concentration) :

The uncertainty components are to be considered as approximations of the corresponding standard deviations (see also *Metrologia*, 1981, **17**, 73 and *Guide to expression of uncertainty in measurement*, ISO, corrected and reprinted 1995).

Dilution factor:

• Solid sources

Type of film (VYNS,...) and typical thickness:

Number of sources prepared

Typical mass :

• Weighing

Balance used:

Date of last calibration:

Traceability to SI:

Weighing conditions: (temperature humidity)

Typical value of buoyancy correction:

b Method and instrumentation

SCINTILLATION DETE	ECTOR DESCRIPTION	
Crystal material (Nal, Csl,)		
Number of crystals		
Total solid angle (sr)		
Well type (Y/N):		
If yes Well diameter		
If yes Well depth		
Crystal diameter		
Crystal height		
Window material		
Window thickness		
Resolution at		
FWHM =		
ASSOCIATED	ELECTRONICS	
Discrimination level or		
window		
Dead time and std		
uncertainty		
Type of dead time (extending		
or non extending) μs		
Method of measurement of		
dead time		
Live time clock Y/N		
Pulser technique Y/N		
loss free counting Y/N		
Pile-up rejector Y/N		

COUNTIN	NG DATA
Typical count rate	
Background rate	
Typical time for one	
measurement	
Number of sources	
measured	
Maximum ¹²⁴ Sb efficiency	
achieved	
Distance between detector	
and source	

c Detailed Uncertainty Budget

Uncertainty components^{*}, in % of the activity concentration, due to

Polativo standard uncortaintios	$u_{\rm i} \ge 10^4$	
	evaluated	method
Contributions due to	А	В
weighing		
dead time		
counting statistics		
background		
Weighting		
Impurities		
Decay corrections		
pile-up		
counting time		
adsorption		
interpolation from calibration curve		
extrapolation due to energy threshold		
Other components		
Relative combined standard uncertainty, <i>u</i> _c		

The uncertainty components are to be considered as approximations of the corresponding standard deviations (see also *Metrologia*, 1981, **17**, 73 and *Guide to expression of uncertainty in measurement,* ISO, corrected and reprinted 1995).

d Final result for a given method

Name(s) of the person(s) who carried out the measurements:

radioactivity concentration of the ¹²⁴ Sb solution on the reference date kBq g ⁻¹ ,	combined uncertainty kBq g ⁻¹	combined uncertainty %

IV. SECONDARY MEASUREMENT METHODS

IV. 1. <u>y</u> Spectrometry

- a Preparation sources
- Dilution

Dilution carried out? (Y/N):

If yes please indicate the diluent use (chemical, concentration) :

Dilution factor:

• Solid sources

Type of film (VYNS,...) and typical thickness:

Number of sources prepared :

Typical mass :

• Weighing

Balance used:

Date of last calibration:

Traceability to SI:

Weighing conditions: (temperature humidity)

Typical value of buoyancy correction:

b Method and instrumentation

SEMI CONDUCTOR DESCRIPTION		
Nature(HPGe, Ge(Li), Si(Li))		
type (p or n)		
Coaxial or planar		
Number of detectors		
Total solid angle		
Well type (Y/N):		
If yes Height Diameter Well		
If yes diameter Well		
depth Window		
material Window		
thickness		
Resolution at X keV		
FWHM =		

ASSOCIATED	ELECTRONICS
Discrimination level or	
window	
Dead time and std	
uncertainty	
Type of dead time (extending	
or non extending) μs	
Method of measurement of	
dead time	
Live time clock Y/N	
Pulser technique Y/N	
loss free counting Y/N	
Pile-up rejector Y/N	
COUNTI	NG DATA
Typical count rate	
Background rate	
Typical time for one	
measurement	
Number of sources	
measured	
Maximum ¹²⁴ Sb efficiency	
achieved	
Distance between detector	
and source	

c Detailed Uncertainty Budget

Uncertainty components^{*}, in % of the activity concentration, due to

Relative standard uncertainties	u₁ x 10 ⁴ evaluated method	
Contributions due to	А	В
Ionization current (111In) including background and geometry uncertainties Reference source Linearity Calibration factor Weighting Impurities Decay corrections		
Quadratic summation		
Other		
Relative combined standard uncertainty, <i>u</i> c		

d Final result for a given method

Name(s) of the person(s) who carried out the measurements:

radioactivity concentration of the ¹²⁴ Sb solution on the reference date kBq g ⁻¹ ,	combined uncertainty kBq g ⁻¹	combined uncertainty %

IV. 2. Ionisation Chamber

- a Preparation sources
- Dilution

Dilution carried out? (Y/N):

If yes please indicate the diluent use (chemical, concentration) :

Dilution factor:

• Liquid sources

Type of ampoule used:

Thickness:

The uncertainty components are to be considered as approximations of the corresponding standard deviations (see also *Metrologia*, 1981, **17**, 73 and *Guide to expression of uncertainty in measurement*, ISO, corrected and reprinted 1995).

Number of sources prepared:

Typical volume :

• Weighing

Balance used:

Date of last calibration:

Traceability to SI:

Weighing conditions: (temperature humidity)

Typical value of buoyancy correction:

IONISATION CHAMI	BER DESCRIPTION
Producer and type	
Shape (U-shape, cylindrical)	
Gas nature	
Gas pressure	
Well diameter	
Thickness of the wall of the well	
IC external diameter	
IC height	
Material of the wall of the well	
Material of the electrodes	
Thickness of lead shielding	
CURRENT ME	ASUREMENT
Reference of the electrometer	
Current measurements	
relative to a long-lived	
reference source (Y/N)	
If yes, reference source	
description	
MEASUREN	MENT DATA
Typical current pA	
Background current pA	
IC response pA/MBq	
Typical integration time	
Number of measurements	

b Method and instrumentation

c Final result for a given method

Name(s) of the person(s) who carried out the measurements:

radioactivity concentration of the ¹²⁴ Sb solution on the reference date kBq g ⁻¹ ,	combined uncertainty kBq g ⁻¹	combined uncertainty %

IV. 3. <u>Other</u>

- a Preparation sources
- b Method and instrumentation (counter description, associated electronics, counting data)
- c Final result for a given method

Name(s) of the person(s) who carried out the measurements:

radioactivity concentration of the ¹²⁴ Sb solution on the reference date kBq g ⁻¹ ,	combined uncertainty kBq g ⁻¹	combined uncertainty %

V. LABORATORY FINAL RESULT

radioactivity concentration of the ¹²⁴ Sb solution on the reference date kBq g ⁻¹ ,	combined uncertainty kBq g ⁻¹	combined uncertainty %

If the final result is obtained by a combination of results from several measurement methods, please explain the procedure below (weighted mean, correlations,...)

Geometry	acronym	Detector	acronym
4π	4P	proportional counter	PC
defined solid angle	SA	pressurized proportional counter	PP
2π	2P	liquid scintillation counting	LS
		Nal(TI)	NA
		Ge(HP)	GH
		Ge-Li	GL
		Si-Li	SL
		Csl	CS
		ionisation chamber	IC
		bolometer	во
		calorimeter	СА
		PIPS detector	PS
		Grid ionisation chamber	GC

Appendix -	Acronyms used	to identify	different	measurement	methods
------------	---------------	-------------	-----------	-------------	---------

Radiation	acronym	Mode	acronym
positron	PO	efficiency tracing	ET
beta particle	BP	internal gas counting	IG
Auger electron	AE	CIEMAT/NIST	CN
conversion electron	CE	sum counting	SC
bremsstrahlung	BS	coincidence	СО
gamma ray	GR	anti-coincidence	AC
x - rays	XR	coincidence counting with efficiency tracing	СТ
alpha - particle	AP	anti-coincidence counting with efficiency tracing	AT
mixture of various radiation e.g. x and gamma	MX	triple-to-double coincidence ratio counting	TD
		selective sampling	SS
		high efficiency	HE

Examples	
method	acronym
4π (PC) β – γ -coincidence counting	4P-PC-BP-NA-GR- CO
$4\pi(PPC)\beta$ - γ -coincidence counting efficiency. tracing	4P-PP-MX-NA-GR- CT
defined solid angle α -particle counting with a PIPS detector	SA-PS-AP-00-00-00
4π (PPC)AX- γ (GeHP)-anticoincidence counting	4P-PP-MX-GH-GR- AC
4π CsI-β,AX,γ counting	4P-CS-MX-00-00-00
calibrated IC	4P-IC-GR-00-00-00

 Table A.1 – Instrumentation - general characteristics, geometry arrangements

Participant	2	2	3	3	5	6	7	8	8
DETECTOR	Coaxial	<u>Planar</u>	Coaxial	<u>Planar</u>	Coaxial	Coaxial Canberra GC 2520	Coaxial Canberra GC4018	Coaxial CANBERRA	<u>Planar</u> CANBERRA
Crystal material	Ge	Ge	Ge	Ge	HP Ge	HP Ge	Ge	HPGe	HPGe
If Ge: material type (P/N)	Р	Р	N	N	N	Р	Р	N-type (Extended range)	
Crystal diameter (mm)	64,9	50,5	48	19,5	50,9	53,4	63	77	25,5
Crystal thickness (mm)	79,7	20	52,7	10	62,5	51	59	78	15
Crystal active volume (cm ³)	264		95,36	2,99	127	112,5	cca 180		500 mm ² active area
Window material	Al	Be	Be	Be	Be	Al	Al	Al	Be
Window thickness (µm)	1000		500	100	500	500	1500	1000	150
SOURCES									
Source-to-detector window distance (cm)	20	20	10,35	8,025	18,5	44,5	25	4, 8, 12	0,325
Nb Sources used d (cm) m (mg)	10 ~ 0,2 ??		6 ~ 0,2 15,5 to 20,6	1 ~ 0,2 15,5	1 ~ 0,3 8,776	3 - 41,640 38,375 61,870	5 ~ 0,6 29 to 31	4 ~ 0,5 2 to 36	
Nature and thickness of cover foil	0,02 mm Polystyrene		12 µm Mylar	12 µm Mylar	2,2 µg cm ⁻² polyester		25 μg cm ⁻² PE	10 and 0,3 mg cm ⁻²	Polyester
COLLIMATOR (Y/N)	Ν	Ν	Ν	Y	Ν	Ν	Ν	N/A	N/A
If Yes: Material				W					
If Yes: Thickness (mm)				2,23					
If Yes: Collimator diameter (mm)				10					
If Yes: Detector window-to-collimator distance (cm)				7,075					

Participant	2	2	3	3	5	6	7	8	8
PREAMPLIFIER			Ortec	Canberra	Ortec	Canberra	Canberra	Canberra	Canberra
Model	A257 P	2002	120-6		EG&ORTEC	2002 CSL	2002 CSL	2002 CSL	2008 SL
		CSL			137CN2				
AMPLIFIER	Ortec	Ortec	Canberra	Canberra	Canberra	Canberra	Canberra	Canberra	Canberra
Model	572	572	2020	2026	2021	2024	DSP 9660	AFT Research Amp	AFT Research
								2025	Amp 2025
Shaping time constant	6 µs	6 µs	6 µs	6 µs	4 µs	4 μs	Equivalent	4 μs	4 μs
							to 4 µs		
Pile-up rejector (Y/N)	N	N	N	N	N	N	Y	Y	Y
ADC	Ortec	Ortec	Ortec	Ortec	Nucler Data	Canberra	Canberra	Silena	Silena
Model	919E	919E	926	926	ND570	Accuspec/A	DSP 9660	7423 UHS	7423 UHS
						Board			
						840632A			
Coding type	SA	SA	Successive-	Successive-		Wilkinson		FDT	FDT
(Wilkinson, ")			approximation	approximation					
			type with	type with					
			sliding-scale	sliding-scale					
OTHED			inearization.	inearization.					
OTHER Cata Integration (V/N)	N	N	N	N	N	Na		LCN Madel DE7722h	LCN Medel
Gate integrator (1/N)	IN	IN	IN	IN	IN	INO		LGN, Model DE/7520	DE7722b
Energy resolution									DE77520
(FWHM)									
(1 w HW)				0.25	1				0.25
- at 122 keV (keV)	1.03	0.662	0.77	0,23	1 06	0.875	0.838	12	0,25
- at 122 KeV (KeV)	1,05	0,002	0,77	0,49	1,00	0,075	0,050	1,2	(0,55 (059,5 keV)
- at 1332 keV (keV)	1.86		1 73		2.07	2.37	1.78	2.5	KC ()
- at 2500 keV (keV)	-,00		2.60		2.71	-	2.6	-,-	

 Table A.1 – Instrumentation - general characteristics, geometry arrangements

EFFICIENCY	CALIBRATION		
Nuclide and en	nergy of the corresponding peak(s), in keV, used		
Participant 2		Participant 7	
¹⁶⁶ Ho ^m	48,8;55,9;80,6;184,4;215,4;280,4;300,7;365,8;410,9;450,6;		The calibration was done by MC calculation and for the validation the
	529,8; 571,0; 670,5; 711,7; 752,3; 810,3; 830,6; 951		following energies were used:
		²⁴¹ Am	59,5
¹⁵² Eu	39,9 ; 45,7 ; 121,8 ; 244,7 ; 344,3 ; 411,1 ; 444 ; 779 ; 867,4 ; 964,1 ; 1112,1 ; 1408	¹⁰⁹ Cd	88
		⁵⁷ Co	122
Participant 3	Next page	¹³⁹ Ce	165
		¹¹³ Sn	391
		⁸⁵ Sr	514
Participant 5	<u>Next pages</u>	¹³⁷ Cs	662
		⁸⁸ Y	898;1836
Participant 6		⁶⁰ Co	1173 ; 1332
⁶⁰ Co	1173 ; 1332	232 Th (208 Tl)	2615
¹³³ Ba	276 ; 302 ; 356 ; 383		
¹³⁷ Cs	661	Participant 8	
¹⁵² Eu	121;778;1408	Coaxial HPGe	(keV)
²⁴¹ Am	59	²⁴¹ Am	59,5
		⁵⁷ Co	122;136
		¹³⁷ Cs	662
		¹³⁴ Cs	605;796
		⁵⁴ Mn	835
		⁶⁵ Zn	1115,5
		⁶⁰ Co	1173 ; 1332
		Planar HPGe	(keV)
		²⁴¹ Am	26,3 ; 59,5
		¹³⁷ Cs	31,8 ; 32,2 ; 36,4 ; 37,3

Table A.2. Instrumentation – efficiency calibration

Code	Efficiency interpolation	Peak analysis
Number		
2	The experimental points of the efficiency curves of germanium spectrometers have been fitted by a polynomial function. The better degree was third degree and evaluated by χ^2 statistical test.	The integral method for liquid peak area evaluation was adopted. This method fits the peak area through an analytical function.
3	Gamma : Polynomial function adjusted with the experimental points, using the software Effigie. Degree of the polynom : 4	The net area of the peaks was determined by using the software MAESTRO. For the multiplets analysis (e.g. 709-713 keV) the computer code COLEGRAM (CEA/LNE-LNHB, Saclay, France) was used, as well as for some weak peaks located over a strongly decreasing slope, e.g. 469 and 481 keV over the Compton front associated to the 602 keV. For the X-ray peaks, the two regions of interest were analyzed with the COLEGRAM computer code.
5	Linear least square fit with cubic spline functions	 a) Gaussian fit, background determined between -3 σ to +3 σ and limited by a linear extrapolation from range : between -6 σ to -3 σ (at lower energy) and linear extrapolation from +3 σ to +6 σ (at higher energy), with a step at peak energy. (PROGRAM "GELI") ref /2/ b) Sum of channel contents, reduced by background determined by method a) (PROGRAM "GELI") ref /2/ c) Gaussian fit with complete curve analysis of the peak region of interest including linear or exponential background fit (PROGRAM "Fit9") ref /3/
6	The method of interpolation was based on the use of EFFIGIE software (J. Morel, "Détermination de la réponse en efficacité des détecteurs gamma-X par lissage de points expérimentaux: code EFFIGIE", LPRI, Saclay, France, 1996).	Give a brief description of your peak analysis method: The net area and the associated uncertainty values were determined by using the ACCUSPEC software (Canberra, USA). The analysis of multiple peaks was performed by using the COLEGRAM software (H. Ruellan, M. C. Lépy, J. Plagnard, "Presentation du logiciel "COLEGRAM"", Note Technique LPRI/95/016, Saclay, France, 1995).
7	The efficiency was calculated by Monte Carlo method for each energy and the results were validated by experimental points.	TPA, continuum subtraction with step function, 1 to 5 channels. (???)
8	The experimental values were used to set the parameters for a Monte Carlo simulation of the detector set-ups. The efficiency for each peak energy was then calculated and no interpolation was involved.	The peak analysis was performed by peak fitting using ORTEC's GammaVision-32 (v.6.01). The program was used to obtain the peak area only. The rest of the calculations were done using a spreadsheet.

Table A.3. Efficiency interpolation

Code Number	Coincidence Summing effect	Absorption by the screens between	Pile-up
i (uniber		sample and detector	
2	No	No	no
3	Yes (ETNA)	Not necessary	no
5	Yes, described in [1] using nuclide data from the Nucléide 2000 database, version july 2007.	no	Yes, Signal from a 50 Hz pulse generator was fed into pre-amplifier test input positioned at upper region of the spectra.
6	No, (the summing effects were negligible, because the measurements were performed at 44,5 cm from the detector window)	No	No
7	Yes, The coincidence probabilities were calculated by Monte Carlo method by own code. The correction factor for photon with energy E1 in cascade with energy E2, E3 (summing-out effect) is: $C_{out}(E1) = 1/(1 - p_{1,2} * T(E2) - p_{13}*T(E3))$ $p_{1,2}$: is probability of the coincidence E1 and E2 T : is total efficiency The correction factor for summing-in effect when photon with energy E0 is crossover by cascade Ea \rightarrow Eb (E0 = Ea+Eb) is: $C_{in}(E0) = 1/(1 + p_{ab}*P(Ea)*P(Eb)/P(E0) +)$ p_{ab} is the probability of the cascade Ea \rightarrow Eb related to the emission probability of E0 P : is peak efficiency and $C1 = C_{out} * C_{in}$	No, included in efficiency calibration	Yes, pile-up and dead-time losses were corrected electronically using DSP and AIM Canberra modules
8	Yes, Correction for the coincidence summing effect was applied to peaks with emission probabilities >2%. Cascading gamma photons only were taken into account and the calculations were performed manually as described in (for example) Debertin and Helmer "Gamma- and X-ray spectrometry with semiconductor detectors" (Elsevier, 1988).	Yes, included in efficiency calibration. The absorbers were also included in the Monte Carlo simulation.	Yes, The pile-up rejection of the spectroscopy amplifiers was turned ON and adjusted. For the dead-time, the live time method was used (non-extending dead-time). The inhibit function losses of the pulsed feedback preamplifier of the low-energy planar detector were taken into account

Participant 3: HPGe detectors efficiency calibration

Nuclide	Energy	Experimental	Relative	Absolute
	(keV)	efficiency	uncertainty (%)	uncertainty
¹⁵² Eu	121.78	8.762E-03	0.63	5.520E-05
⁵⁷ Co	122.06	8.775E-03	0.3	2.633E-05
⁵⁷ Co	136.47	8.298E-03	1.4	1.162E-04
¹³⁹ Ce	165.86	7.495E-03	0.35	2.623E-05
192 Ir	205.79	6.206E-03	1.31	8.130E-05
¹⁵² Eu	244.70	5.281E-03	0.63	3.327E-05
¹³³ Ba	276.40	4.686E-03	0.61	2.858E-05
¹⁹² Ir	295.96	4.336E-03	0.6	2.602E-05
¹³³ Ba	302.85	4.247E-03	0.6	2.548E-05
¹⁹² Ir	308.46	4.168E-03	0.62	2.584E-05
¹⁹² Ir	316.51	4.038E-03	0.42	1.696E-05
¹⁵² Eu	344.28	3.677E-03	0.51	1.875E-05
¹³³ Ba	356.01	3.578E-03	0.5	1.789E-05
¹³³ Ba	383.85	3.298E-03	0.6	1.979E-05
¹⁵² Eu	411.35	3.050E-03	0.86	2.623E-05
¹⁵² Eu	443.97	2.814E-03	0.66	1.857E-05
$^{110}Ag^{m}$	446.81	2.815E-03	1.1	3.097E-05
¹⁹² Ir	468.07	2.656E-03	0.61	1.620E-05
^{134}Cs	475.34	2.599E-03	1.51	3.924E-05
¹⁹² Ir	484.58	2.563E-03	0.9	2.307E-05
^{134}Cs	563.23	2.195E-03	0.47	1.032E-05
^{134}Cs	569.32	2.184E-03	0.42	9.173E-06
¹⁹² Ir	588.58	2.096E-03	0.68	1.425E-05
¹⁹² Ir	604.41	2.045E-03	0.64	1.309E-05
^{134}Cs	604.69	2.053E-03	0.26	5.338E-06
¹⁹² Ir	612.46	2.022E-03	1.57	3.175E-05
$^{110}\text{Ag}^{\text{m}}$	620.36	2.005E-03	1.21	2.426E-05
$^{110}\text{Ag}^{\text{m}}$	657.76	1.902E-03	0.31	5.896E-06
¹³⁷ Cs	661.66	1.890E-03	0.42	7.938E-06
$^{110}Ag^{m}$	677.62	1.869E-03	1.06	1.981E-05
$^{110}Ag^{m}$	687.01	1.855E-03	0.77	1.428E-05
¹⁵² Eu	688.62	1.817E-03	2.2	3.997E-05
$^{110}Ag^{m}$	706.68	1.752E-03	0.75	1.314E-05
$^{110}Ag^{m}$	744.28	1.710E-03	1.43	2.445E-05
$^{110}\text{Ag}^{\text{m}}$	763.94	1.646E-03	0.47	7.736E-06
¹⁵² Eu	778.90	1.605E-03	0.39	6.260E-06
^{134}Cs	795.84	1.580E-03	0.26	4.108E-06
^{134}Cs	801.93	1.571E-03	0.36	5.656E-06
$^{110}Ag^{m}$	818.03	1.547E-03	0.72	1.114E-05
^{54}Mn	834.84	1.513E-03	0.43	6.506E-06
¹⁵² Eu	867.38	1.457E-03	0.73	1.064E-05

Detector G1, energy range 100 – 2000 keV: experimental points

$^{110}\text{Ag}^{\text{m}}$	884.68	1.442E-03	0.5	7.210E-06
⁸⁸ Y	898.04	1.421E-03	0.5	7.105E-06
$^{110}Ag^{m}$	937.49	1.368E-03	0.46	6.293E-06
¹⁵² Eu	964.08	1.323E-03	0.46	6.086E-06
¹³⁴ Cs	1038.56	1.241E-03	0.72	8.935E-06
¹⁵² Eu	1086.41	1.190E-03	0.53	6.307E-06
¹⁵² Eu	1112.04	1.166E-03	0.54	6.296E-06
⁶⁵ Zn	1115.55	1.165E-03	0.48	5.592E-06
¹³⁴ Cs	1167.92	1.121E-03	0.77	8.632E-06
⁶⁰ Co	1173.24	1.120E-03	0.3	3.360E-06
¹⁵² Eu	1212.95	1.082E-03	1.02	1.104E-05
⁶⁰ Co	1332.51	1.001E-03	0.32	3.203E-06
¹³⁴ Cs	1365.16	9.677E-04	0.59	5.709E-06
$^{110}Ag^{m}$	1384.30	9.747E-04	0.51	4.971E-06
¹⁵² Eu	1408.01	9.396E-04	0.53	4.980E-06
$^{110}\text{Ag}^{\text{m}}$	1475.79	9.125E-04	0.92	8.395E-06
$^{110}Ag^{m}$	1505.04	8.997E-04	0.58	5.218E-06
$^{110}Ag^{m}$	1562.29	8.682E-04	1.3	1.129E-05
⁸⁸ Y	1836.05	7.314E-04	0.42	3.072E-06

Detector G1, energy range 500 – 3500 keV: experimental point

			Relative	
Nuclide	Energy	Experimental	uncertainty	Absolute
	(keV)	efficiency	(%)	uncertainty
¹³⁷ Cs	661.66	0.001890	0.41	7.82E-06
⁵⁴ Mn	834.838	0.001477	3.03	4.47E-05
⁵⁶ Co	846.77	0.001500	0.62	9.36E-06
⁵⁶ Co	1037.8427	0.001260	0.79	9.93E-06
⁶⁰ Co	1173.228	0.001122	0.20	2.24E-06
⁵⁶ Co	1238.2883	0.001063	0.67	7.11E-06
⁶⁰ Co	1332.492	0.000996	0.18	1.82E-06
⁵⁶ Co	1771.3567	0.000754	0.69	5.23E-06
⁵⁶ Co	2015.2147	0.000660	0.92	6.04E-06
⁵⁶ Co	2034.7907	0.000664	0.70	4.63E-06
⁵⁶ Co	2598.5	0.000514	0.69	3.55E-06
⁵⁶ Co	3202.029	0.000403	1.08	4.34E-06
⁵⁶ Co	3253.503	0.000394	0.83	3.28E-06
⁵⁶ Co	3273.079	0.000393	1.06	4.17E-06

Nuclide	Energy (keV)	Experimental Efficiency	Relative uncertainty (%)	absolue uncertainty
²⁴¹ Am	13.93	0.0009846	1.008	9.93E-06
⁵⁷ Co	14.41	0.0009830	1.884	1.85E-05
²⁴¹ Am	17.51	0.0010712	0.925	9.91E-06
²⁴¹ Am	21.01	0.0011061	1.349	1.49E-05
¹⁰⁹ Cd	22.10	0.0010964	3.561	3.90E-05
¹⁰⁹ Cd	25.07	0.0011078	4.146	4.59E-05
²⁴¹ Am	26.34	0.0011356	1.537	1.75E-05
¹⁴⁰ Ba	29.96	0.0011444	1.795	2.05E-05
¹³³ Ba	30.85	0.0011506	0.916	1.05E-05
¹³⁷ Cs	31.82	0.0011545	1.607	1.85E-05
¹³⁷ Cs	33.19	0.0011464	7.458	8.55E-05
¹³⁹ Ce	33.3	0.0011923	1.048	1.25E-05
¹³³ Ba	35.26	0.0011819	0.943	1.11E-05
¹³⁷ Cs	36.63	0.0011839	2.283	2.70E-05
¹³⁹ Ce	38.06	0.0012219	1.511	1.85E-05
¹⁵² Eu	39.91	0.0011829	1.088	1.29E-05
¹⁵² Eu	42.75	0.0012032	7.990	9.61E-05
¹⁵² Eu	45.73	0.0011758	1.493	1.76E-05
¹³³ Ba	53.16	0.0012357	1.522	1.88E-05
²⁴¹ Am	59.54	0.0012154	0.462	5.61E-06
¹³³ Ba	80.9	0.0012119	0.922	1.12E-05
¹⁰⁹ Cd	88	0.0011586	1.198	1.39E-05
¹⁵² Eu	121.8	0.0008789	0.536	4.71E-06
⁵⁷ Co	122.06	0.0008765	0.302	2.64E-06

Detector G2, energy range 10 – 120 keV: experimental points

Enclosure Participant 5 - G-01

Euromet Project 907

Experimental efficiencies used for the calibration of the spectrometer G8, evaluation method S

Efficency values G80, file 80FES2S.A07 G80S, point soure, position 4, polyester (2005-2007), energies evaluated by the program nuclide E(keV) p_gamma summcorr efficiency uncert.(%) S 2.937E-03 .95 2 SR85 13.6 .5917 1.005 3xC057 2x14.4 .0916 1.006 S 3.065E-03 1.79 2 Y88 14.4 .6160 1.007 S 2.984E-03 1.21 2 3x 4.50 2 NB93M 1x16.6 .0925 1.000 S 3.300E-03 NB93M 1x18.7 .0179 1.000 S 3.345E-03 4.27 2 1.26 2 CD109 1x22.6 1.0137 1.000 S 3.579E-03 1.67 2 28.1 1.3950 1.003 S 3.680E-03 I125 1xBA133 1x31.7 1.2100 1.009 S 3.640E-03 1.93 2 32.1 S 3.664E-03 1.99 2 CS137 1x.0553 1.000 EU152 5x 39.9 .5910 1.012 S 3.695E-03 2.24 2 1.000 46.5 S 3.797E-03 1.47 2 PB210 1x.0424 BA133 53.2 .0220 1.010 S 3.825E-03 1.15 2 $1 \mathrm{x}$ AM241 2x59.5 .3590 1.000 S 3.890E-03 1.31 2 80.2 1.80 2 I131 2x.0262 1.005 S 3.848E-03 80.9 1.008 S 3.819E-03 .96 2 BA133 1x.3668 CD109 1x88.0 .0366 1.000 S 3.796E-03 1.29 2 1.002 S 3.825E-03 1.75 2 SE75 1x96.7 .0342 SE75 1x121.1 .1720 1.007 S 3.505E-03 2.10 2 .84 2 121.8 .2858 EU152 3x 1.011 S 3.470E-03 122.1 .8560 1.001 S 3.502E-03 .71 2 C057 2x1.57 2 SE75 1x136.0 .5820 1.007 S 3.359E-03 C057 136.5 .1068 .997 S 3.338E-03 .96 2 2x.57 2 CE139 165.9 .7990 1.003 S 3.002E-03 1x.0351 RA226 2x186.1 1.008 S 2.749E-03 2.17 2 .82 2 EU152 5x 244.7 .0758 1.013 S 2.214E-03 SE75 264.7 .5890 1.007 S 2.075E-03 1.18 2 1xS 2.007E-03 .66 2 BA133 1x276.4 .0716 1.009 1.23 2 SE75 1x279.5 .2499 1.005 S 1.978E-03 IR192 1x296.0 .2872 1.010 S 1.871E-03 .98 2 302.9 S 1.839E-03 .65 2 BA133 1x.1833 1.008 .99 2 IR192 308.5 .2968 1.010 S 1.816E-03 1xIR192 $1 \mathrm{x}$ 316.5 .8279 1.007 S 1.752E-03 .85 2 S 1.748E-03 CR51 320.1 .0987 1.000 .79 2 1x344.3 1.006 S 1.622E-03 2.41 2 EU152 5x .2650 BA133 356.0 .6205 1.006 S 1.579E-03 .64 2 1x1.000 S 1.538E-03 1.14 2 I131 2x 364.5 .8170 BA133 1x383.9 .0894 1.003 S 1.471E-03 .68 2 S 1.400E-03 1.45 2 SE75 1x400.7 .1147 .971 444.0 S 1.268E-03 1.03 2 EU152 5x .0315 1.012 IR192 $1 \mathrm{x}$ 468.1 .4781 1.005 S 1.196E-03 .94 2 S 1.120E-03 2.09 2 RU106 1x 511.9 .2040 1.003 SR85 3x 514.0 .9850 1.002 S 1.106E-03 .73 2 CS134 567.2 .2375 1.008 S 1.019E-03 .53 2 1xBI207 1x569.7 .9774 1.004 S 1.013E-03 .82 2

Enclosure Participant 5 - G-01 *Euromet Project 907*

CS134	1x	604.7	.9763	1.005	S	9.595E-04	.42	2
RU106	1x	621.5	.1068	1.005	S	9.339E-04	2.49	2
I131	2x	637.2	.0739	1.000	S	9.157E-04	.42	2
CS137	1x	661.7	.8500	1.000	S	8.846E-04	.59	2
EU152	5x	778.9	.1294	1.007	S	7.725E-04	1.40	2
CS134	1x	796.4	.9409	1.005	S	7.620E-04	.59	2
CO58	2x	810.8	.9945	1.002	S	7.438E-04	.67	2
MN54	1x	834.8	.9998	1.000	S	7.270E-04	.42	2
EU152	5x	867.4	.0424	1.014	S	7.029E-04	.91	2
Y88	3x	898.0	.9400	1.005	S	6.865E-04	.65	2
EU152	5x	964.1	.1473	1.009	S	6.392E-04	.77	2
BI207	1x	1063.7	.7450	1.005	S	5.952E-04	.88	2
EU152	5x	1112.0	.1383	1.008	S	5.740E-04	.71	2
ZN65	1x	1115.5	.5022	1.001	S	5.733E-04	.78	2
C060	2x	1173.2	.9985	1.003	S	5.480E-04	.42	2
NA22	2x	1274.5	.9994	1.009	S	5.066E-04	.45	2
CO60	2x	1332.5	.9998	1.003	S	4.921E-04	.46	2
EU152	5x	1408.0	.2100	1.008	S	4.677E-04	.77	2
RA226	1x	1509.2	.0208	1.005	S	4.397E-04	2.68	2
RA226	2x	1764.5	.1510	1.001	S	3.828E-04	2.30	2
BI207	1x	1770.2	.0687	1.005	S	3.822E-04	1.14	2
Y88	3x	1836.1	.9933	1.006	S	3.699E-04	.54	2
RA226	2x	2204.2	.0498	1.001	S	2.985E-04	2.71	2
RA226	2x	2447.9	.0155	1.001	S	2.716E-04	2.84	2
Y88	3x	2734.0	.0061	.858	S	2.141E-04	3.62	2

Enclosure Participant 5 - G-02

Euromet Project 907

Experimental efficiencies used for the calibration of the spectrometer G8, evaluation method U

Efficency values G80, file 80FES2U.A07 G80U, point soure, position 4, polyester (2005-2007), energies evaluated by the program nuclide E(keV) p_gamma summcorr efficiency uncert.(%) U 2.999E-03 .95 2 SR85 13.6 .5917 1.005 3xC057 2x14.4 .0916 1.006 U 3.115E-03 1.79 2 Y88 14.4 .6160 1.007 U 3.073E-03 1.24 2 3x 3.09 2 NB93M 1x16.9 .1104 1.000 U 3.388E-03 CD109 1x22.6 1.0137 1.000 U 3.737E-03 1.14 2 1.48 2 I125 1x28.1 1.3950 1.003 U 3.783E-03 1.39 2 31.7 1.2100 1.009 U 3.769E-03 BA133 1xCS137 $1 \mathrm{x}$ 32.1 .0553 1.000 U 3.736E-03 1.98 2 39.9 2.21 2 EU152 5x .5910 1.012 U 3.680E-03 PB210 1x46.5 .0424 1.000 U 3.765E-03 1.46 2 53.2 U 3.830E-03 1.13 2 BA133 1x.0220 1.010 AM241 59.5 .3590 1.000 U 3.864E-03 1.30 2 2xI131 2x80.2 .0262 1.005 U 3.827E-03 1.75 2 .94 2 BA133 1x80.9 .3668 1.008 U 3.810E-03 CD109 U 3.800E-03 1.28 2 1x88.0 .0366 1.000 SE75 1x96.7 .0342 1.002 U 3.725E-03 1.78 2 121.1 1.007 U 3.495E-03 2.10 2 SE75 1x.1720 EU152 3x 121.8 .2858 1.011 U 3.494E-03 .81 2 .56 2 .8560 U 3.488E-03 C057 2x122.1 1.001 .5820 1.57 2 SE75 136.0 1.007 U 3.361E-03 $1 \mathrm{x}$.95 2 CO57 2x136.5 .1068 .997 U 3.327E-03 165.9 .7990 1.003 U 3.033E-03 .52 2 CE139 1x2.06 2 RA226 2x186.1 .0351 1.008 U 2.820E-03 .0758 EU152 5x 244.7 1.013 U 2.221E-03 .83 2 1.18 2 SE75 1x 264.7 .5890 1.007 U 2.080E-03 BA133 276.4 .0716 1.009 U 2.012E-03 .65 2 1x.2499 1.005 U 1.978E-03 1.21 2 SE75 1x279.5 .93 2 IR192 1x296.0 .2872 1.010 U 1.874E-03 BA133 1x302.9 .1833 1.008 U 1.838E-03 .64 2 308.5 U 1.796E-03 .93 2 IR192 1x.2968 1.010 U 1.752E-03 IR192 316.5 .8279 1.007 .85 2 1xCR51 1x320.1 .0987 1.000 U 1.748E-03 .78 2 EU152 344.3 .2650 1.006 U 1.631E-03 2.41 2 5x 356.0 1.006 U 1.580E-03 .64 2 BA133 1x.6205 I131 364.5 .8170 1.000 U 1.538E-03 1.08 2 2x.65 2 1.003 BA133 1x383.9 .0894 U 1.476E-03 .971 SE75 1x400.7 .1147 U 1.399E-03 1.40 2 .0315 1.00 2 EU152 5x 444.0 1.012 U 1.277E-03 468.1 1.005 U 1.198E-03 .99 2 IR192 1x.4781 RU106 $1 \mathrm{x}$ 511.9 .2040 1.003 U 1.116E-03 2.07 2 U 1.114E-03 .71 2 SR85 3x 514.0 .9850 1.002 CS134 1x 567.2 .2375 1.008 U 1.019E-03 .53 2 .81 2 BI207 569.7 .9774 1.004 U 1.014E-03 1x

U 9.601E-04

.42 2

1.005

CS134

1x

604.7

.9763

Enclosure Participant 5-G-02 *Euromet Project 907*

RU106	1x	621.5	.1068	1.005	U ö	9.310E-04	2.58	2
I131	2x	637.2	.0739	1.000	U (9.155E-04	.41	2
CS137	1x	661.7	.8500	1.000	U (8.868E-04	.59	2
EU152	5x	778.9	.1294	1.007	' U	7.768E-04	1.39	2
CS134	1x	796.4	.9409	1.005	5 U	7.600E-04	.58	2
CO58	2x	810.8	.9945	1.002	2 U	7.457E-04	.66	2
MN54	1x	834.8	.9998	1.000	U (7.300E-04	.42	2
EU152	5x	867.4	.0424	1.014	U	7.064E-04	.92	2
Y88	3x	898.0	.9400	1.005	U ö	6.903E-04	.67	2
EU152	5x	964.1	.1473	1.009	U U	6.418E-04	.77	2
BI207	1x	1063.7	.7450	1.005	U	5.978E-04	.87	2
EU152	5x	1112.0	.1383	1.008	5 U	5.760E-04	.71	2
ZN65	1x	1115.5	.5022	1.001	. U	5.760E-04	.77	2
C060	2x	1173.2	.9985	1.003	U 8	5.527E-04	.44	2
NA22	2x	1274.5	.9994	1.009	U U	5.099E-04	.43	2
CO60	2x	1332.5	.9998	1.003	U 8	4.961E-04	.44	2
EU152	5x	1408.0	.2100	1.008	5 U	4.705E-04	.77	2
RA226	1x	1509.2	.0208	1.005	U ö	4.462E-04	2.64	2
RA226	2x	1764.5	.1510	1.001	. U	3.875E-04	2.28	2
BI207	1x	1770.2	.0687	1.005	U ö	3.889E-04	1.12	2
Y88	3x	1836.1	.9933	1.006	U ö	3.737E-04	.53	2
RA226	2x	2204.2	.0498	1.001	. U	3.032E-04	2.71	2
RA226	2x	2447.9	.0155	1.001	. U	2.769E-04	2.84	2
Y88	3x	2734.0	.0061	.858	5 U	2.196E-04	3.49	2

Erratum Participant 6

In May 2009, participant 6 sent a revised set of the photon emission intensities. Since the deadline was March, 2008, the values of participant 6 used in the evaluation are the previous ones. These new values are given here as *erratum*.

Since the new values sent by participant 6 remain significantly different of those of the other participants and affected by higher uncertainties in most cases, their introduction in the evaluation would not change most of the adopted values.

As example, a new calculation was done for the most intense gamma line with 602 keV energy, with $I\gamma = 95,1 \pm 1,32$ instead of $I\gamma = 95 \pm 1$:

						Unc (if			New
602 keV	Value	Uc	Min Unc. ?	R. WGHT	Chi2/N-1	w>0,5)	Outlier ?	New Val.	Unc.
2	97,52	0,70		0,136892	0,03141			97,5198	0,70
3	97,84	0,85		0,092470	0,00152			97,8354	0,85
5	97,60	0,66		0,153219	0,01615			97,6	0,66
6	95,10	1,32					OUT(CHV)		
7	97,84	0,34	MIN	0,587607	0,01151			97,841	0,33
8	98,14	1,50		0,029810	0,01507			98,138	1,5
			CONSISTENT DATA					Nb Inp.	
			SET	Chi2 :	0,07568	Chi2 crit.:	3,32	Val.	5
Chi2	0,1		UWM :	97,78698	uc(UWM) :	0,1084			
Chi2 crit:	3,3		WM :	97,76871	uc(WM)int.	0,2598	uc(WM)ext.	0,07147	
UWM:	97,78698		LWM :	97,76871	uc(LWM) :	0,2598			
WM:	97,76871		LWM :	97,77	uc(LWM) :	0,26			
Uc (int):	0,2598								
			LWM has used weighte	ed average an	d internal				
Uc (ext) :	0,07147		uncertainty						
LWM :	97,77	0,26							

124Sb exercise, table 3 : result of measurements, absolute values, erratum participant 6 (In this table, all the g rays quoted in the literature are listed, fill in the corresponding line or not)

Energy (keV)	Efficiency (%)	Correction Factors C5: decay correction (the other corrections C1 C4 were considered negligible)	Photon emission intensity per 100 disintegrations	Elementary components of the uncertainty (% -1 σ) (U1+U2), U3, U7 (the other components were considered negligible)	Relative combined standard (1 σ) uncertainty (%)	Comment
Κα : 27,3						
Κβ: 30,9 – 31,8						
148,2						
158,9						
189,6						
210,3						
254,3						
291,4						
335,8						
346,1						
370,4						
385,9						
400,0	0.0285	2.629252	0.160	38, 1.03, 0.55	38	
443,9	0.0258	2.629252	0.195	36, 1.11, 0.55	36	

Energy (keV)	Efficiency	Correction Factors	Photon	Elementary components	Relative	Comment
	(%)	C5: decay correction (the other	emission	of the uncertainty $(\% -1 \sigma)$	combined	
		corrections C1 C4 were	intensity per	(U1+U2), U3, U7 (the other	standard (1 σ)	
		considered negligible)	disintegrations	components were considered	uncertainty (%)	
			disintegrations			
468,6						
476,85						
481,1	0.0239	2.629252	0.151	34, 1.17, 0.55	34	
498,4						
525,4	0.0219	2.629252	0.051	84, 1.21, 0.55	84	
530,3						
553,8						
571,6						
602,7	0.01896	2.629252	95.1	0.22, 1.26, 0.55	1.39	
632,3						
645,8	0.0176	2.629252	7.40	1.11, 1.27, 0.55	1.77	
662,4						
709,3	0.01586	2.629252	1.46	4.64, 1.28, 0.55	4.84	
713,7	0.01575	2.629252	2.29	3.72, 1.29, 0.55	3.98	
722,7	0.01553	2.629252	10.80	1.19, 1.29, 0.55	1.84	
735,7	0.01522	2.629252	0.22	27, 1.29, 0.55	27	
765,8						

Energy (keV)	Efficiency	Correction Factors	Photon	Elementary components	Relative	Comment
	(%)	C5: decay correction (the other	emission	of the uncertainty (% -1 σ)	combined	
		corrections C1 C4 were	intensity per	(U1+U2), U3, U7 (the other	standard (1σ)	
		considered negligible)	100	components were considered	uncertainty (%)	
			disintegrations	negligible)		
775,2						
790,7	0.01403	2.629252	0.83	8.73, 1.32, 0.55	8.85	
816,8						
856,9						
899,6						
937,9						
968,1	0.01117	2.629252	3.03	3.88, 1.34, 0.55	4.14	
976,2						
997,7						
1014,5						
1045,1	0.01029	2.629252	2.15	5.4, 1.29, 0.55	5.58	
1053,8						
1086,3						
1097						
1163,2						
1198						
1205,5						

Energy (keV)	Efficiency	Correction Factors	Photon	Elementary components	Relative	Comment
	(%)	C5: decay correction (the other	emission	of the uncertainty (% -1 σ)	combined	
		corrections C1 C4 were	intensity per	(U1+U2), U3, U7 (the other	standard (1 σ)	
		considered negligible)	100	components were considered	uncertainty (%)	
			disintegrations	negligible)		
1235						
1253,4						
1263,1						
1269						
1301,3						
1325,5	0.00835	2.629252	1.70	8.55, 1.07, 0.55	8.63	
1355,1	0.008234	2.629252	1.09	11.2, 1.16, 0.55	11.3	
1368,1	0.008187	2.629252	2.83	4.55, 1.21, 0.55	4.74	
1376,1	0.00816	2.629252	0.92	19.2, 1.24, 0.55	19.3	
1385,1	0.00813	2.629252	0.32	40, 1.29, 0.55	40	
1428,7						
1436,5	0.007984	2.629252	1.25	8.6, 1.61, 0.55	8.77	
1445,0	0.007963	2.629252	0.41	24, 1.67, 0.55	24.1	
1453,2						
1488,8	0.007873	2.629252	0.73	13.1, 2.03, 0.55	13.3	
1505,6						
1528,1	0.007815	2.629252	0.49	19.2, 2.41, 0.55	19.4	

Energy (keV)	Efficiency	Correction Factors	Photon	Elementary components	Relative	Comment
	(%)	C5: decay correction (the other	emission	of the uncertainty (% -1 σ)	combined	
		corrections C1 C4 were	intensity per	(U1+U2), U3, U7 (the other	standard (1 σ)	
		considered negligible)	100	components were considered	uncertainty (%)	
			disintegrations	negligible)	• • •	
1557						
1565,8						
1579,7						
1622,4	0.007757	2.629252	0.21	23, 3.5, 0.55	23.3	
1657						
1690,9	0.007785	2.629252	45.6	0.5, 4.44, 0.55	4.50	
1720,3	0.007815	2.629252	0.12	32.7, 4.87, 0.55	33.1	
1757,9	0.007869	2.629252	0.0064	291, 5.45, 0.55	291	
1851,5	0.008082	2.629252	0.28	17.7, 7.04, 0.55	19.1	
1918,8	0.008306	2.629252	0.059	49.4, 8.31, 0.55	50.1	
1950,4						
2015,7						
2039,3						
2078,6						
2090,9						
2099,1						

PART B

¹²⁴Sb – Decay Data Evaluation

¹²⁴Sb - Comments on evaluation of decay data by M.M. Bé and V. Chisté

This evaluation was completed in December 2008. The literature available by this date was included as well as the results obtained as a part of a specific exercise dedicated to the ¹²⁴Sb activity and γ -ray emission intensity measurements organized by the Euramet organisation (Project 907, full report to be published). In the following, the participants in the Euramet 907 project will be referred as E907- *n*, where *n* is a serial number.

1. Decay Scheme

This decay scheme is complete and is based on those proposed by Goswamy (1993Go10), Patil (2006Pa16) and the results obtained in the Euramet-907 project.

A good agreement was found between the effective Q value of 2906 (8) keV computed from the decay scheme data and the adopted Q value of 2904,3 (15) keV from the mass adjustment of Audi *et al.*

2. Nuclear Data

The Q value is from the atomic mass evaluation of Audi et al. (2003Au03).

Reference	T _{1/2}	Uc	Comments
Macklin (1957Ma50)	60,4	0,2	
C.H.Johnson (1958Jo01)	59,9	0,5	
J.P.Cali (1959Ca12)	60,1	0,3	
S.A.Reynolds (1968Re04)	60,3	0,2	
D.M.Fleming (1966Fl01)	60,20	0,03	calorimetry
I.A.Kharitonov (2000Kh04)	60,11	0,07	$4\pi\beta-\gamma$ coincidence method
* E907- 8	60,212	0,011	Ionization chamber
Adopted	60,208	0,011	Reduced $\chi^2 = 1$; critical $\chi^2 = 4,6$

Experimental half-life values (in days) are listed below:

*Euramet 907 participant number 8

The adopted value is the weighted mean of the three most precise values with the external uncertainty.

2.1 Beta transitions

 β - transition energies have been energies are calculated from the Q value and the level energies.

The β - transition probabilities were deduced from the γ transition probability balance at each level of the decay scheme. The adopted values are compared with the measured values in the following table:

	(0,1) 2301 keV	(0,3) 1579	(0,5) 946 keV	(0,10) 610 keV	(0,20) 210
	%	keV	%	%	keV
		%			%
Langer	21	7	9	49	14
(1953La35)					
Moreau	22	7	9	53	9
(1954Mo83)					
Azuma	22	6	4	56	12
(1955Az29)					
Hsue	23	5			
(1965Hs02)					
Zolotavin	28	10	4	49	9
(1956Zo06)					
Adopted	23,44 (28)	4,815 (29)	2,295 (7)	51,21 (19)	8,663 (27)
Nature	1st				
	S(q2+(1p2+16(2)				
	(Hsue)				
	S=k(1-0,25W-				
	0,06/W+0,041W2)(Hsue)				
	$S=q_2+(1p_2+7)$				
	(Calliy) S=0.9a2±n2 (Johnson)				
(1955Az29) Hsue (1965Hs02) Zolotavin (1956Zo06) Adopted Nature	23 23 28 23,44 (28) 1 st S(q2+(1p2+16(2) (Hsue) S=k(1-0,25W- 0,06/W+0,041W2)(Hsue) S=q2+(1p2+7(2) (Canty) S=0,9q2+p2 (Johnson)	5 10 4,815 (29)	4 2,295 (7)	49	9 8,663 (27)

The weak beta transition probabilities are based on the γ transition probability balance at each level of the decay scheme, especially in the upper part of the decay scheme (from level 2886-keV to level 2483-keV) where there are only gamma transitions depopulating these levels. In this evaluation, only the gamma rays observed in several independent experiments have been retained (see § 4.2) so the corresponding levels can be considered definitely established.

2.2 Gamma transitions and internal conversion coefficients

 γ -ray measurements carried out by Doll *et al.* (2000Do11) confirmed the doublet structure of the 2039 level ; one with J^{π} assignment 2⁺ and the second with 3⁺ ; with a spacing of 129 eV.

The γ transitions with energy : 2039,4- ; 790,8- ; 1436,7- ; 713,9-keV start from level with $J^{\pi} = 2^+$ and, those with energy : 790,7- ; 1436,6- ; 713,8-keV from level with $J^{\pi} = 2^+$. They are shown as doublets in the following table.

Internal conversion coefficients

Multipolarity and multipole mixing ratio (δ) for some transitions were determined using the techniques of directional correlation and nuclear orientation measurements, these are summarized in Table 1 :

Transition energy	multipole mixing ratio	Multipolarity	Reference
(keV)	(d)		
444	0,57 (17) or 0,06 (8)		Robinson et al. 1983
646	0,013 (9)	E2, M3	Goswamy et al. 1993
	0,000 (1)		Baker et al. 1972
709	- 0,8 (+3, -4)	M1, E2	Goswamy et al. 1993
	- 1 (+6, -8)		Goswamy et al. 1993
	- 0,18 (5)		Robinson et al. 1983
	0,04 (3, -5)		Grabowski <i>et al.</i> 1971
714	- 0,65 (+38, - 0,54)	M1, E2	Goswamy et al. 1993
	1,15 (16, - 25)		Subrahmanyeswara et al. 1990
	1,5 (7)		Robinson et al. 1983
	1,5 (6)		Baker et al. 1972

Table 1 :

Transition energy	multipole mixing ratio	Multipolarity	Reference
(keV)	(d)		
	0,98 (19)		Grabowski et al. 1971
723	3,74 (12)	M1, E2	Goswamy et al. 1993
	- 3,8 (2)		Subrahmanyeswara et al. 1990
	- 3,4 (3)		Robinson et al. 1983
	- 3,3 (2)		Baker et al. 1972
	- 3,4 (1)		Grabowski <i>et al.</i> 1971
	- 7,5 (20)		Sites et al. 1970
	- 3,4 (6)		Stelson, 1967
791	- 0,15 (+5, -2)	E2, M3	Goswamy et al. 1993
	- 0,3 (+52, -14)		Goswamy et al. 1993
968	0,038 (3)	E1, M2	Goswamy et al. 1993
	- 0,35 (8)		Subrahmanyeswara et al. 1990
	- 0,02 (2)		Robinson et al. 1983
	- 0,03 (6, -5)		Baker et al. 1972
	- 0,02 (8)		Sites et al. 1970
1045	- 0,14 (+3, -4)	E1, M2	Goswamy et al. 1993
	- 0,03 (2)		Robinson et al. 1983
	0,041 (47, -41)		Baker et al. 1972
	- 0,1 (1)		Sites et al. 1970
1356	- 0,32 (+25, -18)	E2, M1	Goswamy et al. 1993
1368	- 0,28 (6)		Subrahmanyeswara et al. 1990
	- 0,02 (1)		Robinson et al. 1983
	- 0,045 (90)		Baker et al. 1972
	- 0,01 (8)		Sites et al. 1970
1376	0,26 (11)	E1, M2	Goswamy et al. 1993
	< 0,29		Goswamy et al. 1993
1.107	- 0,01 (3)	N/4 50	Robinson <i>et al.</i> 1983
1437	0,51 (+13, -11)	M1, E2	Goswamy et al. 1993
	1,5 (8)		Robinson <i>et al.</i> 1983
1 4 4 5	3,7 (27, -20)	F1 M0	Baker et al. 1972
1445	0,015 (80)	E1, M2	Goswamy et al. 1993
1 400	0,10 (9)		Robinson <i>et al.</i> 1983
1489	0,10(23)		Robinson <i>et al.</i> 1983
1001	-3,4(9,-13)	E1 M9	Baker et al. 1972
1691	- 0,009 (22)	E1, M2	Goswamy <i>et al.</i> 1993
	- 0,00 (3) 0.02 (1)		Subrammanyeswara <i>et al.</i> 1990 Pakor et al. 1079
	-0.02(1)		Daket et al. 1972 Sites at al. 1070
2001	U,UU (J) 0.021 (G)	E1 M9	Converse et al. 1970
2031	U,U31 (D) 0 039 (39)		Guswalliy et al. 1995 Subrahmanyasuara et al. 1000
	0,032 (32) 0 00 (2 _3)		Subrallinallyeswala et al. 1990 Bakar at al. 1079
	0,00 (2, -3)		Darei el al. 1972 Sitos at al. 1970
	0,07(0)		SILES EL dI . 1310

Moreover, there are also available two sets of measured values of the conversion electron intensities (Ice_i): by Grigor'eev *et al.* (1968), and by Johnson (1974Jo03). These values as well as their weighted means are summarized in where α_{K602} is the theoretical K conversion coefficient interpolated from Band's tables using the program BrIcc with the "frozen orbital approximation" (Kibédi *et al.* 2008Ki07) for an E2 transition ; Ice_i are the conversion electron intensities, and I γ_i , the relative gamma-ray emission probabilities as summarized in Table 3.

The experimental α_{Ki} conversion coefficients have been compared with the theoretical ICC, the deduced mixing ratios δ are in good agreement with those determined by directional correlation and nuclear orientation measurements summarized in Table 1.

Table 2. Then, the experimental K conversion coefficients α_{Ki} were deduced from the relation :

 $\alpha_{\rm Ki} = \alpha_{\rm K602} \times Ice_i/I\gamma i$

where α_{K602} is the theoretical K conversion coefficient interpolated from Band's tables using the program BrIcc with the "frozen orbital approximation" (Kibédi *et al.* 2008Ki07) for an E2 transition ; Ice_i are the conversion electron intensities, and I γ_i , the relative gamma-ray emission probabilities as summarized in Table 3.

The experimental α_{Ki} conversion coefficients have been compared with the theoretical ICC, the deduced mixing ratios δ are in good agreement with those determined by directional correlation and nuclear orientation measurements summarized in Table 1.
Table 2:

						<mark>α_k (602)=</mark>	0,00420	0,00006							
	John	son	Grigo	r'eev					α _k =						
Energy	lec	Uc	Ice	Uc	Ice WM	Uc dopt.	lg rel.	Uc Ig	Ice/Ig * ak602	$\text{uc} \; \alpha_k$	Multipolarity	delta	%	α_k theo	α_T theo.
159	2,3	0,2			2,3	0,2	0,0050	0,0006	1,93	0,29					
254	0,10	0,08			0,1	0,08	0,0145	0,0009	0,0290	0,0232	E1 ?			0,01269 (18)	0,01465 (21)
336	0,12	0,08			0,12	0,08	0,0741	0,0009	0,0068	0,0045	E1			0,00611 (9)	0,00704 (10)
371	0,1	0,08			0,1	0,08	0,0292	0,0011	0,0144	0,0115					
400	0,45	0,08			0,45	0,08	0,128	0,0027	0,0148	0,0027	E2			0,01323 (2)	0,01566 (2)
444	0,35	0,15			0,35	0,15	0,192	0,009	0,0077	0,0033	M1+E2	0,06	26,5	0,01092 (16)	0,01261 (18)
469	< 0,14				< 0,14		0,0469	0,0027			E1			0,00268 (4)	0,00309 (5)
481	< 0,07				< 0,07		0,0237	0,0032							
525	0,14	0,08			0,14	0,08	0,1484	0,0036	0,0040	0,0023	M1+E2	1	50	0,0066 (3)	0,0077 (3)
602	100		100		100		100		0,00420	0,00006	E2			0,00420 (6)	0,00490 (7)
646	5,4	0,5	6,6	0,3	6,28	0,53	7,591	0,015	0,0035	0,0003	E2+M3	0,006	0,0036	0,00351 (5)	0,00409 (6)
709	1,4	0,5	1,2	0,1	1,21	0,10	1,3941	0,0046	0,0036	0,0003	M1+E2	-0,18	3,1	0,00349 (5)	0,00402 (6)
713	1,6	0,5	1,6	0,2	1,60	0,19	2,325	0,007	0,0029	0,0003	M1+E2	1	50	0,0031 (4)	0,0036 (4)
722	5,7	0,5	7,5	0,3	7,02	0,79	10,952	0,022	0,0027	0,0003	M1+E2	-3,4	92	0,00271 (4)	0,00314 (5)
735	0,04	0,02			0,04	0,02	0,1342	0,0016	0,0013	0,0006					
766	0,035	0,02	0,06	0,02	0,048	0,014	0,0105	0,0009	0,0190	0,0059	E0, M1			0,019 (6)	0,021 (7)
790	0,44	0,08	0,44	0,03	0,440	0,028	0,7584	0,0025	0,0024	0,0002	E2			0,00214 (6)	0,00248 (8)
968	0,24	0,08	0,33	0,03	0,319	0,030	1,93	0,01	0,0007	0,0001	E1(+M2)	-0,2	3,8	0,000569 (9)	0,000653 (11)
1045	0,18	0,08	0,25	0,03	0,241	0,028	1,894	0,014	0,0005	0,0001	E1(+M2)	-0,03	0,09	0,000494 (9)	0,000567 (10)
1325	0,35	0,1	0,30	0,03	0,304	0,029	1,623	0,007	0,0008	0,0001	E2			0,000693 (10)	0,000827 (12)
1355	0,17	0,1	0,20	0,02	0,199	0,020	1,0649	0,0039	0,0008	0,0001	E2(+M3)	-0,32	9,3	0,0009 (5)	0,0011 (5)
1368	0,14	0,05	0,22	0,03	0,199	0,035	2,680	0,008	0,0003	0,0001	E1(+M2)	-0,02	0,04	0,000303 (5)	0,000478 (7)
1376	0,035	0,03			0,035	0,03	0,5113	0,0044	0,0003	0,0002	E1(+M2)	-0,01	0,01	0,000300 (5)	0,000479 (7)
1418	0,25	0,1			0,25	0,1	0	0							
1436	0,28	0,1	0,17	0,03	0,18	0,03	1,262	0,008	0,0006	0,0001	M1+E2	1,5	69,23	0,00063 (5)	0,00078 (5)
1489	0,14	0,1	0,13	0,02	0,13	0,02	0,6924	0,0038	0,0008	0,0001	M1+E2	0,1	0,9901	0,000659 (14)	0,000829 (16)
1526	0,035	0,03	< 0,04		0,035	0,03	0,4232	0,0048	0,0003	0,0003	E1			0,000252 (6)	0,000535 (8)
1657	0,2	0,1			0,2	0,1	0,00	0,00							
1691	2,7	0,4	2,5	0,2	2,54	0,18	48,54	0,19	0,00022	0,00002	E1+M2	0,01	0,01	0,000213 (4)	0,000615 (9)
2090,9	0,24	0,06	0,20	0,04	0,212	0,033	5,618	0,025	0,00016	0,00002	E1(+M2)	0,03	0,1	0,0001522 (23)	0,000838 (12)

3. Atomic Data

The fluorescence yield data are from 1996Sc06 (Schönfeld and Janssen).

3.1 X Radiations

The relative K x-ray emission probabilities are from 1996Sc06.

3.2 Auger Electrons

The ratios P(KLX)/P(KLL) and P(KXY)/P(KLL) are from 1996Sc06.

4. Radiation Emissions

4.1 Electron Emissions

The β - emission energies and intensities were deduced from γ transition probabilities (§ 2.1).

The conversion electron emission intensities have been calculated from the γ -ray emission intensities in sects. 4.2, and the internal-conversion coefficients in sect. 2.2.

The Auger electron emission intensities were calculated by the EMISSION program from PTB using the γ -ray emission probabilities, the atomic data of sect. 3, and the internal-conversion coefficients of sect. 2.2.

4.2 **Photon Emissions**

The X-ray absolute emission intensities were calculated using the EMISSION program and the γ -ray emission intensities, the atomic data given in sect. 3, and the internal-conversion coefficients in sect. 2.2. They are compared with the three sets of absolute values measured by participants in the Euramet exercise. They are, in general, in good agreement.

	E907- 2			- 3	E907	- 8	Calculated			
Energy (keV)	I %	Uc	I %	Uc	I %	Uc	I %	Uc		
27,2 (Kα2)			0,128	0,002	0,130	0,003	0,1252	0,0018		
27,5 (Kα1)			0,264	0,004	0,230	0,006	0,233	0,003		
30,9 (Kβ'2)			0,068	0001	0,063	0,002	0,0667	0,0012		
31,7 (Kβ'1)			0,0170	0,0005	0,0136	0,0006	0,0145	0,0005		
Κα	0,35	0,07	0,392	0,0045	0,359	0,007	0,358	0,0035		
Κβ	0,087	0,018	0,085	0,0011	0,076	0,0018	0,081	0,0013		
K X Total	0,437	0,072	0,476	0,005	0,436	0,007	0,439	0,004		

The X-ray relative emission intensities given by Euramet participants 2 and 8 are compared, in the following table, with the published values of Patil (2006) and Goswamy (1993).

	E907	- 2	E907	- 3	Patil (20	06)	Goswamy (1993)		
Energy (keV)	Rel. Int. Uc		Rel. Int.	Uc	Rel. Int.	Uc	Rel. Int.	Uc	
Κα: 27,3	0,361	0,076	0,4000	0,0046	0,3681	0,0066	0,366	0,017	
Κβ : 30,9 – 31,8	0,089	0,018	0,0864	0,0014	0,0852	0,0017	0,084	0,050	

gray energies

The γ -ray energies in the following table, are from Helmer (2000He14). The other energies were deduced from the level energy differences.

E (keV)	Uc (keV)	E (keV)	Uc (keV)
602,7260	0,0023	1045,125	0,004
645,8520	0,0019	1325,504	0,004
713,776	0,004	1368,157	0,005
722,782	0,003	1436,554	0,007
790,706	0,007	1690,971	0,004
968,195	0,004	2090,930	0,007

gray emission intensities

The 6 participants in the Euramet project sent their γ -ray emission intensities in both relative and absolute scales, since they also carried out activity measurements of the solution.

Moreover, eight sets of measured values published in the literature are available. All of them are relative to the most intense 602-keV γ -ray line (Table 3).

Among the 111 γ rays mentioned before or in this exercise, some weak lines were observed once and not confirmed by other measurements, these are summarized below:

- Weak gamma rays of weak intensities observed by one Euramet participant often described being "barely visible" and then not adopted in the decay scheme :

2871-keV ; 2274-keV ; 2253-keV ; 2151-keV ; 1970-keV (just detection limits) ; 1950-keV ; 1657-keV ; 1557-keV ; 1428-keV ; 1269-keV ; 1202-keV ; 1198-keV ; 1180-keV ; 1163-keV ; 669-keV ;

- Weak gamma rays of weak intensities observed by Patil but by none of the Euramet participant and not adopted in the decay scheme :

2814-keV ; 2746-keV ; 2515-keV ; 2490-keV ; 2386-keV (just detection limits) ; 2373-keV ; 2256-keV ; 2232-keV ; 2145-keV ; 1418-keV ; 795-keV ; 743-keV ; 592-keV ; 186-keV.

A number of weak gamma rays were observed by some Euramet participants or by others :

- 2224-keV, 2204-keV the reported intensities are quite discrepant so they were omitted ;

- 1453-keV could be between levels 2701,6 and 1248,5-keV, but the reported intensities are quite discrepant so this γ -ray was omitted ;

- 476-keV could be between levels 2701,6 and 2224,8-keV, but the reported intensities are quite discrepant so this ray has not been retained ;

- 1757-keV ; 1509-keV ; 1253-keV ; 1097-keV ; 1014-keV ; 937-keV ; 553-keV ; 498,4-keV; 385-keV ; 346-keV ; do not correspond to levels differences, they have not been retained.

- 1235-keV ; 997-keV ; 159,8-keV were accepted but not placed in the decay scheme.

602-keV absolute gray emission intensity

1) A first attempt was made to determine the 602-keV line absolute emission intensity using the results of the absolute measurements carried out in the framework of the Euramet project :

Participant	lg ₅₀₂ in %	Uc
E907- 2	97,5	0,7
E907- 3	97,8	0,9
E907- 5	97,6	0,7
E907- 6	91	1
E907- 7	97,84	0,34
E907- 8	98,1	1,5

The value of participant 6 was found to be an outlier based on Chauvenet's criterion. Value of participant 7 contributes to 58 % to the weighted mean (WM). The set of the five remaining values is consistent, then the evaluated value (LWM) is the weighted mean with the internal uncertainty.

All absolute γ -ray emission intensities measured by the Euramet participants are summarized in Table 4.

2) A second attempt using all the available measurements was done. Since the Euramet participating laboratories sent their results as relative values also, these six sets of results were used as well as the previous measurements published in the literature. So, 14 sets of data were included in the evaluation (Table 3).

In the Euramet project, the participants sent their results as values relative to the reference line $I_{\gamma_{602}} = 100$; with its uncertainty included in the uncertainties of the other γ -ray lines.

In the other publications, when an author gave an uncertainty on this $I_{\gamma_{602}}$ reference line, then this uncertainty was included into each individual value using the relation : Uc = sqrt (Uc_{rel}*Uc_{rel} + Uc_{I_γ602} *Uc_{I_γ602}). So, all gamma rays have been treated with emission intensities relative to $I_{\gamma_{602}} = 100$ (with no uncertainty).

Since no beta transition populating the ground state level in Tellurium 124 is expected, the sum of the gamma transition probabilities with energy 2807-, 2693-, 2681-, 2455-, 2323-, 2294-, 2182-, 2039-, 1657-, 1325-, 602- keV which populate the ground state must be equal to 100. That is:

$$\sum_{i} I_{g_i} \left[1 + \boldsymbol{a}_{T_i} \right] = \frac{100}{N}$$

Where: $I_{\gamma i}$ is the relative emission probability of the gamma-ray, α_{Ti} is its total conversion coefficient, and N is a normalisation factor between the relative and absolute scales.

N, the normalization factor, is then deduced from the measured relative $I_{\gamma i}$ values:

$$N = \frac{100}{\sum_{i} I_{g_i} [1 + \boldsymbol{a}_{T_i}]} \qquad \text{and} \qquad dN^2 = \sum_{i} \left(\frac{\partial N}{\partial I_{g_i}} dI_{g_i} \right)^2 + \sum_{i} \left(\frac{\partial N}{\partial \boldsymbol{a}_{T_i}} d\boldsymbol{a}_{T_i} \right)^2$$

The α_T coefficients are theoretical values interpolated from Band's tables (2002Ba85) using the program BrIcc with the "frozen orbital approximation" (Kibédi *et al.* 2008Ki07). All transitions with a measured multipolarity are E2.

This leads to N = 0.97775 (20).

The absolute emission intensity of the 602-keV gray is then deduced to be: 97,775 (20) %.

This value is in full agreement with the above value of 97,77 (26) %. However, because of the normalization procedure used, its uncertainty is ten times smaller.

Having in mind that the energies of the involved transitions are relatively high and their respective multipolarities are E2, the conversion coefficient values deduced from theoretical calculations can be considered very reliable. Hence, this second absolute intensity value and its associated uncertainty were adopted here.

All the measured relative gamma emission intensities are summarized in Table 3, with the unweighted mean for each set of values given, as well as the weighted mean, the reduced χ^2 and the internal and external uncertainties, the adopted relative emission intensity value and its uncertainty and the deduced and adopted absolute values.

All the absolute gamma-ray emission intensities measured by the participants in the Euramet 907 project are summarized in Table 4. The most intense lines are compared to those obtained from relative values and conversion coefficients (Table 3) in the following table. The agreement is very good.

gray energy keV	From absolute measurements (Table 4)	From relative measurements and ICC (Table 3)
602	97,77 (26)	97,775 (20)
645	7,414 (21)	7,422 (15)
709	1,3635 (43)	1,363 (5)
713	2,269 (11)	2,273 (7)
722	10,712 (31)	10,708 (22)
968	1,880 (6)	1,887 (10)
1045	1,835 (6)	1,852 (14)
1325	1,583 (6)	1,587 (7)
1368	2,615 (9)	2,620 (8)
1690	47,39 (22)	47,46 (19)
2090	5,491 (26)	5,493 (24)

7. References

1953La35 - L.M. LANGER, N.H.Lazar, R.J.D.Moffat. Phys. Rev. 91 (1953) 338. Beta emission probabilities

1954Mo83 - J.Moreau. Comp. Rend. Acad. Sci. (Paris) 239 (1954) 800. Beta emission probabilities

1955Az29 - T.Azuma. J. Phys. Soc. Jpn 10 (1955) 167. Beta emission probabilities

1956Zo06 - A.V.Zolotavin, E.P.Grigoriev, M.A.Abrovian. Izvest.Akad.Nauk SSSR, Ser.Fiz.20 (1956) 289. Columbia Tech.Transl. 20, 271 (1957)

1957Ma50 - R.L.Macklin. Nucl. Instrum. Methods 1 (1957) 335. Half-life

1956Zo06 - A.V.Zolotavin, E.P.Grigoriev, M.A.Abrovian. Izvest.Akad.Nauk SSSR, Ser.Fiz. 20, 289 (1956) ; Columbia Tech.Transl.20 (1957) 271. Beta emission probabilities

1958Jo01 - C.H.Johnson, A.Galonsky, J.P.Ulrich. Phys. Rev.109 (1958) 1243. Half-life

1959Ca12 - J.P.Cali, L.F.Lowe. Nucleonics 17, 10 (1959) 86. Half-life

1965Hs02 - S.T.HSUE, L.M.Langer, S.M.Tang, D.A.Zollman. Nucl. Phys. 73 (1965) 379. Beta emission probabilities

1966FI01 - D.M.FLEMING, I.T. MYERS. Int. J. Appl. Radiat. Isotop. 17 (1966) 251. Half-life

1967ST05 - P.H. Stelson. Phys. Rev. 157 (1967) 1098. ICC

1968Gr24 - E.P.GRIGORIEV, A.V. ZOLOTAVIN, V.O. SERGEEV, M.I. SOVTSOV. Izv. Akad. Nauk SSSR. Ser. Fiz. 32 (1968) 733. K-Conv. Elec. emission probabilities

1968Re04 - S.A.Reynolds, J.F.Emery, E.I.Wyatt. Nucl. Sci. Eng. 32 (1968) 46. Half-life

1968Gr24 - E.P.GRIGORIEV, A.V. ZOLOTAVIN, V.O. SERGEEV, M.I. SOVTSOV. Bull. Ac. Sci. USSR. Phys. Ser. 32 (1968) 711. K-Conv. Elec. emission probabilities

1969Ra31 - R.C.RAGAINI, W.B. WALTERS, R.A. MEYER. Phys. Rev. 187 (1969) 1721. K ICC, Mixing Ratio

1969Me04 - R.A. MEYER, W.B. WALTERS, R.C. RAGAINI. Nucl. Phys. A127 (1969) 595. Gamma-ray emission probabilities, Spin and Parity, Gamma-ray energies

1970Si17 - J.R.SITES, W.A. STEYERT. Nucl. Phys. A156 (1970) 19. Mixing Ratio

1971GR14 - Z.W.Grabowski, K.S.Krane, R.M.Steffen. Phys. Rev. C3 (1971) 1649. Mixing Ratio

1972BA38 - K.R.Baker, J.H.Hamilton, A.V.Ramayya, G.Highland. Nucl. Phys. A186 (1972) 493. Mixing Ratio

1974Jo03 - J.R.JOHNSON, K.C. MANN. Can. J. Phys. 52 (1974) 406. Gamma-ray emission probabilities, K-Conv. Elec. emission probabilities

1979Sh08 - A.K.SHARMA, R.KAUR, H.R. VERNA, K.K. SURI, P.N. TREHAN. J. Phys. Soc. Jap. 46 (1979) 1057. Gamma-ray emission probabilities

1983RO13 - S.J.Robinson, W.D.Hamilton, D.M.Snelling. J. Phys. (London) G9 (1983) 921. Mixing ratio

1984MA13 - G.Mardirosian, N.M.Stewart. Z.Phys. A315, 213 (1984) Gamma-ray emission probabilities

1984Iw03 - Y.IWATA, M.YASUHARA, K.MAEDA, Y.YOSHIZAWA. Nucl. Instrum. Methods 219 (1984) 123. Gamma-ray emission probabilities

1988YO05 - You Jianming, Liu Yunzuo, Hu Dailing. Z. Physik A331 (1988) 391. Gamma-ray emission probabilities 1990ME15 - R.A. Meyer. Fizika 22 (Zagreb) (1990) 153. Gamma-ray emission probabilities

1990Su10 - S.Subrahmanyeswara Rao, K.Bhaskara Rao, V.Seshagiri Rao, H.C.Padhi. Nuovo Cim. 103A (1990) 803. ICC

1993Go10 - J.Goswamy, B.Chand, D.Mehta, N.Singh, P.N.Treha. Appl. Rad. Isotopes 44 (1993) 541. Gamma-ray emission probabilities

1996Sc06 - E. Schönfeld, H. Janssen. Nucl. Instrum. Meth. Phys. Res. A369 (1996) 527. Atomic data

2000Kh04 - I.A.Kharitonov, T.E.Sazonova, S.V.Sepman, T.I.Shilnikova, A.V.Zanevsky. Appl. Rad. Isotopes 52 (2000) 415. Half-life

2000He14 - R.G. Helmer, C. van der Leun. Nucl. Instrum. Methods Phys. Res. A450 (2000) 35. Gamma energy

2000Do11 - C.Doll, H.Lehmann, H.G.Borner, T.von Egidy. Nucl. Phys. A672 (2000) 3. Nuclear structure

2002Ba85 - I.M.Band, M.B.Trzhaskovskaya. At. Data. Nucl. Data Tables 88,1 (2002). Theoretical ICC

2003AU03 - G. Audi, A.H.Wapstra, C. Thibault. Nucl. Phys. A729 (2003) 337-676. Q value

2006Pa16 - A.Patil, D.Santhosh, K.V.Sai, M.Sainath, K.Venkataramaniah. Appl. Rad. Isotopes 64 (2006) 693. Gamma-ray emission probabilities

2008Ki07 - T.Kibédi, T.W.Burrows, M.B.Trzhaskovskaya, P.M.Davidson, C.W.Nestor, Jr. Nucl. Instrum. Methods Phys. Res. A589 (2008) 202. ICC

(14,12) (20,14) (23,14) 148keV (-1, 1) 159keV (16.12)186keV (14,10) 189 keV 209keV (10.6) 254keV 291keV Uc Value Value Uc Value Uc Value Uc Value Uc Uc Value Uc Value ^(o) 0.0088 E907-2 0,005 0,001 0,0012 0,001 0,0046 0,0008 0,012 0,005 0,001 0,002 0,009 0.002 DL=0.0053 E907-3 DL=0,0041 DL=0,0043 DL=0,0031 DL=0,0032 DL=0,0042 0,014 E907-5 E907-6 E907-7 0.0047 0.0159 0.0092 0.0053 0.0012 0.0070 0.0014 0.010 0.006 0.0024 0.0015 0.0010 E907-8 0,0028 0,0008 0,0045 0,0007 0,0049 0.0005 0.0054 0,0010 0.0165 0,0014 0.0059 0,0012 ^(o) 0,0147 Patil (2006Pa16) 0,0020 0.0036 0.0005 0,0137 0.0006 0,0070 0.0006 0,0008 Goswamy (1993Go10) 0,0037 0,0007 0,0037 0.0007 0,0055 0,0010 0,0163 0.0008 8800,0 Jianming (1988Yo05) 0,006 0,002 0,006 0.002 0,0062 0.0028 0.0214 0,0041 0,012 0,006 ^(o) 0,030 Mardirosian (1984Ma13) 0,007 Iwata (1984Iw03) Johnson (1974Jo03) Meyer (1990Me15) Sharma (1979Sh08) Chi2 1,4 1,2 1,6 0,1 4,3 4,0 Chi2 crit: 3.8 4.6 3.3 3,8 2.8 3,0 UWM: 0,00552 0,00449 0,00530 0,00546 0,01520 0,00795 WM: 0,00543 0,00713 0,00382 0,00504 0,00441 0,01447 Uc (int): 0,00047 0.00054 0,00039 0.00066 0.00041 0.00036 0,00086 0,00073 Uc (ext) : 0,00057 0,00060 0,00049 0,00015 LWM : 0,0071 0,0038 0,0006 0,0050 0.0006 0,00441 0.00049 0,0054 0,0007 0,0145 0,0009 0,0007 I Abs.* 0,0037 0,0043 0,0069 0,0006 0,0049 0,0006 omitted 0.0005 0,0053 0,0007 0,0142 0,0009 0,0007

Table 3 : Relative gamma ray intensities and absolute values calculated with $(*)$ Ig602 = 97,775 (20) %.	
(i, j) refers to initial and final levels, (-1, n) transition not placed in the decay scheme. DL = Detection Li	mi

^(o) Outlier

	(10,5)	336 keV		346,5 keV	(20,11)	371 keV		385 keV	(20,10)	400 keV	(14,6)	444 keV	(20,9)	469 keV
	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc
E907-2	0,079	0,008	0,0034	0,0016	0,034	0,011	0,038	0,026	0,128	0,008	0,190	0,004	0,053	0,009
E907- 3	0,073	0,004	DL=0,0064		0,033	0,006	DL=0,0078		0,120	0,006	0,198	0,006	0,038	0,003
E907- 5	0,072	0,021			>0,0217	<0,0338			0,146	0,011	0,198	0,011	0,045	0,006
E907- 6									0,175	0,066	0,211	0,076		
E907-7	0,0733	0,0016	0,0018	0,0018	0,0295	0,0027	DL=0,0024		0,130	0,007	0,1981	0,0024	0,0518	0,0028
E907- 8	0,0708	0,0026	0,0036	0,0025	0,0333	0,0022			0,1246	0,0037	0,1901	0,0047	0,0449	0,0021
Patil (2006Pa16)	0,076	0,002			0,0257	0,0015			0,125	0,007	0,1830	0,0021	0,0364	0,0023
Goswamy (1993Go10)	0,0750	0,0021	0,0060	0,0013	0,034	0,008			0,124	0,013	0,1920	0,0028	0,047	0,003
Jianming (1988Yo05)	^(o) 0,086	0,006	0,013	0,005	0,036	0,006			0,155	0,013	0,204	0,010	0,053	0,003
Mardirosian (1984Ma13)	0,078	0,007			0,024	0,006			0,168	0,012	0,226	0,015	^(o) 0,079	0,005
lwata (1984lw03)					^(o) 0,051	0,009			0,129	0,016	0,205	0,010	0,058	0,008
Johnson (1974Jo03)					0,03	0,01			0,132	0,015	0,173	0,015	0,031	0,010
Meyer (1990Me15)									0,15	0,01	0,20	0,01		
Sharma (1979Sh08)					0,0315	0,0025			^(o) 0,215	0,006	0,221	0,006	0,064	0,003
Chi2	0,5		1,3		1,4				2,2		4,7		7,4	
Chi2 crit:	2,6		3,8		2,4				2,2		2,1		2,3	
UWM:	0,07459		0,00369		0,03103				0,13890		0,19929		0,04749	
WM:	0,07407		0,00414		0,02925				0,12934		0,19237		0,04685	
Uc (int):	0,00094		0,00083		0,00097				0,00219		0,00116		0,00098	
Uc (ext) :	0,00069		0,00096		0,00115				0,00323		0,00252		0,00267	
LWM :	0,0741	0,0009	0,0041	0,0010	0,0292	0,0011			0,1293	0,0032	0,199	^(e) 0,016	0,0469	0,0027
I Abs.*	0,0725	0,0009	omitted		0,0286	0,0011	omitted		0,1264	0,0031	0,195	0,016	0,0459	0,0026

Table 3 (Con't) : Relative gamma ray intensities and absolute Values calculated with ^(*) Ig602 = 97,775 (20)	%.
(i, j) refer to initial and final levels, $(-1, n)$ transition not placed in the decay scheme. $DL = Detection Limit$	it

^(o) Outlier ^(e) expanded uncertainty so range to include the most precise Value

	?(21,9)	476 keV	(23,10)	481 keV		498 keV	(14, 5)	525 keV	(26, 12)	530 keV		553 keV	(26, 10)	572 keV
	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc
E907- 2	0,046	0,017	0,024	0,007	0,038	0,014	0,1428	0,005	0,043	0,006	0,019	0,005	0,020	0,004
E907- 3	DL=0,0069		^(o) 0,015	0,006			0,140	0,005	0,022	0,003			0,013	0,004
E907- 5			>0,0197	<0,0298			0,182	0,009						
E907- 6			^(o) 0,163	0,055			^(o) 0,055	0,046						
E907- 7	DL=0,0018		0,0253	0,0014	DL=0,0018		0,140	0,005	0,0281	0,0012	0,0019	0,0008	0,0153	0,0017
E907- 8	0,0020	0,0009	0,0269	0,0014	0,0007	0,0005	0,1451	0,0034	0,0431	0,0015				
Patil (2006Pa16)			0,0205	0,0010			0,1429	0,0076	0,0421	0,0013			0,0184	0,0010
Goswamy (1993Go10)			0,024	0,0020			0,14	0,02	0,043	0,002			0,0193	0,0013
Jianming (1988Yo05)			0,029	0,0080			0,165	0,010	0,047	0,011			0,025	0,010
Mardirosian (1984Ma13)			0,030	0,005			0,178	0,012						
lwata (1984lw03)							0,117	0,012						
Johnson (1974Jo03)							0,132	0,010						
Meyer (1990Me15)							0,16	0,01						
Sharma (1979Sh08)							0,162	0,004						
Chi2	3,5		3,1				4,5		20,6				1,3	
Chi2 crit:	6,6		2,8				2,2		2,8				3,0	
UWM:	0,02402		0,02567				0,14975		0,03828				0,01843	
WM:	0,02402		0,02367				0,14837		0,03675				0,01799	
Uc (int):	0,01171		0,00065				0,00168		0,00068				0,00070	
Uc (ext) :	0,02198		0,00115				0,00357		0,00310				0,00080	
LWM :	0,024	0,022	0,0237	0,0032			0,1484	0,0036	0,037	^(e) 0,009			0,018	0,0008
I Abs.*	omitted		0,0232	0,0031	omitted		0,1451	0,0035	0,036	0,009	omitted		0,0176	0,0008

Table 3 (Cont'd) : Relative gamma ray intensities and absolute Values calculated with (*)Ig602 = 97,775 (20)	%
(i, j) refer to initial and final levels, $(-1, n)$ transition not placed in the decay scheme. $DL = Detection Limit$	

^(o) Outlier ^(e) expanded uncertainty so range to include the most precise Value

		592 keV	(1, 0)	602 keV	(5, 3)	632 keV	(2, 1)	646 keV	(21, 6)	662 keV		669 keV	(5, 2)	709 keV
	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc
E907- 2			100		0,100	0,008	7,57	0,07	0,041	0,004			1,358	0,019
E907- 3			100		0,098	0,004	7,59	0,10	DL=0,0063				1,388	0,016
E907- 5			100		0,109	0,007	7,603	0,027	<0,0157				1,396	0,007
E907- 6			100				7,69	0,14					1,484	0,072
E907- 7			100		0,1073	0,0010	7,58	0,03	0,0139	0,0009			1,397	0,006
E907- 8			100		0,1053	0,0028	^(o) 7,35	0,16	0,0227	0,0012	0,180	0,004	1,36	0,03
Patil (2006Pa16)	0,014	0,002	100		0,0990	0,0013	7,69	0,09	0,0148	0,0010			1,39	0,02
Goswamy (1993Go10)			100		0,1070	0,0015	7,55	0,11	0,032	0,002			1,34	0,02
Jianming (1988Yo05)			100		0,101	0,006	7,55	0,13	0,035	0,011			1,38	0,04
Mardirosian (1984Ma13)			100		0,118	0,007	^(o) 7,82	0,22	0,043	0,005			1,49	0,07
Iwata (1984Iw03)			100		0,114	0,006	7,61	0,04	0,016	0,005			1,399	0,012
Johnson (1974Jo03)			100		0,12	0,03	7,53	0,16	0,015	0,003			1,38	0,09
Meyer (1990Me15)			100		0,10	0,01	7,55	0,05					1,38	0,02
Sharma (1979Sh08)			100		0,111	0,003	7,52	0,15	0,0148	0,0015			1,465	0,029
Chi2					3,4		0,3		18,7				1,6	
Chi2 crit:					2,2		2,2		2,4				2,1	
UWM:					0,10692		7,5861		0,02480				1,4008	
WM:					0,10524		7,5911		0,01790				1,3941	
Uc (int):					0,00064		0,0152		0,00050				0,0036	
Uc (ext) :					0,00118		0,0084		0,00217				0,0046	
LWM :					0,1052	^(e) 0,0021	7,591	0,015	^(u) 0,025	^(e) 0,011			1,394	0,005
I Abs.*	Omitted		97,775	0,020	0,1029	0,0021	7,422	0,015	0,024	0,011	omitted		1,363	0,005

Table 3 (Cont'd) : Relative gamma ray intensities and absolute Values calculated with (*)Ig602 = 97,775 (20) %.
(i, j) refer to initial and final levels, $(-1, n)$ transition not placed in the decay scheme. $DL = Detection Limit$

^(o) Outlier ^(e) expanded uncertainty so range to include the most precise Value ^(u) unweighted mean

	(6, 3)	713 keV	(3, 1)	722 keV	(23, 6)	735 keV		743 keV	(7, 3)	766 keV	(25, 6)	775 keV	(6, 2)	790 keV
	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc
E907- 2	2,26	0,03	10,81	0,10	0,137	0,005			0,012	0,003	0,0104	0,0017	0,756	0,008
E907- 3	2,324	0,026	10,95	0,13	0,132	0,006			DL=0,0072		0,000		0,753	0,012
E907- 5	2,327	0,012	10,950	0,037	0,125	0,013			>0,0177	<0,0268			0,756	0,008
E907- 6	2,33	0,09	10,95	0,20	^(o) 0,22	0,06							^(o) 0,824	0,073
E907-7	2,33	0,01	10,96	0,04	0,1338	0,0016			0,0080	0,0012	0,0097	0,0005	0,758	0,004
E907- 8	2,26	0,05	10,73	0,24	0,1245	0,0030			0,0089	0,0014	0,0100	0,0014	^(o) 0,733	0,016
Patil (2006Pa16)	2,29	0,03	10,88	0,16	0,1399	0,0024	0,0058	0,0011	^(o) 0,0039	0,0003	0,0119	0,0012	0,766	0,012
Goswamy (1993Go10)	2,27	0,04	10,77	0,18	0,129	0,002			0,0124	⁽ⁱ⁾ 0,0002	0,0093	0,0018	0,752	0,012
Jianming (1988Yo05)	2,29	0,05	10,99	0,19	0,145	0,021			0,0092	0,0041	0,0112	0,0041	0,753	0,013
Mardirosian (1984Ma13)	2,46	0,09	^(o) 11,46	0,16	0,142	0,005			0,009	0,005	0,0112	0,0041	0,766	0,008
lwata (1984lw03)	2,338	0,015	11,02	0,06	0,133	0,009							0,758	0,009
Johnson (1974Jo03)	2,43	0,10	11,16	0,20	0,14	0,03							0,763	0,015
Meyer (1990Me15)	2,32	0,03	11,0	0,2	0,14	0,01							0,76	0,01
Sharma (1979Sh08)	2,42	0,05	^(o) 11,31	0,22	0,146	0,004							0,734	0,016
Chi2	1,4		0,6		2,8				2,2		0,5		0,2	
Chi2 crit:	2,1		2,2		2,2				3,0		2,8		2,3	
UWM:	2,3317		10,9300		0,1359				0,00986		0,01052		0,75832	
WM:	2,3250		10,9525		0,1342				0,01053		0,01002		0,75842	
Uc (int):	0,0061		0,0224		0,0010				0,00059		0,00041		0,00247	
Uc (ext) :	0,0072		0,0176		0,0016				0,00089		0,00030		0,00120	
LWM :	2,325	0,007	10,952	0,022	0,1342	0,0016			0,0105	0,0009	0,01002	0,00041	0,7584	0,0025
I Abs.*	2,273	0,007	10,708	0,022	0,1312	0,0016	omitted		0,0103	0,0009	0,0098	0,0004	0,7415	0,0024

Table 3 (Cont'd) : Relative gamma ray intensities and absolute Values calculated with $^{(*)}Ig602 = 97,775$ (20) %. (i, j) refer to initial and final levels, (-1, n) transition not placed in the decay scheme. DL = Detection Limit

^(o) Outlier

⁽ⁱ⁾ This original uncertainty was increased in order to limit the relative weight to 50%

		795 keV	(23, 5)	817 keV	(8, 3)	856 keV	(9, 3)	899 keV		937 keV	(10, 3)	968 keV	(9, 2)	976 keV
	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc
E907- 2			0,081	0,007	0,0203	0,006	0,023	0,009	0,0206	0,005	1,907	0,055	0,084	0,005
E907- 3			0,074	0,007	0,017	0,006	0,026	0,016	DL=0,0085		1,921	0,024	^(o) 0,095	0,011
E907- 5			0,076	0,013	>0,0187	<0,0288	<0,0187				1,909	0,013	0,088	0,013
E907- 6											^(o) 2,857	0,118		
E907-7			0,0735	0,0037	0,0228	0,0007	0,0175	0,0010	DL=0,0012		1,926	0,008	0,0862	0,0011
E907- 8			0,0745	0,0021	0,0243	0,0017	0,0176	0,0015	0,0032	0,0012	1,873	0,042	0,0833	0,0023
Patil (2006Pa16)	0,0368	0,0012			0,0216	0,0011	0,020	0,001			^(o) 2,105	0,031	0,0841	0,0013
Goswamy (1993Go10)			0,074	0,002	0,024	0,001	0,0175	0,0014			1,92	0,028	0,0845	0,0019
Jianming (1988Yo05)			0,074	0,007	0,032	0,006	0,020	0,006			1,945	0,030	0,088	0,005
Mardirosian (1984Ma13)			0,086	0,008	0,027	0,006					2,038	0,024	0,088	0,012
Iwata (1984Iw03)			0,079	0,006	0,029	0,007	0,016	0,009			1,919	0,015	0,088	0,008
Johnson (1974Jo03)			^(o) 0,065	0,006	0,022	0,006	0,011	0,004			2,03	0,04	^(o) 0,102	0,020
Meyer (1990Me15)											1,93	0,03	0,09	0,01
Sharma (1979Sh08)			0,083	0,003	0,029	0,003	0,028	0,004			2,03	0,04	^(o) 0,097	0,004
Chi2			1,1		1,2		1,4		5,3		3,5		0,4	
Chi2 crit:			2,4		2,3		2,4		6,6		2,2		2,4	
UWM:			0,0775		0,02447		0,01962		0,0119		1,9457		0,08639	
WM:			0,0761		0,02315		0,01825		0,0119		1,9304		0,08512	
Uc (int):			0,0012		0,00048		0,00059		0,0038		0,0053		0,00070	
Uc (ext) :			0,0012		0,00052		0,00071		0,0087		0,0099		0,00042	
LWM :			0,0761	0,0012	0,0232	0,0005	0,0183	0,0007	0,012	0,009	1,93	0,01	0,0851	0,0007
I Abs.*	omitted		0,0744	0,0012	0,0227	0,0005	0,0179	0,0007	omitted		1,887	0,010	0.0832	0,0007

Table 3 : Relative gamma ray intensities and absolute Values calculated with $(*)$ Ig602 = 97,775 (20) %
(i, j) refer to initial and final levels, (-1, n) transition not placed in the decay scheme. DL = Detection Limit

^(o) Outlier

	(-1,2)	997 keV		1014 keV	(10, 2)	1045 keV	(4, 1)	1053 keV	(12, 2)	1086 keV	1097 keV			1163 keV
	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc
E907-2	0,025	0,007			1,884	0,036			0,041	0,004	0,034	0,008		
E907-3	DL=0,0091		DL=0,0093		1,867	0,024	DL=0,0097		0,042	0,009			DL=0,0108	
E907-5					1,861	0,017			0,050	0,008				
E907-6					2,00	0,11								
E907-7	0,0014	0,0014	0,0025	0,0025	1,880	0,008	0,0026	0,0026	0,0369	0,0012	DL=0,0019		DL=0,0019	
F007 0	0.0040	(1)	0.0040	0.001.4	1 0 1 1	0.044	0.0000	0.0040	0.0000	0.004.0	0.0000	0.0040	0.0000	
E907-8	0,0046	0,0009	0,0046	0,0014	1,841	0,041	0,0036	0,0012	0,0368	0,0018	0,0026	0,0012	0,0033	
Palli (2006Pa16)	-				2,020	0,022	0.005	0.000	0,0356	0,0016				
Goswamy (1993G010)					1,87	0,03	0,005	0,002	0,038	0,002				
Jianming (19881005)					1,90	0,03	0.007	0.004	⁽⁰⁾ 0,043	0,005				
Mardirosian (1984Ma13)					2,01	0,02	0,007	0,001	0,058	0,005				
Iwata (1984Iw03)					1,86	0,02			0,038	0,009				
Johnson (1974Jo03)					1,92	0,04			0,031	0,005				
Meyer (1990Me15)	-				1,88	0,04								
Sharma (1979Sh08)	-				1,97	0,04			0,046	0,004				
Chi2	5,8		0,5		6,2		1,9		1,2					
Chi2 crit:	4,6		6,6		2,1		3,8		2,3					
UWM:	0,01033		0,00354		1,9123		0,00457		0,03985					
WM:	0,00343		0,00408		1,8936		0,00538		0,03739					
Uc (int):	0,00099		0,00124		0,0053		0,00070		0,00074					
Uc (ext) :	0,00238		0,00088		0,0133		0,00097		0,00081					
LWM :	0,0034	0,0024	0,0041	0,0012	1,894	0,014	0,0054	0,0010	0,0374	0,0008				
I Abs.*	0,0033	0,0023	Omitted		1,852	0,014	0,0053	0,0010	0,0366	0,0008	omitted		omitted	

Table 3 (Cont'd) : Relative gamma ray intensities and absolute Values calculated with $(*)$ Ig602 = 97,775 (20) %
(i, j) refer to initial and final levels, $(-1, n)$ transition not placed in the decay scheme. DL = Detection Limit

 $^{\rm (o)}$ Outlier $^{\rm (i)}$ This original uncertainty was increased in order to limit the relative weight to 50%

Table 3 : Relative gamma ray intensities and absolute Values calculated with $^{(*)}Ig602 = 97,775$ (20) % (i, j) refer to initial and final levels, (-1, n) transition not placed in the decay scheme. DL = Detection Limit

		1180 keV		1198 keV		1205 keV	(-1,3)	1235 keV		1253 keV	(15, 2)	1263 keV		1269 keV
	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc
E907-2							0,028	0,006	0,042	0,009	0,043	0,004		
E907- 3			DL=0,0112		DL=0,012				DL=0,0117		0,030	0,010	DL=0,0118	
E907- 5											0,031	0,008		
E907-6														
E907- 7			DL=0,002		DL=0,016		0,0047	⁽ⁱ⁾ 0,0010	DL=0,0019		0,0413	0,0015	DL=0,0019	
E907- 8	0,630	0,014	0,0031	0,0009	0,0314	0,0012	0,0094	0,0012			0,0382	0,0018	0,0037	0,0013
Patil (2006Pa16)											0,0482	0,0015		
Goswamy (1993Go10)											0,042	0,002		
Jianming (1988Yo05)											0,043	0,005		
Mardirosian (1984Ma13)											0,054	0,010		
Iwata (1984Iw03)											0,046	0,015		
Johnson (1974Jo03)											0,045	0,010		
Meyer (1990Me15)														
Sharma (1979Sh08)											0,057	0,005		
Chi2							9,8				3,1			
Chi2 crit:							4,6				2,2			
UWM:							0,0141				0,0432			
WM:							0,0075				0,0432			
Uc (int):							0,00086				0,0008			
Uc (ext) :							0,00269				0,0014			
LWM :							0,0075	0,0027			0,0432	^(e) 0,0019		
I Abs.*	Omitted		omitted		omitted		0,0073	0,0026	omitted		0,0422	0,0019	omitted	

^(e) expanded uncertainty so range to include the most precise Value ⁽ⁱ⁾ This original uncertainty was increased in order to limit the relative weight to 50%

	(17, 2)	1301 keV	(3, 0)	1325 keV	(5, 1)	1355 keV	(20, 3)	1368 keV	(21, 3)	1376 keV	(22, 3)	1385 keV		1418 keV
	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc
E907- 2	0,032	0,004	1,637	0,033	1,066	0,031	2,650	0,045	0,521	0,021	0,072	0,006	l	
E907- 3	0,047	0,010	1,599	0,027	1,059	0,022	2,628	0,034	0,481	0,016	0,051	0,012	l	
E907- 5	0,037	0,009	1,603	0,016	1,055	0,022	2,686	0,017	0,505	0,009	0,064	0,008	l	
E907- 6			1,582	0,137	1,011	0,114	2,65	0,13	0,516	0,099	^(o) 0,20	0,08	l	
E907-7	0,0339	0,0021	1,621	0,007	1,062	0,004	2,682	0,011	0,5130	⁽ⁱ⁾ 0,0034	0,070	0,002	l	
E907- 8	0,037	0,003	^(o) 1,768	0,040	1,070	0,024	2,633	0,061	0,493	0,011	0,060	0,002	l	
Patil (2006Pa16)	0,0256	0,0013	1,707	0,026	1,093	0,017	2,7	0,034	0,543	0,007	0,064	0,002	0,005	0,002
Goswamy (1993Go10)	0,035	0,001	1,61	0,03	1,05	0,015	2,64	0,04	0,493	0,008	0,062	0,003	l	
Jianming (1988Yo05)	0,039	0,005	1,645	0,028	1,103	0,021	2,696	0,041	0,496	0,011	0,071	0,006	l	
Mardirosian (1984Ma13)	^(o) 0,061	0,008	1,69	0,29	1,108	0,022	2,758	0,069	0,531	0,046	0,079	0,025	l	
Iwata (1984Iw03)	0,041	0,015	1,584	0,023	1,042	0,027	2,67	0,03	0,50	0,02	0,061	0,026	l	
Johnson (1974Jo03)			1,67	0,04	^(o) 1,14	0,04	2,76	0,06	0,54	0,03	^(o) 0,03	0,01	l	
Meyer (1990Me15)			1,66	0,04	1,06	0,04	2,68	0,05	0,51	0,04			l	
Sharma (1979Sh08)	0,045	0,004	1,71	0,04	1,17	0,02	^(o) 2,82	0,06	^(o) 0,572	0,012	0,053	0,003		
													Ļ	
Chi2	5,5		2,0		1,1		0,7		2,9		3,5		l	
Chi2 crit:	2,4		2,2		2,2		2,2		2,2		2,3		l	
UWM:	0,0372		1,6399		1,06493		2,6794		0,51101		0,06434		l	
WM:	0,0327		1,6233		1,06491		2,6796		0,51128		0,06337		l	
Uc (int):	0,0007		0,0051		0,00363		0,0076		0,00258		0,00096		l	
Uc (ext) :	0,0017		0,0073		0,00387		0,0063		0,00438		0,00180		 	
LWM :	^(u) 0,0372	^(e) 0,0022	1,623	0,007	1,0649	0,0039	2,680	0,008	0,5113	0,0044	0,063	^(e) 0,006	L	
I Abs.*	0,0364	0,0022	1,587	0,007	1,0412	0,0038	2,620	0,008	0,4999	0,0043	0,062	0,006	omitted	

Table 3 : Relative gamma ray intensities and absolute Values calculated with $(*)$ Ig602 = 97,775 (20) %.
(i, j) refer to initial and final levels, $(-1, n)$ transition not placed in the decay scheme. $DL = Detection Limit$

 $^{\rm (o)}$ Outlier $^{\rm (e)}$ expanded uncertainty so range to include the most precise Value $^{\rm (u)}$ unweighted mean

	1428 keV		(6, 1)	1436 keV	(20, 2)	1445 keV	(21, 2) ?	1453 keV	(7, 1)	1489 keV		1509 keV	(23, 2)	1526 keV
	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc
E907- 2	0,049	0,007	1,253	0,026	0,336	0,008	0,032	0,007	0,686	0,018	0,052	0,016	0,421	0,019
E907- 3			1,266	0,022	0,309	0,014	DL=0,0163		0,684	0,016			0,404	0,012
E907- 5			1,244	0,031	0,350	0,012			0,693	0,015			0,443	0,016
E907- 6			1,19	0,10	0,38	0,09			0,71	0,09			0,43	0,08
E907- 7	DL=0,0026		1,257	⁽ⁱ⁾ 0,005	0,336	⁽ⁱ⁾ 0,002	DL=0,0027		0,700	0,007			0,4184	⁽ⁱ⁾ 0,0026
E907- 8	0,0276	⁽ⁱ⁾ 0,0017	1,313	0,030	0,384	0,009	0,080	0,002	0,667	0,015	0,0074	0,0025	0,398	0,010
Patil (2006Pa16)			1,27	0,017	0,335	0,032			0,692	0,009	0,008	0,001	0,451	0,006
Goswamy (1993Go10)			1,25	0,016	0,334	0,005			0,687	0,009			0,414	0,006
Jianming (1988Yo05)			1,236	0,021	0,346	0,011			0,71	0,05			0,434	0,010
Mardirosian (1984Ma13)			1,34	0,27	0,329	0,014			0,72	0,02			0,433	0,008
lwata (1984lw03)			1,225	0,024	0,358	0,017			0,68	0,02			0,41	0,02
Johnson (1974Jo03)			1,38	0,04	0,30	0,03			0,70	0,03			0,45	0,02
Meyer (1990Me15)			1,26	0,05	0,34	0,04			0,71	0,03			0,41	0,03
Sharma (1979Sh08)			1,37	0,03	0,41	0,01			^(o) 0,80	0,02			^(o) 0,49	0,01
Chi2	4,9		2,5		7,2		21,4		0,7		3,7		3,4	
Chi2 crit:	6,6		2,1		2,1		6,6		2,2		4,6		2,2	
UWM:	0,0383		1,2748		0,3460		0,0561		0,6959		0,0225		0,4241	
WM:	0,0383		1,2619		0,3423		0,0561		0,6924		0,0081		0,4232	
Uc (int):	0,0049		0,0051		0,0021		0,0052		0,0038		0,0009		0,0022	
Uc (ext) :	0,0107		0,0080		0,0057		0,0241		0,0031		0,0018		0,0040	
LWM :	0,038	0,011	1,262	0,008	0,342	^(e) 0,007	0,056	0,024	0,6924	0,0038	0,0081	0,0018	0,423	0,005
I Abs.*	omitted		1,234	0.008	0,334	0,007	omitted		0,677	0,0037	omitted		0,414	0,005

Table 3 (Cont'd) : Relative gamma ray intensities and absolute Values calculated with $^{(*)}Ig602 = 97,775$ (20) %. (i, j) refer to initial and final levels, (-1, n) transition not placed in the decay scheme. DL = Detection Limit

^(o) Outlier
 ^(e) expanded uncertainty so range to include the most precise Value
 ⁽ⁱ⁾ This original uncertainty was increased in order to limit the relative weight to 50%

	1557 keV		(25, 2)	1565 keV	(8, 1)	1580 keV	(9, 1)	1622 keV	(4, 0)	1657 keV	(10, 1)	1691 keV	(11, 1)	1720 keV
	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc
E907- 2					0,441	0,018	0,041	0,003			46,72	1,16	0,098	0,005
E907- 3			DL=0,0105		0,422	0,015	0,043	0,008	DL=0,0089		48,08	0,57	^(o) 0,090	0,004
E907- 5			>0,0197	<0,0298	0,414	0,008					48,28	0,21	0,100	0,007
E907- 6							^(o) 0,22	0,05			49,12	0,94	^(o) 0,135	0,044
E907- 7	DL=0,0017		0,012	0,001	^(r) 0,145	0,001	0,041	0,001	DL=0,0012		48,70	0,18	0,0967	0,0007
E907- 8	0,014	0,007	0,006	⁽ⁱ⁾ 0,001	0,354	0,009	0,042	0,001	0,0086	0,0034	46,35	1,13	0,0963	0,0025
Patil (2006Pa16)					0,460	0,006	0,0477	0,0013			46,63	0,65	0,097	0,0180
Goswamy (1993Go10)			0,015	0,004	0,427	0,007	0,042	0,001			49,32	0,74	0,096	0,0022
Jianming (1988Yo05)			0,013	0,004	0,42	0,04	0,040	0,004			48,73	0,78	0,102	0,0041
Mardirosian (1984Ma13)					^(r) 0,238	0,007	0,047	0,004			50,88	0,88	0,101	0,005
Iwata (1984Iw03)					^(r) 0,155	0,012	0,035	0,012			48,58	0,25	0,097	0,0070
Johnson (1974Jo03)					^(r) 0,15	0,05	^(o) 0,03	0,01			51,3	1,0	0,096	0,007
Meyer (1990Me15)					0,42	0,03					48,4	0,8		
Sharma (1979Sh08)					0,49	0,01	0,047	0,003			50,6	1,0	^(o) 0,104	0,003
Chi2			2,9		0,4		3,0				3,0		0,3	
Chi2 crit:			3,8		3,3		2,4				2,1		2,4	
UWM:			0,0114		0,4203		0,0425				48,692		0,09794	
WM:			0,0111		0,4217		0,0425				48,545		0,09684	
Uc (int):			0,0007		0,0047		0,0005				0,108		0,00063	
Uc (ext) :			0,0012		0,030		0,0009				0,186		0,00035	
LWM :			0,0111	0,0012	0,422	0,005	0,0425	^(e) 0,0019			48,54	0,19	0,0968	0,0006
I Abs.*	omitted		0,0109	0,0012	0,412	0,005	0,0416	0,0019			47,46	0,19	0,0946	0,0006

Table 3 : Relative gamma ray intensities and absolute Values calculated with $^{(*)}Ig602 = 97,775$ (20) %. (i, j) refer to initial and final levels, (-1, n) transition not placed in the decay scheme. DL = Detection Limit

⁽ⁱ⁾ This original uncertainty was increased in order to limit the relative weight to 50%
 ^(r) Removed from analysis
 ^(o) Outlier

^(e) expanded uncertainty so range to include the most precise Value Table 3 (Cont'd) : Relative gamma ray intensities and absolute Values calculated with ^(*)Ig602 = 97,775 (20) %.

(i, j) refer to initial and final levels, (-1, n) transition not placed in the decay scheme. DL = Detection Limit

		1757 keV	(13, 1)	1852 keV	(16, 1)	1918 keV		1950 keV		1970 keV	(18, 1)	2016 keV	(6, 0)	2039 keV
	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc
E907-2					0,056	0,005					0,013	0,002	0,0633	0,004
E907- 3			DL=0,0077		0,051	0,003			DL=0,016		0,008	0,002	0,064	0,003
E907- 5					0,054	0,008							0,064	0,006
E907- 6	0,007	0,021	^(o) 0,341	0,061	^(o) 0,077	0,038								
E907-7	DL=0,0009		0,0054	0,0006	0,0537	0,0005	DL=0,0006				0,0092	⁽ⁱ⁾ 0,0003	0,0636	0,0006
E907- 8			0,0008	⁽ⁱ⁾ 0,0001	0,0529	0,0019	0,053	0,011			0,0098	0,0011	^(o) 0,0753	0,0020
Patil (2006Pa16)			0,0026	0,0001	0,058	0,016					0,0090	0,0009	0,0661	0,0020
Goswamy (1993Go10)	0,0049	0,0023	0,0062	0,0009	0,055	0,002					0,0112	0,0010	0,066	0,0021
Jianming (1988Yo05)			^(o) 0,0112	0,0031	0,06	0,03					0,0124	0,0007	0,068	0,0021
Mardirosian (1984Ma13)	0,0188	0,0035	0,0025	0,0025	0,055	0,003					0,0112	0,0025	0,068	0,003
Iwata (1984Iw03)					0,052	0,004					0,0093	0,0026	^(o) 0,0589	0,0029
Johnson (1974Jo03)					0,058	0,004					0,007	0,002	0,067	0,004
Meyer (1990Me15)					0,05	0,01							0,07	0,01
Sharma (1979Sh08)					0,059	0,002					0,012	0,001	0,067	0,003
Chi2	4,0		10,5		0,8						2,9		0,9	
Chi2 crit:	4,6		3,3		2,2						2,3		2,3	
UWM:	0,01032		0,00350		0,05494						0,01017		0,06611	
WM:	0,01170		0,00314		0,05405						0,00999		0,06446	
Uc (int):	0,0024		0,0003		0,00046						0,00026		0,00051	
Uc (ext) :	0,0049		0,0009		0,00042						0,00044		0,00049	
LWM :	0,0117	0,0049	0,0031	0,0009	0,0541	0,0005					0,0100	^(e) 0,0008	0,0645	0,0005
I Abs.*	omitted		0,0030	0,0009	0,0529	0,0005	omitted		omitted		0,0098	0,0008	0,0631	0,0005

^(o) Outlier ^(e) expanded uncertainty so range to include the most precise Value ⁽ⁱ⁾ This original uncertainty was increased in order to limit the relative weight to 50%

2079 keV 2151 keV 2172 keV (19.1) (20.1)2090.9 keV (21.1)2099 keV (22.1)2108 keV 2145 keV (23, 1) Uc Uc Value Uc Uc Value Uc Value Value Uc Value Value Value Uc E907-2 0,003 5.28 0,20 0.046 0,003 0.052 0,003 0.0289 E907-3 0.024 5.56 0.08 0.058 0.048 0.001 0,001 0.001 0.0030 0.0003 E907-5 0.057 0.018 0,002 5,59 0,05 0.054 0.003 0,004 E907-6 ⁽ⁱ⁾ 0,0004 ⁽ⁱ⁾ 0.0001 E907-7 ⁽ⁱ⁾ 0.0003 0.0206 0.0006 5.63 0.02 0.0448 0.0430 DL=0.0002 0.0014 E907-8 0.0213 0.0008 5,34 0,0532 0,0457 0,0013 0.0010 0.0005 0.14 0,0016 0.0057 0,0002 ^(o) 0,0741 Patil (2006Pa16) 0.0019 5.40 0.07 0.0572 0.0013 0.0501 0.0009 0.00068 0.0000 Goswamy (1993Go10) 0.0268 0,0014 5,74 0.09 0.047 0.001 0.045 0,002 0.0021 0.0005 Jianming (1988Yo05) 0.0163 0.0025 5.69 0.046 0,002 0,044 0,002 0,11 Mardirosian (1984Ma13) 0.037 0.009 5.92 0.1 0.037 0.005 0.035 0.005 0.0046 0.0010 Iwata (1984Iw03) 0.0163 0.0025 5,59 0,03 0.045 0.006 0.0438 0.0027 ^(r) 0.081 Johnson (1974Jo03) 0.056 0.010 5.86 0,14 0.051 0.020 Meyer (1990Me15) 5.7 0.1 0.04 0.01 0.04 0.01 Sharma (1979Sh08) 0.0305 0.0010 5.75 0.12 0.04 0.01 0.047 0.002 Chi2 12,2 2,8 12,0 6.1 74.3 Chi2 crit: 2.4 2.2 2.2 2.2 3.3 UWM: 0.02393 5,6195 0.04762 0.04667 0.00337 WM: 0,02286 5.6176 0.04824 0.04540 0.00301 Uc (int): 0.00036 0.0150 0.00042 0.00037 0.00011 Uc (ext) : 0,0012 0,025 0.00146 0.00092 0.00094 ^(e) 0,0023 ^(e) 0.0034 ^(e) 0.0024 ^(e) 0.0016 LWM : 0.0229 0.0482 0.0454 0,0030 5.618 0.025 I Abs.* 0,0224 0,0022 5,493 0,024 0,0471 0,0033 0,0444 0,0023 omitted omitted 0,0029 0,0016

Table 3 (Cont'd) : Relative gamma ray intensities and absolute Values calculated with $^{(*)}Ig602 = 97,775$ (20) %. (i, j) refer to initial and final levels, (-1, n) transition not placed in the decay scheme. DL = Detection Limit

⁽ⁱ⁾ This original uncertainty was increased in order to limit the relative weight to 50%

^(o) Outlier

^(e) expanded uncertainty so range to include the most precise Value

Table 3 (Cont'd) : Relative gamma ray intensities and absolute Values calculated with $^{(*)}Ig602 = 97,775$ (20) %.

(i, j) refer to initial and final levels, (-1, n) transition not placed in the decay scheme. DL = Detection Limit

	(8, 0)	2182 keV	? (24, 1)	2204 keV	? (9, 0)	2224 keV		2232 keV		2253 keV		2256 keV		2274 keV
	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc
E907-2	0,043	0,003	0,030	0,002	0,021	0,013								
E907-3	0,041	0,001												
E907- 5	0,042	0,008												
E907- 6														
E907- 7	0,0422	0,0004	0,0004	0,0002	0,0002	0,0001			DL=0,00014		DL=0,00015			
E907- 8	0,0424	0,0011	0,0051	0,0002	0,0020	0,0003			0,0006	0,0001			0,0008	0,0003
Patil (2006Pa16)	^(o) 0,036	0,007	0,0310	0,0007			0,001	0,003			0,0006	0,0002		
Goswamy (1993Go10)	0,044	0,001												
Jianming (1988Yo05)	0,045	0,002												
Mardirosian (1984Ma13)	^(o) 0,048	0,002												
Iwata (1984Iw03)	0,0398	0,0019												
Johnson (1974Jo03)	0,041	0,003												
Meyer (1990Me15)	0,04	0,01												
Sharma (1979Sh08)	0,044	0,001												
Chi2	1,0		706,4		11,9									
Chi2 crit:	2,3		3,8		4,6									
UWM:	0,04217		0,01671		0,00773									
WM:	0,04241		0,00415		0,00109									
Uc (int):	0,00032		0,00015		0,00020									
Uc (ext) :	0,00031		0,00392		0,00068									
LWM :	0,04241	0,00032	0,017	0,016	0,0011	0,0009								
I Abs.*	0,04147	0,00031	Omitted		omitted		omitted		omitted		omitted		omitted	

^(o) Outlier

	(27, 1)	2283 keV	(10, 0)	2294 keV	(11, 0)	2323 keV		2373 keV		2386 keV	(13, 0)	2455 keV		2490 keV
	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc
E907- 2	^(o) 0,024	0,015	^(r) 0,082	0,007	^(o) 0,0098	0,0044					0,0093	0,0034		
E907- 3	0,0051	0,0004	0,029	⁽ⁱ⁾ 0,001	DL=0,005		DL=0,0049		DL=0,004					
E907- 5			0,032	0,002										
E907- 6														
E907-7	0,0046	0,0006	0,0342	0,0010	0,0020	⁽ⁱ⁾ 0,0001					0,0015	0,0002		
E907- 8	0,0064 (o)	0,0004	^(o) 0,413	0,011	0,0037	0,0003					0,0019	0,0003		
Patil (2006Pa16)	0,0422	0,0010	0,056	0,023	^(o) 0,0060	0,0003	0,0009	0,0003	0,00024	0,00002	^(r) 0,0092	0,0001	0,0020	0,0010
Goswamy (1993Go10)	0,0101	0,0008	^(r) 0,076	0,005	0,0027	0,0003					0,0018	0,0002		
Jianming (1988Yo05)	0,0076	0,0014	0,031	0,005	0,0025	0,0007					0,0016	0,0006		
Mardirosian (1984Ma13)	0,010	0,002	0,045	0,002	0,004	0,001					0,0010	0,0005		
lwata (1984lw03)	0,0041	0,0013	0,031	0,010										
Johnson (1974Jo03)	0,007	0,002	0,025	0,005										
Meyer (1990Me15)	0,008	0,001	0,031	0,001										
Sharma (1979Sh08)	0,0051	0,0006	0,059	0,002										
Chi2	5,8		43,2		5,7						0,8			
Chi2 crit:	2,4		2,4		3,3						3,3			
UWM:	0,00677		0,0374		0,00298						0,00156			
WM:	0,00596		0,03335		0,00260						0,00164			
Uc (int):	0,00020		0,00042		0,00014						0,00012			
Uc (ext) :	0,00048		0,0027		0,00034						0,00011			
LWM :	0,0060	0,0005	0,0334	^(e) 0,0042	0,0026	^(e) 0,0006					0,00164	0,00012		
I Abs.*	0.0059	0.0005	0.0327	0.0041	0,0025	0.0006	omitted		omitted		0.00160	0.00012	omitted	

Table 3 (Cont'd) : Relative gamma ray intensities and absolute Values calculated with $^{(*)}Ig602 = 97,775$ (20) %. (i, j) refer to initial and final levels, (-1, n) transition not placed in the decay scheme. DL = Detection Limit

^(o) Outlier

(e) expanded uncertainty so range to include the most precise Value
 (r) removed from analysis
 (i) This original uncertainty was increased in order to limit the relative weight to 50%

Table 3 (Cont'd) : Relative gamma ray intensities and absolute Values calculated with (*)Ig602 = 97,775 (20) %	•
(i, j) refer to initial and final levels, (-1, n) transition not placed in the decay scheme. DL = Detection Limit	

		2515 keV	(19, 0)	2682 keV	(20, 0)	2693 keV		2746 keV	(24, 0)	2807 keV		2814 keV		2871 keV
	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc	Value	Uc
E907-2			^(o) 0,007	0,003	0,0048	0,0021			^(o) 0,0069	0,003				
E907-3					0,0019	0,0001								
E907- 5					0,0025	0,0003								
E907-6														
E907-7			0,0017	0,0001	0,0033	0,0001			0,0007	0,0002			0,0002	0,0001
E907-8			0,0019	0,0001	^(o) 0,0433	0,0012			0,0016	0,0002				
Patil (2006Pa16)	0,00049	0,00001			0,0003	0,0001	0,0010	0,0001			0,0035	0,0002		
Goswamy (1993Go10)			0,0020	0,0004	0,0047	0,0005			0,0015	0,0002				
Jianming (1988Yo05)			0,0018	0,0006	0,0026	0,0016			0,0020	0,0008				
Mardirosian (1984Ma13)			^(o) 0,0025	0,0010	0,0056	0,0010								
Iwata (1984Iw03)					0,0027	0,0019								
Johnson (1974Jo03)					0,0024	0,0005								
Meyer (1990Me15)					0,0026	0,0003								
Sharma (1979Sh08)					0,0066	0,0005								
Chi2			0,7		48,2				4,5					
Chi2 crit:			3,8		2,2				3,8					
UWM:			0,00187		0,00334				0,00145					
WM:			0,00180		0,00186				0,00121					
Uc (int):			0,00006		0,00005				0,00011					
Uc (ext) :			0,00005		0,00038	(-)			0,00024	(-)				
LWM :			0,00180	0,00006	^(u) 0,0033	^(e) 0,0014			0,0012	^(e) 0,0005				
I Abs.*	omitted		0,00176	0,00006	0,0032	0,0014	omitted		0,0012	0,0005	omitted		omitted	

^(o) Outlier ^(e) expanded uncertainty so range to include the most precise Value ^(u) unweighted mean

		148 keV		158 keV	185 keV			189 keV		210 keV		254 keV		291 keV
	I (%)	Uc	l (%)	Uc	l (%)	Uc	I (%)	Uc	l (%)	Uc	I (%)	Uc	l (%)	Uc
E907- 2	0,012	0,005	0,005	0,001			0,002	0,001	0,0086	0,0012	0,0089	0,0014	0,0045	0,0008
E907- 3	DL=0,0031		DL=0,0032		DL=0,0041		DL=0,0042		DL=0,0043		0,013	0,002	DL=0,0053	
E907- 5														
E907-6														
E907- 7	0,0052	0,0011	0,0069	0,0014			0,0096	0,0058	0,0046	0,0023	0,0155	0,0015	0,0090	0,0009
E907- 8	0,0029	0,00084	0,0046	0,0007			0,0053	0,0005	0,0054	0,0010	0,0159	0,0014	0,0054	0,0011
Chi2	3,0		1,1				2,5		2,5		5,2		6,9	
Chi2 crit:	4,6		4,6				4,6		4,6		3,8		4,6	
UWM:	0,00669		0,00549				0,0056		0,00621		0,01340		0,00631	
WM:	0,00385		0,00506				0,0050		0,00649		0,01345		0,00618	
Uc (int):	0,00067		0,00055				0,0005		0,00073		0,00076		0,00053	
Uc (ext) :	0,00115		0,00056				0,0008		0,00114		0,00172		0,00140	
LWM :	0,0038	0,0012	0,0051	0,0006			0,005	0,0008	0,0065	0,0011	0,0135	^(e) 0,0025	0,0062	^(e) 0,0017

Table 4 : Absolute gamma ray intensity Value measured by the participants in the Euramet project 907; in %.

		335 keV		346 keV		370 keV		385 keV		400 keV		443 keV		468 keV
	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc
E907- 2	0,077	0,007	0,0033	0,0016	0,033	0,011	0,037	0,025	0,124	0,008	^(o) 0,186	0,004	0,052	0,009
E907-3	0,071	0,004	DL=0,0064		0,032	0,006	DL=0,0078		0,117	0,006	0,194	0,005	0,037	0,003
E907- 5	0,071	0,020			>0,0217	<0,0328			0,143	0,011	0,193	0,011	0,044	0,006
E907-6									0,16	0,06	0,192	0,069		
E907-7	0,0717	0,0015	0,0018	0,0018	0,0289	0,0026	DL=0,0023		0,13	0,01	0,1938	0,0023	0,0507	0,0027
E907-8	0,0710	0,0024	0,0034	0,0023	0,0334	0,0021			0,125	0,003	0,1899	0,0037	0,0467	0,0021
Chi2	0,0		0,3		0,6				1,0		0,2		3,0	
Chi2 crit:	3,8		4,6		3,8				3,0		3,3		3,3	
UWM:	0,07116		0,00280		0,03186				0,13252		0,19259		0,04617	
WM:	0,07144		0,00276		0,03167				0,12446		0,19289		0,04577	
Uc (int):	0,00123		0,00105		0,00155				0,00236		0,00183		0,00139	
Uc (ext) :	0,00020		0,00053		0,00122				0,00232		0,00086		0,00239	
LWM :	0,0714	0,0012	0,0028	0,001	0,0317	0,0016			0,1245	0,0024	0,1929	0,0018	0,0458	0,0024

		476 keV		481 keV		498 keV		525 keV		530 keV		553 keV		571 keV
	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc
E907-2	0,045	0,016	0,023	0,007	0,037	0,014	0,1393	0,0045	0,042	0,005	0,019	0,005	0,020	0,004
E907-3	DL=0,0069		0,014	0,006			0,1367	0,0044	0,022	0,003			0,012	0,004
E907- 5			>0,0187	<0,0288			^(o) 0,178	0,009						
E907-6			0,148	0,050			^(o) 0,050	0,042						
E907- 7	DL=0,0018		0,0248	0,0013	DL=0,0018		0,1372	0,0050	0,0275	⁽ⁱ⁾ 0,0012	0,0019	0,0008	0,0149	0,0017
E907- 8	0,0019	0,0008	0,0250	0,0012	0,0007	⁽ⁱ⁾ 0,0005	0,1449	0,0026	0,0433	0,0014				
Chi2	3,5		1,0		3,5		1,3		31,2				0,9	
Chi2 crit:	6,6		3,8		6,6		3,8		3,8				4,6	
UWM:	0,02344		0,02175		0,01884		0,13952		0,03348				0,01575	
WM:	0,02344		0,02465		0,01884		0,14138		0,03338				0,01508	
Uc (int):	0,01146		0,00088		0,00968		0,00187		0,00086				0,00143	
Uc (ext) :	0,02156		0,00087		0,01816		0,00211		0,00480				0,00136	
LWM :	0,023	0,022	0,0247	0,0009	0,019	0,018	0,1414	0,0021	0,033	^(e) 0,006			0,0151	0,0014

		602 keV		632 keV		645 keV		662 keV		669 keV		709 keV		713 keV
	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc
E907- 2	97,5	0,7	0,098	0,007	7,386	0,058	0,040	0,004			^(o) 1,325	0,019	2,205	0,029
E907-3	97,8	0,9	0,096	0,004	7,42	0,07	DL=0,0063				1,358	0,009	2,273	0,015
E907- 5	97,6	0,7	0,106	0,007	7,420	0,053	<0,0157				1,362	0,011	2,270	0,018
E907-6	^(o) 91	1			^(o) 7,00	0,11					1,350	0,065	^(o) 2,12	0,08
E907-7	97,84	0,34	0,1050	0,0012	7,417	0,028	0,0136	0,0009			1,3671	0,0056	2,28	0,01
E907-8	98,1	1,5	0,1052	0,0023	^(o) 7,33	0,11	0,0229	0,0011	0,1793	0,0029	1,350	0,021	2,21	0,04
Chi2	0,1		1,4		0,1		36,6				0,3		2,1	
Chi2 crit:	3,3		3,3		3,8		4,6				3,3		3,3	
UWM:	97,787		0,10209		7,4117		0,0254				1,35734		2,2475	
WM:	97,769		0,10440		7,4137		0,0190				1,36347		2,2694	
Uc (int):	0,260		0,00098		0,0214		0,0007				0,00426		0,0074	
Uc (ext) :	0,071		0,00115		0,0065		0,0045				0,00246		0,0107	
LWM :	97,77	0,26	0,1044	0,0011	7,414	0,021	0,019	0,005			1,3635	0,0043	2,269	0,011

		722 keV		735 keV		765 keV		775 keV		790 keV		816 keV		856 keV
	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc
E907-2	^(o) 10,538	0,084	0,134	0,005	0,0114	0,003	0,0101	0,002	0,737	0,007	^(o) 0,079	0,006	0,020	0,001
E907-3	10,713	0,072	0,129	0,005	DL=0,0072		DL=0,0073		0,737	0,010	0,072	0,007	0,017	0,005
E907- 5	10,680	0,075	0,122	0,012	>0,0167	<0,0258			0,737	0,009	0,074	0,013	>0,0177	<0,0278
E907-6	^(o) 9,96	0,16	^(o) 0,200	0,054					0,750	0,066				
E907-7	10,72	0,04	0,1309	⁽ⁱ⁾ 0,0016	0,0078	0,0012	0,0095	0,0005	0,742	0,004	0,0719	0,0036	0,0223	0,0007
E907- 8	10,71	0,17	0,1173	0,0022	0,0085	0,0013	0,0093	0,0013	0,727	0,012	0,0745	0,0017	0,0230	0,0015
Chi2	0,1		6,4		0,6		0,1		0,3		0,2		3,1	
Chi2 crit:	3,8		3,3		4,6		4,6		3,0		3,8		3,8	
UWM:	10,7067		0,1266		0,00925		0,00962		0,73830		0,07313		0,02049	
WM:	10,7122		0,1260		0,00836		0,00951		0,73906		0,07398		0,02108	
Uc (int):	0,0310		0,0013		0,00084		0,00044		0,00293		0,00150		0,00044	
Uc (ext) :	0,0087		0,0033		0,00063		0,00013		0,00173		0,00060		0,00078	
LWM :	10,712	0,031	0,126	^(e) 0,005	0,0084	0,0008	0,00951	0,00044	0,7391	0,0029	0,0740	0,0015	0,0211	0,0008

		899 keV		937 keV		968 keV		976 keV		997 keV		1014 keV		1045 keV
	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc
E907-2	0,022	0,009	0,020	0,005	1,86	0,05	0,082	0,005	0,024	0,007	0		1,837	0,033
E907-3	0,026	0,016	DL=0,0085		1,880	0,018	^(o) 0,093	0,011	DL=0,0091		DL=0,0093		1,826	0,019
E907- 5	<0,0187				1,863	0,017	0,086	0,013					1,816	0,020
E907-6					⁽ⁱ⁾ 2,60	0,11							1,82	0,10
E907-7	0,0171	0,0010	DL=0,0012		1,88	0,01	0,0843	0,0010	0,0014	0,0014	0,0025	0,0025	1,839	0,008
E907-8	0,0171	0,0014	0,0030	⁽ⁱ⁾ 0,0012	1,87	0,03	0,0824	0,0019	0,0037	⁽ⁱ⁾ 0,0007	0,0068	0,0021	1,836	0,029
Chi2	0,2		5,3		0,4		0,4		5,4		1,8		0,3	
Chi2 crit:	3,8		6,6		3,3		3,8		4,6		6,6		3,0	
UWM:	0,02041		0,0116		1,8711		0,08360		0,00971		0,00463		1,8292	
WM:	0,01717		0,0116		1,8797		0,08379		0,00300		0,00497		1,8350	
Uc (int):	0,00079		0,0037		0,0064		0,00090		0,00097		0,00161		0,0063	
Uc (ext) :	0,00035		0,0085		0,0040		0,00054		0,00224		0,00215		0,0034	
LWM :	0,0172	0,0008	0,012	0,009	1,880	0,006	0,0838	0,0009	0,0030	0,0022	0,0050	0,0021	1,835	0,006

										1180				
-		1053 keV		1086 keV		1097 keV		1163 keV		keV		1198 keV		1205 keV
	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc
E907- 2			0,040	0,004	0,0335	0,008								
E907-3	DL=0,0097		0,041	0,009			DL=0,0108				DL=0,0112		DL=0,0115	
E907- 5			^(o) 0,049	0,008										
E907-6														
E907-7	0,0026	0,0026	0,0361	0,0012	DL=0,0019		DL=0,0019				DL=0,002		DL=0,016	
E907- 8	0,0038	0,0013	0,0369	0,0017	0,0028	⁽ⁱ⁾ 0,0013	0,0016	0,0010	0,606	0,010	0,0030	0,0009	0,0142	0,0005
Chi2	0,2		0,4		7,3									
Chi2 crit:	6,6		3,8		6,6									
UWM:	0,00319		0,03850		0,01815									
WM:	0,00357		0,03660		0,01815									
Uc (int):	0,00116		0,00095		0,00569									
Uc (ext) :	0,00051		0,00057		0,01535									
LWM :	0,0036	0,0012	0,0366	0,0009	0,018	0,015								

		1235 keV		1253 keV		1263 keV		1269 keV		1301 keV		1325 keV		1355 keV
	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc
E907-2	0,027	0,006	0,041	0,009	0,042	0,004			0,031	0,004	1,597	0,030	1,039	0,029
E907-3			DL=0,0117		0,029	0,010	DL=0,0118		0,046	0,010	1,565	0,024	1,036	0,020
E907- 5					0,030	0,008			0,036	0,009	1,564	0,018	1,029	0,022
E907-6											^(o) 1,440	0,124	^(o) 0,92	0,10
E907-7	0,0046	⁽ⁱ⁾ 0,0009	DL=0,0018		0,0404	0,0014	DL=0,0019		0,0332	0,0021	1,586	0,006	1,0394	0,0043
E907-8	0,0116	0,0015	0,0005			0,0016	0,0028	0,0010	0,0376	0,0030	^(o) 1,76	0,03	^(o) 1,06	0,02
Chi2	10,8		11,0						0,9		0,7		0,1	
Chi2 crit:	4,6		6,6						3,3		3,8		3,8	
UWM:	0,01456		0,0208						0,03669		1,57771		1,03606	
WM:	0,00867		0,0208						0,03442		1,58251		1,03894	
Uc (int):	0,00105		0,0061						0,00154		0,00581		0,00405	
Uc (ext) :	0,00346		0,0202						0,00144		0,00480		0,00112	
LWM :	0,0087	0,004	0,021			0,0014			0,0344	0,0015	1,583	0,006	1,0389	0,0040

		1368 keV		1376 keV		1385 keV		1428 keV		1436 keV		1445 keV		1453 keV
	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc
E907- 2	2,585	0,041	0,508	0,020	0,070	0,006	0,048	0,007	1,222	0,024	0,328	0,007	0,031	0,007
E907-3	2,571	0,025	0,471	0,015	0,050	0,011			1,238	0,019	0,303	0,013	DL=0,0163	
E907- 5	2,621	0,023	0,493	0,009	0,062	0,008			1,210	0,031	0,342	0,012		
E907- 6	^(o) 2,41	0,11	0,47	0,09	^(o) 0,18	0,07			^(o) 1,08	0,09	0,35	0,08		
E907- 7	2,624	0,011	0,5019	⁽ⁱ⁾ 0,0033	0,0682	0,0018	DL=0,0025		1,230	0,005	0,3286	⁽ⁱ⁾ 0,0023	DL=0,0026	
E907- 8	2,63	0,05	0,465	0,008	0,059	0,002	0,0262	0,0016	^(o) 1,31	0,02	0,367	0,006	0,077	⁽ⁱ⁾ 0,002
Chi2	1,1		3,4		3,4		5,3		0,2		6,8		20,6	
Chi2 crit:	3,3		3,0		3,3		6,6		3,8		3,0		6,6	
UWM:	2,6058		0,4849		0,06202		0,0371		1,2249		0,3364		0,0539	
WM:	2,6154		0,4904		0,06416		0,0371		1,2296		0,3363		0,0539	
Uc (int):	0,0088		0,0038		0,00125		0,0048		0,0048		0,0030		0,0050	
Uc (ext) :	0,0093		0,0070		0,00232		0,0109		0,0024		0,0078		0,0229	
LWM :	2,615	0,009	0,490	^(e) 0,012	0,0642	^(e) 0,0041	0,037	0,011	1,2296	0,0048	0,336	0,008	0,054	0,023

		1488 keV		1505 keV		1526 keV		1557 keV		1565 keV		1579 keV		1622 keV
	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc
E907-2	0,669	0,017	0,051	0,016	0,410	0,018					0,430	0,018	0,040	0,003
E907-3	0,669	0,014			0,395	0,012			DL=0,0105		0,413	0,014	0,042	0,008
E907- 5	0,676	0,015			^(o) 0,432	0,016			>0,0187	<0,0298	0,404	0,008		
E907-6	0,65	0,09			0,39	0,07							^(o) 0,200	0,046
E907-7	0,685	0,007	DL=0,002		0,4094	0,0025	DL=0,0017		0,0114	⁽ⁱ⁾ 0,0007	^(o) 0,1420	0,0012	0,0397	0,0008
E907-8	0,666	0,012	0,0084	0,0028	0,398	0,007	0,013	0,007	0,0053	0,0009	^(o) 0,353	0,007	0,0397	0,0012
Chi2	0,5		3,6		0,9				20,9		0,9		0,0	
Chi2 crit:	3,0		6,6		3,3				6,6		4,6		3,8	
UWM:	0,6692		0,02970		0,4005				0,00838		0,4155		0,04028	
WM:	0,6770		0,02970		0,4077				0,00838		0,4091		0,03971	
Uc (int):	0,0051		0,01118		0,0023				0,00066		0,0066		0,00066	
Uc (ext) :	0,0037		0,02130		0,0022				0,00304		0,0064		0,00012	
LWM :	0,677	0,005	0,030	0,021	0,4077	0,0023			0,0084	0,0030	0,409	0,007	0,0397	0,0007

		1657 keV		1690 keV		1720 keV		1757 keV		1851 keV		1918 keV		1950 keV
	I (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc
E907-2			45,56	1,09	0,095	0,005					0,055	0,004		
E907-3	DL=0,0089		47,04	0,40	^(o) 0,088	0,004			DL=0,0077		0,049	0,003		
E907- 5			47,10	0,35	0,098	0,006					0,052	0,008		
E907-6			^(o) 44,70	0,77	^(o) 0,123	0,041	0,007	0,019	^(r) 0,31	0,06	^(o) 0,070	0,035		
E907-7	DL=0,0012		47,65	0,18	0,0946	0,0007	DL=0,0009		0,0053	0,0006	0,0526	0,0005	DL=0,0006	
E907-8	0,009	0,003	46,03	0,87	0,0955	0,0020			0,0008	0,0001	0,0527	0,0017	0,0528	0,0110
Chi2			2,2		0,2				28,9		0,3			
Chi2 crit:			3,3		3,8				6,6		3,3			
UWM:			46,68		0,09581				0,00304		0,05235			
WM:			47,39		0,09475				0,00304		0,05254			
Uc (int):			0,15		0,00065				0,00042		0,00049			
Uc (ext) :			0,22		0,00025				0,00225		0,00028			
LWM :			47,39	0,22	0,0947	0,0006			0,0030	0,0023	0,0525	0,0005		

		1970 keV		2015 keV		2039 keV		2078 keV		2090 keV		2099 keV		2108 keV
	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc
E907-2			0,013	0,002	0,062	0,004	0,028	0,003	5,15	0,19	0,045	0,003	0,051	0,003
E907-3	DL=0,0016		0,008	0,001	0,063	0,003	0,023	0,001	5,44	0,08	0,056	0,001	0,047	0,001
E907-5					0,062	0,006	0,017	0,002	5,45	0,06	0,052	0,003	0,056	0,004
E907-6														
E907-7			0,0090	0,0003	0,0622	0,0006	0,0201	0,0006	5,511	0,022	0,0439	⁽ⁱ⁾ 0,0004	0,0421	⁽ⁱ⁾ 0,0003
E907-8			0,0092	0,0010	^(o) 0,0751	0,0016	0,0212	0,0007	5,33	0,11	0,0525	0,0013	0,0456	0,0011
Chi2			1,8		0,0		4,5		1,8		17,8		6,2	
Chi2 crit:			3,8		3,8		3,3		3,3		3,3		3,3	
UWM:			0,00968		0,06220		0,02198		5,3766		0,04998		0,0482	
WM:			0,00907		0,06221		0,02120		5,4909		0,04849		0,0444	
Uc (int):			0,00026		0,00058		0,00039		0,0193		0,00062		0,0006	
Uc (ext) :			0,00034		0,00010		0,00082		0,0256		0,00260		0,0014	
LWM :			0,00907	0,00034	0,0622	0,0006	0,0212	^(e) 0,0011	5,491	0,026	0,0485	^(e) 0,0046	0,048	^(e) 0,006

		2151 keV		2172 keV		2182 keV		2203 keV		2224 keV		2253 keV		2274 keV
	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc
E907- 2					0,042	0,003	0,030	0,002	0,020	0,013				
E907-3			0,0029	0,0003	0,040	0,001								
E907- 5					0,040	0,008								
E907- 6														
E907- 7	DL=0,0002		0,0014	0,0001	0,0413	0,0004	0,0004	⁽ⁱ⁾ 0,0002	0,0002	⁽ⁱ⁾ 0,0001	DL=0,0001		DL=0,0002	
E907- 8	0,0016	0,0008	0,0057	0,0002	0,0435	0,0010	0,0063	0,0003	0,0020	0,0003	0,0005	0,0001	0,0008	0,0003
Chi2			172,2		1,6		241,5		12,0					
Chi2 crit:			4,6		3,3		4,6		4,6					
UWM:			0,00335		0,04131		0,01210		0,00738					
WM:			0,00317		0,04145		0,00359		0,00107					
Uc (int):			0,00011		0,00035		0,00018		0,00019					
Uc (ext) :			0,00140		0,00045		0,00280		0,00067					
LWM :			0,0032	^(e) 0,0018	0,04145	0,00045	0,0036	^(e) 0,0032	0,0011	^(e) 0,0009				

		2283 keV		2293 keV		2323 keV		2454 keV		2682 keV		2693 keV		2808 keV
	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc	l (%)	Uc
E907-2	0,023	0,014	0,080	0,007	0,0096	0,0043	0,0091	0,0034	0,0071	0,0033	0,0047	0,0020	0,0067	0,0027
E907-3	0,0049	0,0004	0,028	⁽ⁱ⁾ 0,001	DL=0,005		DL=0,0049		DL=0,004		0,0019	⁽ⁱ⁾ 0,0001	DL=0,0033	
E907- 5			0,032	0,002							0,0024	0,0003		
E907-6														
E907-7	0,0045	0,0006	0,0335	0,0010	0,0020	⁽ⁱ⁾ 0,0001	0,0015	0,0002	0,0017	0,0001	0,0032	0,0001	0,0007	0,0002
E907-8	0,0062	0,0003	^(o) 0,414	0,009	0,0042	0,0003	0,0018	0,0003	0,0019	0,0001	^(o) 0,0434	0,0010	0,0009	0,0001
Chi2	3,5		20,3		12,0		2,9		2,8		22,6		2,9	
Chi2 crit:	3,8		3,8		4,6		4,6		4,6		3,8		4,6	
UWM:	0,00966		0,04337		0,00526		0,00413		0,00358		0,00305		0,00277	
WM:	0,00545		0,03123		0,00311		0,00159		0,00177		0,00251		0,00084	
Uc (int):	0,00023		0,00065		0,00024		0,00016		0,00006		0,00008		0,00010	
Uc (ext) :	0,00043		0,00295		0,00082		0,00028		0,00010		0,00038		0,00018	
LWM :	0,00545	0,00043	0,0312	0,0029	0,0031	^(e) 0,0011	0,00159	0,00028	0,00177	0,00010	0,0025	0,0006	0,00084	0,00018

¹²⁴ Sb	

		2871 keV
	l (%)	Uc
E907- 2		
E907-3		
E907- 5		
E907-6		
E907-7	0,0002	0,0001
E907- 8		
Chi2		
Chi2 crit:		
UWM:		
WM:		
Uc (int):		
Uc (ext) :		
LWM :		

(¹⁾ This original uncertainty was increased in order to limit the relative weight to 50 % (^{o)} Outlier (^{e)} expanded uncertainty so range to include the most precise I (%) (^{r)} removed from analysis



1 Decay Scheme

L'antimoine 124 se désintègre par émission bêta moins vers des niveaux excités du tellure 124. Sb-124 disintegrates by beta minus emissions to excited levels in Te-124.

2 Nuclear Data

 $\begin{array}{rll} T_{1/2}(^{124}{\rm Sb}\) &:& 60,\!208 & (11) & {\rm d} \\ Q^-(^{124}{\rm Sb}\) &:& 2904,\!3 & (15) & {\rm keV} \end{array}$

2.1 β^- Transitions

	Energy keV	$\begin{array}{c} {\rm Probability} \\ \times \ 100 \end{array}$	Nature	$\lg ft$
$\beta_{0,27}^{-}$	$17,9\ (15)$	0,0059~(5)	Allowed	6,9
$\beta_{0,26}^{-}$	$38,\! 6\ (15)$	0,054~(9)	Allowed	6,9
$\beta_{0.25}^{-}$	89,7~(15)	0,0207~(12)		8,4
$\beta_{0.24}^{-}$	96,8~(15)	0,0012~(5)	1st Forbidden	$_{9,8}$
$\beta_{0.23}^{-}$	129,2(15)	$0,\!653~(6)$		7,5
$\beta_{0.22}^{-,-3}$	193,3(15)	0,106(6)	1st Forbidden	8,8
$\beta_{0,21}^{-,}$	202,7(15)	0,571 (25)	Allowed	8
$\beta_{0,20}^{-,-1}$	210,6(15)	8,663(27)	Allowed	7
$\beta_{0.19}^{-19}$	221,8(15)	0,0242 (22)	1st Forbidden	$9,\!6$
$\beta_{0.18}^{-10}$	285,2(15)	0,0098 (8)		10,4
$\beta_{0.17}^{-10}$	354,6(15)	0,0364(22)		10
$\beta_{0.16}^{-16}$	382,8(15)	0,0529(5)	1st Forbidden	10
$\beta_{0.15}^{-15}$	392,3(15)	0,0422 (19)	1st Forbidden	10,2
$\beta_{0.14}^{-14}$	421,0 (15)	0,332(10)	1st Forbidden	9,4
$\beta_{0.13}^{-13}$	449,3 (15)	0,0050(26)	1st Forbidden	11,3
$\beta_{0.11}^{-10}$	580,9(15)	0,0686(14)	1st Forbidden	10,5
$\beta_{0,10}^{-10}$	610, 6 (15)	51,21 (19)	Allowed	7,7
$\beta_{0,9}^{-}$	679,5(15)	0,0967(34)	1st Forbidden	10,6

	Energy keV	$\begin{array}{c} {\rm Probability} \\ \times \ 100 \end{array}$	Nature	$\lg ft$
$\begin{array}{c} \beta_{0,8}^{-} \\ \beta_{0,7}^{-} \\ \beta_{0,6}^{-} \\ \beta_{0,5}^{-} \\ \beta_{0,4}^{-} \\ \beta_{0,3}^{-} \\ \beta_{0,2}^{-} \\ \beta_{0,1}^{-} \end{array}$	$\begin{array}{c} 721,9 \ (15) \\ 812,6 \ (15) \\ 865,0 \ (15) \\ 946,4 \ (15) \\ 1247,7 \ (15) \\ 1578,8 \ (15) \\ 1655,7 \ (15) \\ 2301,6 \ (15) \end{array}$	$\begin{array}{c} 0,47 \ (30) \\ 0,688 \ (38) \\ 4,143 \ (18) \\ 2,295 \ (7) \\ 0,0053 \ (10) \\ 4,815 \ (29) \\ 2,472 \ (33) \\ 23,44 \ (28) \end{array}$	1st Forbidden 1st Forbidden 1st Forbidden 3rd Forbidden 1st Forbidden 1st Forbidden 1st Forbidden	$10 \\ 10 \\ 9,4 \\ 9,8 \\ 12,8 \\ 10,3 \\ 10,7 \\ 10,3$

2.2 Gamma Transitions and Internal Conversion Coefficients

	Energy keV	$P_{\gamma+ce} \times 100$	Multipolarity	α_K	$lpha_L$	$lpha_M$	α_T
$\gamma_{14,12}({\rm Te})$	148,02(5)	0,0037~(6)	E1+M2				
$\gamma_{(-1,1)}(\mathrm{Te})$	159,867 (35)	0,0049~(6)					
$\gamma_{14,10}(\mathrm{Te})$	189,565 (18)	0,0043~(5)					
$\gamma_{20,14}(\text{Te})$	210,402 (19)	0,0053~(7)					
$\gamma_{10,6}(\text{Te})$	254,424 (6)	0,0144~(9)	(E1)	$0,01269\ (18)$	0,001575 (22)	0,000312 (5)	0,01465~(21)
$\gamma_{23,14}(\text{Te})$	291,793 (25)	0,0069~(7)					
$\gamma_{10,5}(\text{Te})$	335,797 (16)	0,073~(1)	E1	0,00612 (9)	0,000754 (11)	0,0001495~(21)	$0,00706\ (10)$
$\gamma_{20,11}(\text{Te})$	370,269 (30)	$0,0286\ (11)$					
$\gamma_{20,10}({ m Te})$	399,967~(6)	0,1284 (31)	E2	0,01323 (19)	0,00196 (3)	0,000394~(6)	0,01566~(22)
$\gamma_{14,6}(\text{Te})$	443,989 (18)	0,197~(16)	M1 + 26% E2	0,01092 (16)	0,001360 (19)	0,000271 (4)	0,01261 (18)
$\gamma_{20,9}(\text{Te})$	468,840 (25)	0,0460 (26)	E1	0,00268 (4)	0,000327 (5)	0,0000648 (9)	0,00309 (5)
$\gamma_{23,10}(\text{Te})$	481,36(2)	0,0232 (31)					
$\gamma_{14,5}(\text{Te})$	525,362(24)	0,1462 (35)	M1 + 50% E2	0,0066 (3)	0,000867 (18)	0,000173 (4)	0,0077 (3)
$\gamma_{26,12}(Te)$	530,46(7)	0,036(9)					
$\gamma_{26,10}(\text{Te})$	572,01 (5)	0,0176 (8)		(-)	(-)		()
$\gamma_{1,0}(\mathrm{Te})$	602,7278 (21)	98,254(21)	${ m E2}$	0,00420 (6)	0,000566 (8)	0,0001132(16)	0,00490(7)
$\gamma_{5,3}(\text{Te})$	632,403 (16)	0,1029(21)				()	(-)
$\gamma_{2,1}$ (Te)	645,8542 (37)	7,452(15)	E2+0,004%M3	0,00351 (5)	0,000468 (7)	0,0000935(14)	0,00409(6)
$\gamma_{21,6}$ (Te)	662,334(10)	0,024(11)					
$\gamma_{5,2}(\text{Te})$	709,333 (16)	1,368(5)	M1+3%E2	0,00349(5)	0,000429(7)	0,0000853(13)	0,00402(6)
$\gamma_{6,3}$ (Te)	713,776(5)	2,281(7)	M1+50%E2	0,0031(4)	0,00039(4)	0,000078(7)	0,0036(4)
$\gamma_{3,1}$ (Te)	722,7842 (37)	10,742 (22)	M1 + 92% E2	0,00271 (4)	0,000352(5)	0,0000702(10)	0,00314(5)
$\gamma_{23,6}$ (Te)	735,782(17)	0,1312(16)		0.010 (0)			0.001 (=)
$\gamma_{7,3}$ (Te)	766,168(21)	0,0105(9)	E0,M1	0,019(6)			0,021(7)
$\gamma_{25,6}(\text{Te})$	775,27(7)	0,0098(4)	120	0.00014 (0)			0.000.00 (0)
$\gamma_{6,2}(\text{Te})$	790,706 (5)	0,7433(24)	E2	0,00214 (6)	0,000276 (8)	0,000055(2)	0,00248 (8)
$\gamma_{23,5}(Te)$	817,155 (23)	0,0744 (12)					
$\gamma_{8,3}(\text{Te})$	856,878 (30)	0,0227 (5)					
$\gamma_{9,3}(Te)$	899,327 (25)	0,0179(7)	D1 + 407 MO	0.000500 (0)	0.0000070 (11)	0.00001040 (00)	0.000050 (11)
$\gamma_{10,3}(Te)$	968,200(5)	1,888(10)	E1+4%M2	0,000569 (9)	0,0000678 (11)	0,00001343(22)	0,000653(11)
$\gamma_{9,2}(\text{Te})$	976,257 (25)	0,0832(7)					
$\gamma_{(-1,2)}(1e)$	997,80 (3)	0,0033(23)		0.000404 (0)	0.0000505 (11)	0.00001160.(01)	0.000565 (10)
$\gamma_{10,2}(1e)$	1045,130(5)	1,853(14)	E1+0,09%M2	0,000494(9)	0,0000587(11)	0,00001163(21)	0,000567(10)
$\gamma_{4,1}(Te)$	1053,87(30)	0,0053(10)	E2	0,001117(16)	0,0001394(20)	0,0000277(4)	0,001290(18)
$\gamma_{12,2}(1e)$ (T-)	1080,08(5)	0,0367(9)	EI	0,000457(7)	0,0000543(8)	0,00001074(15)	0,000524(8)
$\gamma_{(-1,3)}(1e)$	1235(1) 1962 46 (7)	0,0073(20)					
$\gamma_{15,2}(1e)$	1203,40 (7)	0,0422 (19)					
$\gamma_{17,2}(1e)$	1301,15(9)	0,0364 (22)	EO	0 000009 (10)	0 0000040 (10)	0 00001685 (04)	0.000997 (19)
$\gamma_{3,0}(1e)$	1323,312(3)	1,388(1)	E2 E9+0.907M9	0,000093(10)	0,0000848(12)	0,00001085(24)	0,000827 (12)
$\gamma_{5,1}(1e)$	1999,187 (10)	1,0423 (38)	E2+9,3%M3	0,0009 (5)	0,00011 (6)	0,000023 (11)	0,0011(5)

LNE – LNHB/CEA Table de Radionucléides

124	Ch	
51	SD	73

	Energy keV	$P_{\gamma+ce} \times 100$	Multipolarity	α_K	α_L	α_M	α_T
$\gamma_{20,3}(\text{Te})$	1368,167(6)	2,621 (8)	E1+0,04%M2	0,000303(5)	0,0000358 (6)	0,00000709(10)	0,000478 (7)
$\gamma_{21,3}(\text{Te})$	1376,110 (9)	0,5001 (43)	E1+0,01%M2	0,000300 (5)	0,0000354~(6)	0,00000701 (12)	0,000479 (7)
$\gamma_{22,3}(\text{Te})$	1385,500 (21)	0,062~(6)					
$\gamma_{6,1}(\mathrm{Te})$	1436,5602 (45)	1,235~(8)	M1 + 69% E2	0,00063~(5)	0,000076~(6)	$0,0000151\ (11)$	0,00078~(5)
$\gamma_{20,2}(\text{Te})$	1445,097~(6)	0,334~(7)	E1+M2	0,00029 (4)	0,000034 (4)	0,0000067 (8)	0,00052~(4)
$\gamma_{7,1}(\mathrm{Te})$	1488,952 (21)	0,6776 (37)	M1+1%E2	0,000659 (14)	0,0000792~(16)	0,0000157 (3)	0,000829 (16)
$\gamma_{23,2}(\text{Te})$	1526,488 (17)	0,414~(5)	${ m E1}$	0,000252 (4)	0,0000296 (5)	0,00000586 (9)	0,000535 (8)
$\gamma_{25,2}(\text{Te})$	1565,98 (7)	0,0109(12)					
$\gamma_{8,1}(\mathrm{Te})$	1579,662 (30)	0,412~(5)	M1+E2	0,00054 (5)	0,000065~(6)	0,0000128(11)	0,00072~(5)
$\gamma_{9,1}({ m Te})$	1622,111 (25)	0,0416~(19)	E2	0,000467~(7)	0,0000564 (8)	0,00001118(16)	0,000664 (10)
$\gamma_{4,0}({ m Te})$	1656, 6 (3)		E0				
$\gamma_{10,1}(\text{Te})$	1690,9842 (45)	47,49(19)	E1+0,01%M2	0,000213 (4)	0,0000250 (4)	0,00000494 (8)	0,000615 (9)
$\gamma_{11,1}(\text{Te})$	1720,682 (30)	0,0947~(6)	M1+E2	0,00045~(4)	0,000054 (4)	0,0000107 (8)	0,00068 (4)
$\gamma_{13,1}(\text{Te})$	1852,23 (7)	0,0030 (9)	M1+E2	0,00039 (3)	0,000047~(4)	0,0000093 (7)	0,00067~(3)
$\gamma_{16,1}(\text{Te})$	1918,75~(6)	0,0529 (5)	M1(+E2)	0,000364 (24)	0,000043 (3)	0,0000086~(6)	0,00067~(3)
$\gamma_{18,1}(\text{Te})$	2016, 36(6)	0,0098 (8)					
$\gamma_{6,0}(\mathrm{Te})$	2039,288 (4)	0,0631 (5)	E2	0,000305 (5)	0,0000364 (5)	0,00000721(10)	0,000667~(10)
$\gamma_{19,1}(\text{Te})$	2079,77 (13)	0,0224~(22)	M1+E2	0,000311 (18)	0,0000371 (21)	0,0000073 (4)	0,000691 (20)
$\gamma_{20,1}(\text{Te})$	2090,951 (5)	5,498(24)	E1+0,1%M2	0,0001522 (23)	0,0000178 (3)	0,00000352~(6)	0,000838(12)
$\gamma_{21,1}(\text{Te})$	2098,894 (9)	0,0471 (33)					
$\gamma_{22,1}(\text{Te})$	2108,284 (21)	0,0444~(23)					
$\gamma_{23,1}(\text{Te})$	2172,342 (17)	0,0029(16)					
$\gamma_{8,0}({ m Te})$	2182,39 (3)	0,04147(31)					
$\gamma_{27,1}({\rm Te})$	2283,64 (6)	0,0059 (5)	E1+M2	0,00033~(21)	0,000040 (25)	0,000008 (5)	0,00091 (5)
$\gamma_{10,0}(\text{Te})$	2293,712 (4)	0,0327~(41)					
$\gamma_{11,0}(\text{Te})$	2323,41 (3)	0,0025~(6)					
$\gamma_{13,0}(\text{Te})$	2454,96 (7)	0,00160(12)	E2	0,000219 (3)	0,0000259 (4)	0,00000513 (8)	0,000768(11)
$\gamma_{19,0}(\text{Te})$	2682,50 (15)	0,00176 (6)					
$\gamma_{20,0}(\text{Te})$	2693,679 (10)	0,0032 (14)					
$\gamma_{24,0}(\text{Te})$	2807,55 (24)	0,0012 (5)	E2	0,0001730 (25)	0,0000204 (3)	0,00000404 (6)	0,000878 (13)

3 Atomic Data

3.1 Te

ω_K	:	$0,\!875$	(4)
$\bar{\omega}_L$:	$0,\!0862$	(35)
n_{KL}	:	0,917	(4)

3.1.1 X Radiations

	${ m Energy}\ { m keV}$	Relative probability
$\begin{array}{c} \mathbf{X}_{\mathbf{K}} \\ \mathbf{K}\alpha_{2} \\ \mathbf{K}\alpha_{1} \\ \mathbf{K}\beta_{3} \\ \mathbf{K}\beta_{1} \\ \mathbf{K}\beta_{5}^{\prime\prime} \end{array}$	27,202 27,4726 30,9446 30,996 31,236	53,7 100 } } 28,6

		$\begin{array}{c} {\rm Energy} \\ {\rm keV} \end{array}$		Relative probability
	$K\beta_2$	31,7008	}	
	$\mathrm{K}eta_4$	31,774	}	6,2
	$\mathrm{KO}_{2,3}$	$31,\!812$	}	
$\mathbf{X}_{\mathbf{L}}$				
	$\mathrm{L}\ell$	$3,\!3348$		
	$L\alpha$	$3,\!7595 - 3,\!7697$		
	$\mathrm{L}\eta$	$3,\!6052$		
	$\mathrm{L}eta$	$4,\!0299 - 4,\!3661$		
	$\mathrm{L}\gamma$	$4,\!4448 - 4,\!8228$		

3.1.2 Auger Electrons

	${ m Energy}\ { m keV}$	Relative probability
Auger K KLL KLX KXY	21,804 - 22,989 25,814 - 27,470 29,80 - 31,81	$100 \\ 45,3 \\ 5,13$

4 Electron Emissions

		Energy keV	Electrons per 100 disint.
e_{AL}	(Te)	2,3 - 4,9	0,4829 (26)
e _{AK}	(Te) KLL KLX KXY	$\begin{array}{rrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrr$	0,0628 (22) } } }
$\begin{array}{c} ec_{1,0 \ \rm K} \\ ec_{1,0 \ \rm L} \\ ec_{1,0 \ \rm M} \\ ec_{2,1 \ \rm K} \\ ec_{3,1 \ \rm K} \\ ec_{10,1 \ \rm K} \end{array}$	(Te) (Te) (Te) (Te) (Te) (Te)	$\begin{array}{rrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrr$	$\begin{array}{c} 0,411 \ (6) \\ 0,0553 \ (8) \\ 0,01107 \ (16) \\ 0,02605 \ (37) \\ 0,02902 \ (43) \\ 0,01011 \ (19) \end{array}$

	${ m Energy}\ { m keV}$		rgy V	Electrons per 100 disint.
$\beta_{0,27}^-$	max:	17,9	(15)	0,0059 (5)
$\beta_{0,27}^{-}$	avg:	4,5	(4)	
$\beta_{0,26}^{-}$	max:	$38,\! 6$	(15)	0,054 (9)
$\beta_{0,26}^{-}$	avg:	$_{9,8}$	(4)	
$\beta_{0,25}^-$	max:	89,7	(15)	0,0207~(12)
$\beta_{0,25}^{-}$	avg:	$23,\!4$	(4)	
$\beta_{0,24}^{-}$	max:	$96,\!8$	(15)	0,0012~(5)
$\beta_{0,24}^{-}$	avg:	$25,\!3$	(4)	
$\beta_{0,23}^{-}$	max:	129,2	(15)	$0,\!653~(6)$
$\beta_{0,23}^{-}$	avg:	$34,\!4$	(4)	
$\beta_{0,22}^{-}$	max:	$193,\!3$	(15)	$0,\!106~(6)$
$\beta_{0,22}^{-}$	avg:	$52,\!9$	(5)	
$\beta_{0,21}^{-}$	max:	202,7	(15)	$0,571\ (25)$
$\beta_{0,21}^{-}$	avg:	55,7	(5)	
$\beta_{0.20}^{-}$	max:	$210,\!6$	(15)	8,663~(27)
$\beta_{0,20}^{-,-0}$	avg:	58,0	(5)	
$\beta_{0.19}^{-}$	max:	221,8	(15)	0,0242 (22)
$\beta_{0,19}^{-,10}$	avg:	$61,\!5$	(5)	
$\beta_{0.18}^{-}$	max:	285,2	(15)	0,0098 (8)
$\beta_{0.18}^{-}$	avg:	81,0	(5)	
$\beta_{0.17}^{-17}$	max:	$354,\! 6$	(15)	0,0364 (22)
$\beta_{0.17}^{-}$	avg:	$103,\!6$	(5)	
$\beta_{0.16}^{-}$	max:	382,8	(15)	0,0529(5)
$\beta_{0.16}^{-16}$	avg:	113,0	(5)	, , ,
$\beta_{0,15}^{-}$	max:	392.3	(15)	0.0422(19)
$\beta_{0,15}^{-}$	avg:	116,0	(5)	, ()
$\beta_{0,14}^{-14}$	max:	421.0	(15)	0.332(10)
$\beta_{0,14}^{-14}$	avg:	126,0	(5)	-) (-)
$\beta_{0,14}^{-10}$	max:	449.3	(15)	0.0050(26)
$\beta_{0,12}^{-12}$	avg:	135.8	(6)	2,0000 (20)
$\beta_{0,13}^{-}$	max:	580.9	(15)	0.0686(14)
$\beta_{0,11}^{-1}$	avg:	182.8	(6)	0,0000 (11)
$\beta_{0,10}^{-10}$	max	610.6	(15)	51 21 (19)
$\beta_{0,10}^{-10}$	avg:	193.8	(6)	51,21 (10)
β_{-c}^{-}	max	679.5	(15)	0 0967 (34)
$\beta_{0,9}^{-}$	avg:	219.5	(6)	0,0001 (04)
β_{-}^{-}	mav	721 0	(15)	0 47 (30)
$\beta_{0,8}^{-}$	avg.	236.0	(6)	0,41 (00)
/~0,8 ∂	mav.	-30,0 812.6	(15)	0 688 (38)
$\beta_{0,7}^{-}$	avo.	271.0	(6)	0,000 (00)
$\sim 0,7$ β^{-}	max.	865.0	(15)	1 112 (10)
$\beta_{0,6}$ β_{-}^{-}	max.	202,0 202	(10) (1)	4,143 (18)
$\rho_{0,6}$	avg.	494	(1)	

		${ m Energy}\ { m keV}$		Electrons per 100 disint.
$\beta_{0,5}^-$	max:	946,4	(15)	2,295 (7)
$\beta_{0,5}^-$	avg:	324	(1)	
$\beta_{0,4}^{-}$	max:	1247,7	(15)	0,0053~(10)
$\beta_{0,4}^{-}$	avg:	450	(1)	
$\beta_{0,3}^-$	max:	$1578,\!8$	(15)	4,815~(29)
$\beta_{0,3}^{-}$	avg:	593	(1)	
$\beta_{0,2}^{-}$	max:	1655,7	(15)	2,472 (33)
$\beta_{0,2}^{\circ,-}$	avg:	627	(1)	
$\beta_{0,1}^{-1}$	max:	$2301,\! 6$	(15)	23,44 (28)
$\beta_{0.1}^{-,1}$	avg:	918	(1)	, , ,
- / -				

5 Photon Emissions

5.1 X-Ray Emissions

		${ m Energy}\ { m keV}$		Photons per 100 disint.	
XL	(Te)	3,3348 - 4,8228		0,0449 (9)	
$XK\alpha_2$	(Te)	27,202		0,1252 (18)	$K\alpha$
$XK\alpha_1$	(Te)	$27,\!4726$		0,233 (3)	}
$egin{array}{c} { m XK}eta_3 \ { m XK}eta_1 \ { m XK}eta_5^{\prime\prime} \end{array}$	(Te) (Te) (Te)	30,9446 30,996 31,236	} } }	0,0667~(12)	$\mathrm{K}'eta_1$
$\begin{array}{l} \mathrm{XK}\beta_2\\ \mathrm{XK}\beta_4\\ \mathrm{XKO}_{2,3} \end{array}$	(Te) (Te) (Te)	31,7008 31,774 31,812	} } }	0,0145~(5)	$\mathrm{K}'eta_2$

5.2 Gamma Emissions

	Energy keV	Photons per 100 disint.
$\begin{array}{c} \gamma_{14,12}(\text{Te}) \\ \gamma_{(-1,1)}(\text{Te}) \\ \gamma_{14,10}(\text{Te}) \\ \gamma_{20,14}(\text{Te}) \\ \gamma_{10,6}(\text{Te}) \\ \gamma_{23,14}(\text{Te}) \end{array}$	$\begin{array}{c} 148,02 \ (5) \\ 159,867 \ (35) \\ 189,57 \ (2) \\ 210,40 \ (2) \\ 254,42 \ (1) \\ 291,79 \ (3) \end{array}$	$\begin{array}{c} 0,0037\ (6)\\ 0,0049\ (6)\\ 0,0043\ (5)\\ 0,0053\ (7)\\ 0,0142\ (9)\\ 0,0069\ (7) \end{array}$
	Energy	Photons
--------------------------------------------	-----------------	-------------------
	keV	per 100 disint.
$\gamma_{10.5}(\text{Te})$	335.80(2)	0.0725(9)
$\gamma_{20,11}(Te)$	370.27(3)	0.0286(11)
$\gamma_{20,11}(Te)$	399.97(1)	0.1264(31)
$\gamma_{14.6}(Te)$	444.00(2)	0.195(16)
$\gamma_{20.9}(Te)$	468.84(3)	0.0459(26)
$\gamma_{23,10}(Te)$	481.36 (2)	0.0232(31)
$\gamma_{14,5}(Te)$	525.36(3)	0.1451(35)
$\gamma_{26,12}(Te)$	530.46(7)	0.036(9)
$\gamma_{26,12}()$ $\gamma_{26,10}(Te)$	572.01(5)	0.0176(8)
$\gamma_{10}(Te)$	602,7260 (23)	97.775(20)
$\gamma_{5,3}(\text{Te})$	632.40(2)	0.1029(21)
$\gamma_{2,1}(\text{Te})$	645.8520(19)	7,422(15)
$\gamma_{21.6}(Te)$	662.33(1)	0.024(11)
$\gamma_{5,2}(\text{Te})$	709.33(2)	1.363(5)
$\gamma_{6,3}(\text{Te})$	713.776(4)	2,273(7)
$\gamma_{3,1}(\text{Te})$	722.782(3)	10.708(22)
γ_{23} 6(Te)	735.78(2)	0.1312(16)
$\gamma_{7,3}(\text{Te})$	766.17(2)	0.0103(9)
$\gamma_{25.6}(\text{Te})$	775.27(7)	0.0098(4)
$\gamma_{20,0}$ (Te)	790.706(7)	0.7415(24)
$\gamma_{23,5}(Te)$	817.15 (3)	0.0744(12)
$\gamma_{8,3}(\text{Te})$	856.87(3)	0,0227(5)
$\gamma_{9,3}(\text{Te})$	899.32(3)	0,0179(7)
$\gamma_{10,3}(Te)$	968,195(4)	1,887 (10)
$\gamma_{9,2}(\text{Te})$	976,25(3)	0,0832(7)
$\gamma_{(-1,2)}(\text{Te})$	997.8(3)	0,0033 (23)
$\gamma_{10,2}(Te)$	1045,125(4)	1,852 (14)
$\gamma_{4.1}(\text{Te})$	1053,9(3)	0,0053(10)
$\gamma_{12,2}(Te)$	1086,67(5)	0,0367(9)
$\gamma_{(-1,3)}(\text{Te})$	1235(1)	0,0073 (26)
$\gamma_{15,2}(Te)$	1263,45(7)	0,0422 (19)
$\gamma_{17,2}(Te)$	1301, 14(9)	0,0364 (22)
$\gamma_{3,0}(\text{Te})$	1325,504 (4)	1,587(7)
$\gamma_{5,1}(\text{Te})$	1355,20(2)	1,0412(38)
$\gamma_{20,3}(\text{Te})$	1368,157(5)	2,620(8)
$\gamma_{21,3}(\text{Te})$	1376, 10(1)	0,4999 (43)
$\gamma_{22,3}(\text{Te})$	1385,49(2)	0,062~(6)
$\gamma_{6,1}(\text{Te})$	1436,554(7)	$1,234\ (8)$
$\gamma_{20,2}(\text{Te})$	1445,09(1)	0,334(7)
$\gamma_{7,1}(\text{Te})$	1488,94(2)	$0,\!6770$ (37)
$\gamma_{23,2}(\text{Te})$	$1526,\!48(2)$	0,414~(5)
$\gamma_{25,2}(\text{Te})$	1565,97(7)	0,0109~(12)
$\gamma_{8,1}(\text{Te})$	$1579,\!65$ (3)	0,412~(5)
$\gamma_{9,1}(\text{Te})$	1622, 10 (3)	$0,0416\ (19)$
$\gamma_{10,1}(\text{Te})$	1690,971 (4)	47,46 (19)
$\gamma_{11,1}(\text{Te})$	$1720,\!67$ (3)	0,0946~(6)
$\gamma_{13,1}(\text{Te})$	1852,22 (7)	0,0030 (9)

	Energy keV	Photons per 100 disint.
$\begin{array}{l} \gamma_{16,1}(\mathrm{Te}) \\ \gamma_{18,1}(\mathrm{Te}) \\ \gamma_{6,0}(\mathrm{Te}) \\ \gamma_{19,1}(\mathrm{Te}) \\ \gamma_{20,1}(\mathrm{Te}) \\ \gamma_{21,1}(\mathrm{Te}) \\ \gamma_{22,1}(\mathrm{Te}) \\ \gamma_{23,1}(\mathrm{Te}) \\ \gamma_{23,1}(\mathrm{Te}) \\ \gamma_{27,1}(\mathrm{Te}) \\ \gamma_{10,0}(\mathrm{Te}) \\ \gamma_{11,0}(\mathrm{Te}) \\ \gamma_{13,0}(\mathrm{Te}) \\ \gamma_{19,0}(\mathrm{Te}) \\ \gamma_{20,0}(\mathrm{Te}) \end{array}$	$\begin{array}{c} 1918,74\ (6)\\ 2016,34\ (6)\\ 2039,27\ (1)\\ 2079,75\ (13)\\ 2090,930\ (7)\\ 2098,88\ (1)\\ 2108,27\ (2)\\ 2172,32\ (2)\\ 2172,32\ (2)\\ 2182,37\ (3)\\ 2283,62\ (6)\\ 2293,69\ (1)\\ 2323,39\ (3)\\ 2454,93\ (7)\\ 2682,47\ (13)\\ 2693,65\ (1)\\ \end{array}$	$\begin{array}{c} 0,0529\ (5)\\ 0,0098\ (8)\\ 0,0631\ (5)\\ 0,0224\ (22)\\ 5,493\ (24)\\ 0,0471\ (33)\\ 0,0444\ (23)\\ 0,0029\ (16)\\ 0,04147\ (31)\\ 0,0059\ (5)\\ 0,0327\ (41)\\ 0,0025\ (6)\\ 0,00160\ (12)\\ 0,00176\ (6)\\ 0,0032\ (14) \end{array}$
$\gamma_{24,0}(\text{Te})$	2807,52 (24)	0,0012~(5)

6 Main Production Modes

 $Sb - 123(n,\gamma)Sb - 124$ $\sigma: 3,88$ (12) barns Possible impurities: Sb - 122

7 References

- L.M. LANGER, N.H.LAZAR, R.J.D.MOFFAT. Phys. Rev. 91 (1953) 338 (Beta emission probabilities)
- J.MOREAU. Comp. Rend. Acad. Sci. (Paris) 239 (1954) 800 (Beta emission probabilities)
- T.Azuma. J. Phys. Soc. Jpn 10 (1955) 167 (Beta emission probabilities)
- A.V.ZOLOTAVIN, E.P.GRIGORIEV, M.A.ABROVIAN. Izvest.Akad.Nauk SSSR, Ser.Fiz. 20 (1956) 289 (Columbia Tech.Transl. 20, 271 (1957))
- R.L.MACKLIN. Nucl. Instrum. Methods 1 (1957) 335 (Half-life)
- A.V.ZOLOTAVIN, E.P.GRIGORIEV, M.A.ABROVIAN. Izvest.Akad.Nauk SSSR, Ser.Fiz. 20, 289 (1956); Columbia Tech.Transl. 20 (1957) 271 (Beta emission probabilities)
- C.H.JOHNSON, A.GALONSKY, J.P.ULRICH. Phys. Rev. 109 (1958) 1243 (Half-life)
- J.P.CALI, L.F.LOWE. Nucleonics 17, 10 (1959) 86 (Half-life)
- S.T.HSUE, L.M.LANGER, S.M.TANG, D.A.ZOLLMAN. Nucl. Phys. 73 (1965) 379 (Beta emission probabilities)
- D.M.FLEMING, I.T. MYERS. Int. J. Appl. Radiat. Isotop. 17 (1966) 251 (Half-life)
- P.H. Stelson. Phys. Rev. 157 (1967) 1098 (ICC)
- E.P.GRIGORIEV, A.V. ZOLOTAVIN, V.O. SERGEEV, M.I. SOVTSOV. Izv. Akad. Nauk SSSR. Ser. Fiz. 32 (1968) 733
- (K-Conv. Elec. emission probabilities)
- S.A.REYNOLDS, J.F.EMERY, E.I.WYATT. Nucl. Sci. Eng. 32 (1968) 46 (Half-life)

- E.P.GRIGORIEV, A.V. ZOLOTAVIN, V.O. SERGEEV, M.I. SOVTSOV. Bull. Ac. Sci. USSR. Phys. Ser. 32 (1968)711(K-Conv. Elec. emission probabilities) - R.C.RAGAINI, W.B. WALTERS, R.A. MEYER. Phys. Rev. 187 (1969) 1721 (K ICC Mixing Ratio) - R.A. MEYER, W.B. WALTERS, R.C. RAGAINL. Nucl. Phys. A127 (1969) 595 (Gamma-ray emission probabilities Spin and Parity Gamma-ray energies) - J.R.SITES, W.A. STEYERT. Nucl. Phys. A156 (1970) 19 (Mixing Ratio) - Z.W.GRABOWSKI, K.S.KRANE, R.M.STEFFEN. Phys. Rev. C3 (1971) 1649 (Mixing Ratio) - K.R.BAKER, J.H.HAMILTON, A.V.RAMAYYA, G.HIGHLAND. Nucl. Phys. A186 (1972) 493 (Mixing Ratio) - R.L.HEATH. Report ANCR-1000-2 (1974) (Gamma-ray energies Gamma-ray emission probabilities) - J.R.JOHNSON, K.C. MANN. Can. J. Phys. 52 (1974) 406 (Gamma-ray emission probabilities K-Conv. Elec. emission probabilities) - A.K.SHARMA, R.KAUR, H.R. VERNA, K.K. SURI, P.N. TREHAN, J. Phys. Soc. Jap. 46 (1979) 1057 (Gamma-ray emission probabilities) - S.J.ROBINSON, W.D.HAMILTON, D.M.SNELLING. J. Phys. (London) G9 (1983) 921 (Mixing ratio) - Y.IWATA, M.YASUHARA, K.MAEDA, Y.YOSHIZAWA. Nucl. Instrum. Methods 219 (1984) 123 (Gamma-ray emission probabilities) - YOU JIANMING, LIU YUNZUO, HU DAILING. Z. Physik A331 (1988) 391 (Gamma-ray emission probabilities) - R.A. MEYER. Fizika (Zagreb) 22 (1990) 153 (Gamma-ray emission probabilities) - S.Subrahmanyeswara Rao, K.Bhaskara Rao, V.Seshagiri Rao, H.C.Padhi. Nuovo Cim. 103A (1990) 803 (ICC) - J.GOSWAMY, B.CHAND, D.MEHTA, N.SINGH, P.N.TREHA. Appl. Rad. Isotopes 44 (1993) 541 (Gamma-ray emission probabilities) - E. SCHÖNFELD, H. JANSSEN. Nucl. Instrum. Meth. Phys. Res. A369 (1996) 527 (Atomic data) - I.A.KHARITONOV, T.E.SAZONOVA, S.V.SEPMAN, T.I.SHILNIKOVA, A.V.ZANEVSKY. Appl. Rad. Isotopes 52 (2000) 415(Half-life) - R.G. HELMER, C. VAN DER LEUN. Nucl. Instrum. Methods Phys. Res. A450 (2000) 35 (Gamma energy) - C.DOLL, H.LEHMANN, H.G.BORNER, T.VON EGIDY. Nucl. Phys. A672 (2000) 3 (Nuclear structure) - I.M.BAND, M.B.TRZHASKOVSKAYA. At. Data. Nucl. Data Tables 88,1 (2002) (Theoretical ICC) - G. AUDI, A.H. WAPSTRA, C. THIBAULT. Nucl. Phys. A729 (2003) 337-676 (Q value) - A.PATIL, D.SANTHOSH, K.V.SAI, M.SAINATH, K.VENKATARAMANIAH. Appl. Rad. Isotopes 64 (2006) 693 (Gamma-ray emission probabilities)
 - T.KIBÉDI, T.W.BURROWS, M.B.TRZHASKOVSKAYA, P.M.DAVIDSON, C.W.NESTOR, JR. Nucl. Instrum. Methods Phys. Res. A589 (2008) 202 (ICC)

LNE-LNHB/CEA - Table de Radionucléides



 $\boldsymbol{\gamma}$ Emission intensities per 100 disintegrations



LNE-LNHB/CEA - Table de Radionucléides

N

60,208 (11) d



γ Emission intensities per 100 disintegrations





ÉDITÉ PAR LA DIRECTION DES SYSTÈMES D'INFORMATION

CEA / SACLAY 91191 GIF-SUR-YVETTE CEDEX FRANCE